

CALIBRATION CERTIFICATE

This is to certify that the system maintenance & calibration was performed on **07th APRIL 2021** for the VITROS ECIQ IMMUNOASSAY system installed at **CIHSR** laboratory, 4th mile, Dimapur bearing Serial No.30005029

The system's calibration includes checking the reproducibility performance of the instrument as per the guidelines provided by the manufacture.

Validity: Six months for Optics Calibration.

For *Ortho Clinical Diagnostics*.



Monuj Saikia
Sr. Service Engineer