

DAILY QC TEMPERATURE LOG FOR REFRIGERATOR AND / OR FREEZER

DEPARTMENT: **PATHOLOGY LABORATORY**


(LG FRIDGE)


MONTH/YEAR: **JUNE / 2021, JULY 2021**


		1	2	3	4	5	6	7	8	9	10	11	12	13	14	15	16	17	18	19	20	21	22	23	24	25	26	27	28	29	30	31
JUN 2021	REFRIGERATOR																															
	Morning	2°C	8°C	2°C	2°C	4°C	4°C	6°C	4°C	4°C	6°C	2°C	5°C	4°C	4°C	4.5°C	4°C	4.4°C	4.5°C	4.4°C	4.6°C	4.4°C	4.4°C	4°C	6°C	4°C	4.5°C	5°C	5°C	5°C	5°C	
	Evening	6°C	8°C	4°C	4°C	6°C	4°C	4°C	5°C	6°C	6°C	4°C	4.5°C	4.4°C	4°C	4°C	4°C	4.4°C	4°C	4°C	4.4°C	4°C	4.4°C	4°C	6.5°C	4.9°C	5°C	6°C	6°C	6°C	6°C	
	FREZEER																															
	Morning	-20°C	-18°C	-20°C	-24°C	-20°C	-22°C	-20°C	-22°C	-20°C	-21°C	-22°C	-21°C	-18°C	-19°C	-20°C	-24°C	-23°C	-21°C	-18°C	-22°C	-24°C	-20	-18°C	-20°C	-21°C	-21°C	-22°C	-21°C	-22°C	-21°C	
	Evening	-18°C	-14°C	-26°C	-18°C	-28°C	-29°C	-18°C	-23°C	-26°C	-18°C	-24°C	-20°C	-22°C	-21°C	-21°C	-20°C	-20°C	-18°C	-20°C	-19°C	-20°C	-16	-24°C	-24°C	-21°C	-21°C	-21°C	-21°C	-21°C	-21°C	
INITIAL		Goan	Maun	Maun	Goan	Shubh	Shubh	Anish	Anish	Shubh	Atul	Shubh	Anish	Prathap	Maun	Maun	Prathap	Shubh	Shubh	Shubh	Anish	Shubh	Shubh	Goan	ANILAN	SINHA	NEELAM	G. N. M. M.	SARAD	ANILAN	SARAD	
JUL 2021	REFRIGERATOR																															
	Morning	3°C	4°C	4°C	5°C	5°C	4°C																									
	Evening	4°C	5°C	6°C	6°C	6°C	5°C																									
	FREZEER																															
	Morning																															
	Evening																															
INITIAL		Neelam	Shubh	Anish	Shubh	Anish	Maun																									

ACCEPTABLE LIMITS
 REFRIGERATOR 2-8°C
 FREEZER (-10-30°C)

NOTIFY SUPERVISOR AND MAINTENANCE DEPARTMENT IF READING EXCEEDS THE ACCEPTABLE LIMITS.

Doc No 002	DAILY QC TEMPERATURE LOG FOR Refrigerator & or freezer	Page 1 of 1
Issue Number :	Issue Date : 01.06.2021 Amend No : 00 Amend Date : 00	
Issued By	Dr Vimmy Samant 	





CALIBRATION REPORT

Certificate No: SC / 146 / 2020 - 21

Date: 07 / 10 / 2020

Time: 11.20 am

Temp: 24°C

Customer Name: Victor Hospital, Pathology Laboratory, MADGAON, GOA

Serviced By: Samarth Calibrations, Pune


Calibration Done With: Distilled Water

Balance: CA - 64 Serial No: 2902245

Calibration Validity: Till 06 / 10 / 2021

Sr. No	ID NO	Description	Volume	Mean	S. D.	C. V. %	Accuracy (+ / - UI)	
1	VH/LAB/02	Finnpipette Sr. No: W 24968	5 - 50ul	5ul	5.019	0.063	0.359	0.150ul
				50ul	50.05	0.087	0.174	0.300ul
2	VH/LAB/03	I PETTE Sr. No: NIL	5 - 50ul	5ul	5.034	0.069	1.136	0.150ul
				50ul	50.07	0.112	0.224	0.300ul
3	VH/LAB/05	Finnpipette Sr. No: W 24997	5 - 50ul	5ul	5.03	0.051	1.024	0.150ul
				50ul	50.08	0.147	0.294	0.300ul
4	VH/LAB/08	I PETTE Sr. No: NIL	5 - 50ul	5ul	5.029	0.048	0.964	0.150ul
				50ul	50.11	0.145	0.290	0.300ul
5	VH/LAB/06	Finnpipette Sr. No: J 31133	20 - 200ul	20ul	20.05	0.103	0.498	0.600ul
				200ul	200.05	0.147	0.075	1.200ul
6	VH/LAB/07	I PETTE Sr. No: NIL	100 - 1000ul	100ul	100.05	0.028	0.028	1.500ul
				1000ul	1000.34	0.149	0.014	5.000ul

Note: Micropipette Calibration is done with ISO Certified Electronic balance
Electronic Balance is calibrated by N.A.B.L. Accredited Laboratory.
Please refer details attached



Mark the area

SC SAMARTH CALIBRATIONS © 9822267413

CALIBRATION STATUS

Instrument	Centrifuge
Sr. No / ID No.	202001184-12
Vol. / Model	VH/LAB/CEN
Calib. Date	07/10/2020
Due Date	06/10/2021
For Details Refer Calibration Report	Sign <i>[Signature]</i>

Equipment Name: Centrifuge

Model: Electra BL (12T)

Make:- EURO

Year of Make: _____

Date Placed in Service: _____

Range & Accuracy: 100-4000 Rpm

S/N No. 202001184-12

Installed on: _____

PM Done on: 7.10.2020

PM Due on: 6.10.2021

Cal Done on: 7.10.2020

Cal Due on: 6.10.2021

Calibrated by: Samarth Calibrations