

LAB MONTHLY SUMMARY



Lab Name **ALPHA LIFECARE**
 Month **March**
 Constituent Group **Chemistry II**

Lab No **8870**
 Year **2021**



Date of Result Entered : **18/03/2021**

Date of Report Published : **01/04/2021**

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE II	GOD-POD II	Transasia / Erba	791	160.91	10.47	16.84	157.4 mg/dl	-0.21	1.20
2	UREA II	Urease UV / GLDH II	Transasia / Erba	605	42.26	12.78	5.40	54.2 mg/dl	2.21	0.44
3	CREATININE II	Jaffes End point II	Any Analyser (Automation / Semi Automation)	606	1.64	13.84	0.23	1.68 mg/dl	0.18	0.02
4	T.BILIRUBIN II	Diazonium Salt (Colorimetric) / Jendrassik II	Transasia / Erba	658	3.60	13.31	0.48	0.5 mg/dl	-6.46	0.04
5	T-PROTEIN II	Biuret - Colorimetric II	Transasia / Erba	595	5.89	9.22	0.54	5.64 g/dl	-0.46	0.04
6	ALBUMIN II	BCG - Colorimetric II	Transasia / Erba	599	3.52	9.28	0.33	3.2 g/dl	-0.98	0.03
7	URIC ACID II	Enzymatic / Uricase Colorimetric II	Transasia / Erba	668	7.46	13.58	1.01	8.4 mg/dl	0.93	0.08
8	CHOLESTROL II	CHOD-PAP II	Transasia / Erba	717	118.97	12.88	15.32	111 mg/dl	-0.52	1.14
9	TRIGLYCERIDE II	GPO-PAP / Enzymatic Colorimetric / End Point II	Transasia / Erba	665	105.23	19.47	20.49	77.3 mg/dl	-1.36	1.59

SDI Range	Interpretation
Within -1.0 to +1.0	Excellent.
Between ±1.0 to ±2.0	Good.
Between ±2.0 to ±3.0	Accept with caution. Warning Signal.
Beyond ±3.0	Unacceptable performance. Action Signal.

Homogeneity and Stability of the sample is passed.

Data in CMC EQAS reports is confidential

Contact details:

Email: clinqc@cmcvellore.ac.in

Contact Number: 0416-2283102

Pamela Christudoss

Dr. Pamela Christudoss
CMC EQAS Co-Ordinator
Christian Medical College, Vellore

******* End of Report *******