


	Name of Equipment	Serial No	Make	Range & Accuracy	Calibration date	Calibration due on
1	Micro Pipette	MIZ-55 014		5-50 μ L	07-10-2020	06.10.2021
2	Micro Pipette	J3113	Finnpipette	20-200 μ L	07-10-2020	06.10.2021
3	Micro Pipette	W24968	Finnpipette	5-50 μ L	07-10-2020	6.10.2021
4	Micro Pipette	VH/LA B/03	I PETTE	5-50 μ L	07-10-2020	06.10.2021
5	Micro Pipette	W24997	Finnpipette	5-10 μ L	07-10-2020	6.10.2021
6	Micro Pipette	VH/LA B/07	I PETTE	100-1000 μ L	07-10-2020	06.10.2021

Dr. Eugene D'Souza
M.D. (Pathology)
Laboratory Director
Reg. No.: 2691

Equipment name	Model	Make	S/N	PM done on	PM due on	Cal Done on	Cal due on
Immuno Assay Analyzer	ADVIA Centaur CP	Siemens	ROB3000 6369	4.11.2020	3.11.2021	4.11.2020	3.11.2021
ABG Machine	Cobas b221	Roche	27636	10.11.2020	9.11.2021	10.11.2020	9.11.2021
FULLY Auto Analyzer	Mispa Nano	Agappe	WN-8310 7096	21.11.2020	20.11.2021	21.11.2020	20.11.2021
Coagulometer	Hemostar XF 1.0	Tulip Diagnostics	J2320591 4	14.12.2020	13.12.2021	14.12.2020	13.12.2021
Immuno Assay Analyzer	Mini VIDAS	BIOME RIEUX	IVD5211 137	4.11.2020	3.11.2021	4.11.2020	3.11.2021
Blood & Body fluid Analyzer	BacT/Alert	BIOME RIEUX	710BS57 97	4.11.2020	3.11.2021	4.11.2020	3.11.2021
Hot Air Oven	-	Lawrence & Mayo	100702	7.10.2020	6.10.2021	7.10.2020	6.10.2021
Incubator	60x60x60	Lawrence & Mayo		7.10.2020	6.10.2021	7.10.2020	6.10.2021
Autoclave	VAC-30-16	KEC	034/1020	06.11.2020	05.11.2021	6.11.2020	5.11.2021
Nycocard Reader	Reader II	Reader	130818 71047				
I-Chroma	Weldon	Biotech	1R2PK01 2972	25.9.2020	24.9.2021	25.9.2020	24.9.2021
Centrifuge	Electra BL(12T)	Euro	20200118 412	07.10.2020	6.10.2021	7.10.2021	6.10.2021


Dr. Eugene D'Souza
 M.D. (Pathology)
 Laboratory Director
 Reg. No.: 2691



CERTIFICATE OF CALIBRATION

ADL/BO/MS/C/0032/21-22

Instrument : Mispa Nano Fully Automated Biochemistry Analyser
Customer Name : Victor Hospital, Station Malbhat, Near Carmelite Monastery-Madgaon
Date of installation : 13-05-2018 **Serial No** : WN-83107096

Alignment Check

Motion Check	Checked	Remark
Sample Probe	<input checked="" type="checkbox"/>	Ok
Reagent Probe	<input checked="" type="checkbox"/>	Ok
Mixing Bar	<input checked="" type="checkbox"/>	Ok
Reaction Disk	<input checked="" type="checkbox"/>	Ok
Reagent Disk	<input checked="" type="checkbox"/>	Ok
Sample Disk	<input checked="" type="checkbox"/>	Ok
Fluidic disk	<input checked="" type="checkbox"/>	Ok
Lamp	<input checked="" type="checkbox"/>	Ok

System	Checked	Remark
Main unit	<input checked="" type="checkbox"/>	Ok
Temp. unit	<input checked="" type="checkbox"/>	Ok
Mixing unit	<input checked="" type="checkbox"/>	Ok
Reaction unit	<input checked="" type="checkbox"/>	Ok
Reagent unit	<input checked="" type="checkbox"/>	Ok
Sample unit	<input checked="" type="checkbox"/>	Ok

AGAPPE DIAGNOSTICS LTD. ISO 9001:2015 | EN ISO 13485:2016 CERTIFIED COMPANY | CIN : U24239MH1998PLC115413

CORPORATE OFFICE / REAGENT PLANT
 Agappe Hills, Pattimattom (PC), Dist. Ernakulam, Kerala - 683 562, India.
 Tel. +91 484 286 7000 | Email: agappe@agappe.in

EQUIPMENT PLANT
 X/588-CB, Block No. 32, KINFRA Small Industrial Park,
 Nellaiad, Cochin, Kerala, India - 686 721. Tel: +91 484 236 7477.

KOLKATA OFFICE
 406, Mirfin Matrix, Plot No 10, Block DN, Sector V,
 Salt Lake City, Kolkata - 700 091.
 Tel: +91 33 4503 0431 | Email: kolkataoffice@agappe.in

MUMBAI (REGISTERED OFFICE)
 401 & 402, 4th Floor, Jaisingh Business Centre, 119,
 Sahar Road, Parsiwada, Andheri (East), Mumbai - 400 099, India.
 Tel: +91 22 4300 8000 | Email: mumbai@agappe.in

DELHI OFFICE
 DSM 540, 5th Floor, DLF Tower, Shivaji Marg,
 New Delhi - 110 015, India.
 Tel: +91 11 4358 0416 | Email: delhi@agappe.in

BANGALORE OFFICE
 5-6, 11th Floor, Red Cross Bhavan, No.26 Race Course Road,
 Bangalore - 560 001. Tel: +91 80 2228 8288
 Email: bangalore@agappe.in



OPTICAL SECTION

Filter	Dark Current	Background
340	145	44187
405	134	47751
450	121	50107
510	114	51687
546	119	51458
578	109	53692
630	112	52928
670	111	58689
Reference		

Power Supply : 5V 12V 24V

The equipment is calibrated for the different parameters by using Multi Calibrator

Lot.No: 30110040 Exp: 06/2022

This is to certify that the above mentioned product has been calibrated according to the standard operating procedure given by the MINDRAY Medical Systems.

Date of Calibration : 25-MAR-2021

Next Due Date : 24-SEP-2021

For Agappe Diagnostics Ltd,



Antony
Authorized Signatory
 Name: Antony Jose
 Designation: Engineer

AGAPPE DIAGNOSTICS LTD. ISO 9001:2015 | EN ISO 13485:2016 CERTIFIED COMPANY | CNL : 10429988299APL015473

CORPORATE OFFICE / REAGENT PLANT
 Agappe #19, Pattinattom (PO), Bili, Ernakulam, Kerala - 681 562, India.
 Tel : +91 484 295 7000 | Email: agappe@agappe.in

EQUIPMENT PLANT
 K/505-OR, Block No. 22, KINERA Small Industrial Park,
 Netaji, Cochin, Kerala, India - 686 722, Tel: +91 484 236 7471

KOLKATA OFFICE
 606, Merla Mitra, Plot No-12, Block OR, Sector K,
 Salt Lake City, Kolkata - 700 071
 Tel: +91 33 4361 9451 | Email: kolkataoffice@agappe.in

MUMBAI (REGISTERED OFFICE)
 401 & 402, 4th Floor, Jansingh Business Centre, 129,
 Sakar Road, Purwadas, Andheri (East), Mumbai - 400 049, India.
 Tel: +91 22 4300 8000 | Email: mumbaioffice@agappe.in

DELHI OFFICE
 D5M 545, 5th Floor, BJT Tower, Shivaji Marg,
 New Delhi - 110 025, India.
 Tel: +91 11 4228 8416 | Email: delhioffice@agappe.in

BANGALORE OFFICE
 5-4, 1st Floor, Red Cross Shivan, No.26, Back Cross
 Bangalore - 560 002, Tel: +91 98 1228 6258
 Email: bangaloreoffice@agappe.in