

DAILY QC TEMPERATURE LOG FOR REFRIGERATOR AND / OR FREEZER

SAMSUNG FRIDGE 3


MONTH/YEAR: JUNE, JULY 2021

DEPARTMENT: PATHOLOGY LABORATORY

		1	2	3	4	5	6	7	8	9	10	11	12	13	14	15	16	17	18	19	20	21	22	23	24	25	26	27	28	29	30	31	
JUN 2021	REFRIGERATOR																																
	Morning	4°C	4°C	5°C	3°C	3°C	4°C	4°C	4°C	5°C	4°C	4°C	4°C	4°C	4°C	5°C	4.5°C	5°C	4°C	5°C	6°C	4.5°C	4°C	4°C	4°C	4°C	4°C	4°C	6°C	4°C	4°C	5°C	
	Evening	5°C	5°C	6°C	5°C	4°C	5°C	5°C	4°C	5°C	5°C	5°C	5°C	5°C	6°C	6°C	5°C	5.5°C	5°C	5.5°C	7°C	5°C	5°C	5°C	5°C	5.5°C	5°C	5°C	5°C	5°C	6°C	5°C	
	FREEZER																																
	Morning	-21°C	-22°C	-19°C	-21°C	-21°C	-21°C	-21°C	-21°C	-21°C	-21°C	-21°C	-21°C	-21°C	-21°C	-21°C	-21°C	-21°C	-21°C	-21°C	-21°C	-21°C	-21°C	-21°C	-21°C	-21°C	-21°C	-21°C	-21°C	-21°C	-21°C	-21°C	
	Evening	-21°C	-22°C	-20°C	-21°C	-20°C	-21°C	-21°C	-21°C	-21°C	-21°C	-21°C	-21°C	-21°C	-21°C	-21°C	-21°C	-21°C	-21°C	-21°C	-21°C	-21°C	-21°C	-21°C	-21°C	-21°C	-21°C	-21°C	-21°C	-21°C	-21°C	-21°C	
INITIAL																																	
JUL 2021	REFRIGERATOR																																
	Morning	4°C	5°C	3.5°C	4°C	2.9°C	4°C	4°C	4°C	3.5°C	4.5°C	4°C	5°C	4.4°C																			
	Evening	5°C	5.5°C	4°C	5°C	3.5°C	5°C	5°C	5°C	4°C	5°C	5°C	5.5°C																				
	FREEZER																																
	Morning	-21°C	-19°C	-21°C	-21°C	-21°C	-21°C	-21°C	-21°C	-21°C	-21°C	-21°C	-21°C	-21°C	-21°C	-21°C	-21°C	-21°C	-21°C	-21°C	-21°C	-21°C	-21°C	-21°C	-21°C	-21°C	-21°C	-21°C	-21°C	-21°C	-21°C	-21°C	
	Evening	-21°C	-22°C	-21°C	-21°C	-21°C	-21°C	-21°C	-21°C	-21°C	-21°C	-21°C	-21°C	-21°C	-21°C	-21°C	-21°C	-21°C	-21°C	-21°C	-21°C	-21°C	-21°C	-21°C	-21°C	-21°C	-21°C	-21°C	-21°C	-21°C	-21°C	-21°C	
INITIAL																																	

ACCEPTABLE LIMITS
 REFRIGERATOR 2-8°C
 FREEZER (-10-30°C)

NOTIFY SUPERVISOR AND MAINTENANCE DEPARTMENT IF READING EXCEEDS THE ACCEPTABLE LIMITS.

Doc No 002	DAILY QC TEMPERATURE LOG FOR Refrigerator & or freezer		Page 1 of 1
Issue Number :	Issue Date : 01.06.2021	Amend No : 00	Amend Date : 00
Issued By	Dr Vimmy Samant 		

SAMSUNG

VH | 49B/20-21
015 (glockenicht)

