





VH/L98/20-21
014 (Heart)



DAILY QC TEMPERATURE LOG FOR REFRIGERATOR AND / OR FREEZER

MONTH/YEAR: JUNE | 2021, JULY 2021

DEPARTMENT: PATHOLOGY LABORATORY

(LG FRIDGE)

	1	2	3	4	5	6	7	8	9	10	11	12	13	14	15	16	17	18	19	20	21	22	23	24	25	26	27	28	29	30	31	
JUN 2021	REFRIGERATOR																															
Morning	2°C	8°C	2°C	2°C	4°C	4°C	6°C	4°C	4°C	6°C	2°C	5°C	4°C	4°C	4.5°C	4°C	4.4°C	4.5°C	4.4°C	4.6°C	4.4°C	4.4°C	4°C	6°C	4°C	4.5°C	5°C	5°C	5°C	5°C	5°C	
Evening	6°C	8°C	4°C	4°C	6°C	4°C	4°C	5°C	6°C	6°C	4°C	4.5°C	4.4°C	4°C	4°C	4°C	4.4°C	4.5°C	4°C	4.4°C	4°C	4.4°C	4°C	6.5°C	4.9°C	5°C	6°C	6°C	6°C	6°C		
JUN 2021	FREEZER																															
Morning	-20°C	-18°C	-20°C	-21°C	-20°C	-22°C	-20°C	-22°C	-20°C	-21°C	-22°C	-21°C	-18°C	-19°C	-20°C	-24°C	-23°C	-21°C	-18°C	-22°C	-24°C	-20	-18°C	-20°C	-21°C	-21°C	-22°C	-21°C	-22°C	-21°C		
Evening	-18°C	-14°C	-26°C	-18°C	-28°C	-29°C	-18°C	-28°C	-26°C	-18°C	-24°C	-20°C	-22°C	-21°C	-21°C	-20°C	-20°C	-18°C	-20°C	-19°C	-20°C	-16	-21°C	-24°C	-21°C	-21°C	-21°C	-21°C	-21°C	-21°C		
INITIAL	Gina	Manish	Manish	Gina	Shubha	Shubha	Arish	Arish	Shubha	Arish	Shubha	Arish	Prathna	Meenal	Meenal	Prathna	Shubha	Shubha	Shubha	Arish	Shubha	Shubha	Gina	ANIKAM	Shubha	MEENAM	GIMMA	SANAA	ANIKAM	SHUBHA		
JUL 2021	REFRIGERATOR																															
Morning	3°C	4°C	4°C	5°C	5°C	4°C	5°C	2.9°C	2.8°C																							
Evening	4°C	5°C	6°C	6°C	6°C	5°C	6°C	3.5°C																								
JUL 2021	FREEZER																															
Morning	-21°C	-20°C	-21°C	-20°C	-21°C	-20°C	-21°C	-21°C	-21°C																							
Evening	-21°C	-21°C	-21°C	-21°C	-21°C	-21°C	-21°C	-21°C	-21°C																							
INITIAL	HELEN	SHUBHA	ANIKAM	SHUBHA	ANIKAM	ANIKAM	ANIKAM	ANIKAM	SHUBHA	SHUBHA																						

ACCEPTABLE LIMITS

REFRIGERATOR 2-8°C

FREEZER (-10-30°C)

NOTIFY SUPERVISOR AND MAINTENANCE DEPARTMENT IF READING EXCEEDS THE ACCEPTABLE LIMITS.

Doc No 002

Issue Number :

Issued By

DAILY QC TEMPERATURE LOG FOR Refrigerator & or freezer

Issue Date : 01.06.2021

Amend No : 00

Amend Date : 00

Page 1 of 1

Dr Vimmy Samant

