



Asia

SAMSUNG

VH/L00/2-01
016 (Sunday)



DAILY QC TEMPERATURE LOG FOR REFRIGERATOR AND / OR FREEZER

DEPARTMENT: **PATHOLOGY LAB**

(**SAMSUNG FRIDGE: 1**)

MONTH/YEAR: **JUNE 2021, JULY 2021**

		1	2	3	4	5	6	7	8	9	10	11	12	13	14	15	16	17	18	19	20	21	22	23	24	25	26	27	28	29	30	31
JUN 2021	REFRIGERATOR																															
	Morning	4°C	5°C	4°C	4°C	5°C	7°C	6°C	5°C	4°C	4.5°C	6.5°C	6°C	4°C	6°C	7°C	6°C	3.5°C	5°C	6.6°C	4.9°C	6.7°C	6.5°C	6.6°C	6.5°C	6.5°C	6°C	6°C	6°C	7°C	7°C	
	Evening	5°C	6°C	5°C	6°C	6.5°C	5.5°C	7°C	6°C	5°C	5°C	7°C	7°C	5.5°C	7°C	7.2°C	7°C	4.9°C	6°C	7.0°C	5°C	7°C	7°C	7°C	7°C	7°C	6.5°C	6°C	6.5°C	7°C	6°C	
	FREEZER																															
	Morning	-2°C	-2°C	-2°C	-1°C	-1°C	-2°C	-1°C	-2°C	-2°C	-1°C	-2°C	-2°C	-2°C	-2°C	-2°C	-2°C	-2°C	-2°C	-2°C	-2°C	-2°C	-2°C	-2°C	-2°C	-2°C	-2°C	-2°C	-2°C	-2°C	-2°C	-2°C
	Evening	-2°C	-1°C	-2°C	-2°C	-2°C	-2°C	-2°C	-2°C	-2°C	-2°C	-2°C	-2°C	-2°C	-2°C	-2°C	-2°C	-2°C	-2°C	-2°C	-2°C	-2°C	-2°C	-2°C	-2°C	-2°C	-2°C	-2°C	-2°C	-2°C	-2°C	-2°C
INITIAL		Sneha	Amu	Mehar	Amu	Pran	Mehar	Amu	Sneha	Amu	Pran	Mehar	Mehar	Amu	Amu	Sneha	Amu	Amu	Mehar	Mehar	Sneha	Amu	Sneha	Amu	Amu	Amu	Amu	Amu	Amu	Amu	Sneha	
JUL 2021	REFRIGERATOR																															
	Morning	4°C	5°C	3°C	4°C	5°C	4°C	4°C	5°C	6°C	5°C	4°C	5°C	5.5°C																		
	Evening	5°C	6°C	4°C	5°C	5°C	5°C	6°C	6°C	5.5°C	4.5°C	6°C																				
	FREEZER																															
	Morning	-2°C	-1°C	-2°C	-2°C	-2°C	-2°C	-2°C	-2°C	-2°C																						
	Evening	-2°C	-1°C	-2°C	-2°C	-2°C	-2°C	-2°C	-2°C	-2°C																						
INITIAL		Amu	Mehar	Amu	Mehar	Amu	Amu	Amu	Amu	Sneha																						

ACCEPTABLE LIMITS
REFRIGERATOR 2-8°C
FREEZER (-10-30°C)

NOTIFY SUPERVISOR AND MAINTENANCE DEPARTMENT IF READING EXCEEDS THE ACCEPTABLE LIMITS.

Doc No 002
Issue Number :
Issued By :
DAILY QC TEMPERATURE LOG FOR Refrigerator & or freezer
Issue Date : 01.06.2021
Amend No : 00
Amend Date : 00
Dr Vimmy Samant

