



SC SAMARTH CALIBRATIONS
CALIBRATION STATUS
9822267413
Instrument: MicroPipette
Sr. No / ID No: W24997 S-50ul
Vol. / Model: W11081 QS

W24997

SC SAMARTH CALIBRATIONS
CALIBRATION STATUS
9822267413
Instrument: MicroPipette
Sr. No / ID No: W11081 QS
Vol. / Model: W11081 QS

Vol. Range : 5-50 µl

KN

CALIBRATION REPORT

Certificate No: SC / 146 / 2020 - 21

Date: 07 / 10 / 2020

Time: 11.20 am

Temp: 24°C

Customer Name: Victor Hospital, Pathology Laboratory, MADGAON, GOA

Serviced By: Samarth Calibrations, Pune

Calibration Done With: Distilled Water

Balance: CA - 64 Serial No: 2902245

Calibration Validity: Till 06 / 10 / 2021

Sr. No	ID NO	Description	Volume	Mean	S. D.	C. V. %	Accuracy (+ / - UI)	
1	VH/LAB/02	Finnpipette Sr. No: W 24968	5 - 50ul	5ul	5.019	0.063	0.359	0.150ul
				50ul	50.05	0.087	0.174	0.300ul
2	VH/LAB/03	I PETTE Sr. No: NIL	5 - 50ul	5ul	5.034	0.069	1.136	0.150ul
				50ul	50.07	0.112	0.224	0.300ul
3	VH/LAB/05	Finnpipette Sr. No: W 24997	5 - 50ul	5ul	5.03	0.051	1.024	0.150ul
				50ul	50.08	0.147	0.294	0.300ul
4	VH/LAB/08	I PETTE Sr. No: NIL	5 - 50ul	5ul	5.029	0.048	0.964	0.150ul
				50ul	50.11	0.145	0.290	0.300ul
5	VH/LAB/06	Finnpipette Sr. No: J 31133	20 - 200ul	20ul	20.05	0.103	0.498	0.600ul
				200ul	200.05	0.147	0.075	1.200ul
6	VH/LAB/07	I PETTE Sr. No: NIL	100 - 1000ul	100ul	100.05	0.028	0.028	1.500ul
				1000ul	1000.34	0.149	0.014	5.000ul

Note: Micropipette Calibration is done with ISO Certified Electronic balance
Electronic Balance is calibrated by N.A.B.L. Accredited Laboratory.
Please refer details attached

