

LAB MONTHLY SUMMARY



Lab Name **PRIMARY HEALTH CENTRE** Lab No **9960**
 Month **June** Year **2021**
 Constituent Group **Chemistry II**



Date of Result Entered : 08/06/2021

Date of Report Published : 03/07/2021

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants CV	Participants SD	Your Value	SDI	U
1	GLUCOSE II	GOD-POD II	Agape	270	200.27	9.32	18.66	215.71 mg/dl	0.83	2.27
2	UREA II	Urease UV / GLDH II	Agape	230	77.13	10.18	7.85	75.17 mg/dl	-0.25	1.04
3	CREATININE II	Jaffes Kinetic-Alkaline Picrate II	Agape	186	3.94	10.29	0.40	4.34 mg/dl	0.99	0.06
4	T.BILIRUBIN II	Diazonium Salt (Colorimetric) / Jendrassik II	Agape	216	3.83	13.73	0.52	2.47 mg/dl	-2.59	0.07
5	CHOLESTROL II	CHOD-PAP II	Agape	266	102.59	11.47	11.76	108.44 mg/dl	0.50	1.44

SDI Range	Interpretation
Within -1.0 to +1.0	Excellent.
Between ±1.0 to ±2.0	Good.
Between ±2.0 to ±3.0	Accept with caution. Warning Signal.
Beyond ±3.0	Unacceptable performance. Action Signal.

Homogeneity and Stability of the sample is passed.

Data in CMC EQAS reports is confidential

Contact details:

Email: clinqc@cmcvellore.ac.in

Contact Number: 0416-2283102

Dr. Pamela Christudoss
 CMC EQAS Co-Ordinator
 Christian Medical College, Vellore

***** End of Report *****