

APK300 Calibration Certificate.

This is to certify that installed at NU Hospitals, Ambur. AutoPak300 Analyzer Serial number **20016904** has been duly calibrated on 11/05/2021 and the same was verified with the following tests.

System Tests results:


	Passed	Failed	Comments
● Temperature	<input checked="" type="checkbox"/>	<input type="checkbox"/>
● Stray light	<input checked="" type="checkbox"/>	<input type="checkbox"/>
● Noise	<input checked="" type="checkbox"/>	<input type="checkbox"/>
● Stability	<input checked="" type="checkbox"/>	<input type="checkbox"/>
● Tip pump	<input checked="" type="checkbox"/>	<input type="checkbox"/>
● Level detection	<input checked="" type="checkbox"/>	<input type="checkbox"/>
● Washer hydraulics	<input checked="" type="checkbox"/>	<input type="checkbox"/>
● Washer	<input checked="" type="checkbox"/>	<input type="checkbox"/>
● Clot detector	<input checked="" type="checkbox"/>	<input type="checkbox"/>

Verified the Calibrations and alignment visually.

Performed Tests , all are passed.

Calibration of the instrument was carried out as per the recommendation.

Next Calibration will be performed on 11/05/2022

Name/Address of Lab: KM NU Hospitals, Solur, Ambur Tamilnadu	
Phone#:9751740906	Contact Person: Louis
MAIL:guru.yogesh@siemens-healthineers.com	Contact Email:lab.ambur@nuhospitals.com
Instrument Model: Autopak 300	Instrument Serial #: 20016904
Installation Date: 09.05.2020	Software Revision #: 1.5
Install Technician Name: Guru Yogesh	Signature: 

Guruyogesh.R

11/05/2021

Service Engineer, Bangalore

