



## CALIBRATION CERTIFICATE

**Name of the customer:** ALPHA CLINICAL LAB  
**Instrument type:** Hematology analyzer  
**Model number:** CBC 330 **Serial number:** 2900PET2708  
**Date of Calibration:** 10-08-2020 **Valid till date:** 09-08-2021

**Power checks:**

Power supply	AC 228V-50 Hz
Plug top	5 Amps
Earth voltage	1.1 Volt

**Environmental check:**

Parameter	Working range	Site conditions
Room temperature	5-40°C	27°C
Humidity	10-90%	52%

**Physical check of the equipment:**

Check for damage	Checked	No damage found
Visual and functional inspection	Checked	OK
No damage on internal assembly	Checked	No damage found

**System. configuration check:**

Positioning of the Equipment	flat stable surface	Checked	Found OK
Adequate space around the Sphera		Checked	Found OK
Interference to other equipment		Checked	No interference

**Result**

Installed the hematology analyzer satisfactorily and is ready for operational verification

for Mr. Senthil Kumar.R  
South Service Manager