



COMMITTED TO THE
CUSTOMER SINCE - 1996

VAIDYANATHESHWARA INSTRUMENTS

CERTIFICATE OF CALIBRATION



No. 301/A, 9th Main Road, 3rd Cross, Rajiv Gandhi Nagar, J.B. Kaval, Nandhini Layout Post, Bangalore - 560 096.

Contact : 080-23377266, Mob : 9986586789 / 9448080177 / 9964308118 | Email : info@viplgroup.com Web : www.viplgroup.com

NABL Accredited Calibration Lab as per ISO/IEC 17025 : 2017 With vide Certificate No: CC-2473

Page 1 of 2

1 Name and Address of the Customer : M/s. KM NU HOSPITALS., (A Unit of NU Hospital Pvt Ltd.,)
NH 48, # 75/2F2, M.C. Road, Solur,
Ambur-635 814, Tirupattur(Dt.), Tamil Nadu.

2 Customer Reference

2.1 ULR No : CC247321100008540F
2.2 SRF No : 7798
2.3 Certificate No. : VI/20-21/7798-01
2.4 Format No. : VI-FRM-ME-105
2.5 DC No & Date : 217 & 17-03-2021
2.6 Receipt Date : 18 - 03 - 2021
2.7 Date Of Issue : 20 - 03 - 2021

3 Details Of Device Under Calibration(DUC).

3.1 Nomenclature : Pipette
3.2 Make : P'fact
3.3 Range : 100 - 1000 µl
3.4 SL No. : 307877
3.5 DUC Condition : Satisfactory
3.6 Calibration Procedure No. : SOP-16-71 Based On ISO 8655-6:2002 E
3.7 No.of Pages : 2
3.8 Calibration Date : 20-03-2021
3.9 Calibration Due : 19-03-2022
3.10 Calibration done at : VI Volumetric Lab
3.11 Discipline : Mechanical (Mass & volume)

4 Environmental Condition

Temperature 20.2 °C Humidity 51 %Rh

5 Standards Used for calibration

Sl. No.	Nomenclature	Make & Model	Sl. No	Range	Traceable to /Cert. No.	Validity
1	Digital Weighing Balance	RADWAG/MYA5.4Y	544953	0.1mg to 5g	VI/20-21/INT-ME-121	18-09-2021
2	E1 Class Weights	LCGC	VI-E1-ME-001	1mg to 200g	TVCSP1 19/04/339-01	02 - 04 - 2022

6 Note:

- 6.1. The Calibration Certificate relates only to the above DUC
- 6.2. Publication or reproduction of this Certificate in any form other than by complete set of the whole report & in the language, written, is not permitted without the written consent of VI Lab..
- 6.3. Corrections/erasing, invalidate the Calibration Certificate.
- 6.4. Calibration of the DUC are traceable to National standards/International Standards
- 6.5. Any error in this Certificate should be brought to our knowledge within 30 days from the date of this Cert.
- 6.6. Results Reported are valid at the time of and under the stated conditions of measurements.
- 6.7. The usage of NABL symbol is as per NABL guidelines given in NABL-133.

Calibrated By


Sameer M S

(Calibration Engineer)

Checked By

P. Santhosh Kumar

(Lab- In-Charge)





COMMITTED TO THE
CUSTOMER SINCE - 1996

VAIDYANATHESHWARA INSTRUMENTS

CERTIFICATE OF CALIBRATION



No. 301/A, 9th Main Road, 3rd Cross, Rajiv Gandhi Nagar, J.B. Kaval, Nandhini Layout Post, Bangalore - 560 096.
Contact : 080-23377266, Mob : 9986586789 / 9448080177 / 9964308118 | Email : info@viplgroup.com Web : www.viplgroup.com

NABL Accredited Calibration Lab as per ISO/IEC 17025 : 2017 With vide Certificate No: CC-2473

Certificate No: VII/20-21/7798-01

Page 2 of 2

Range: 100 - 1000 μ l

Results:

Range	Cal Point	Volume observed	Expanded Uncertainty \pm
	μ l	μ l	μ l
100 - 1000 μ l	100.0	99.725	0.62
	500.0	499.021	3.10
	1000.0	998.567	6.20

Note:

1. Visual Inspection : Found Well.
2. Tripple Distilled water is used to Calibrate the MicroPipette.

Conclusion /Remarks:

1. Ref. standard used are traceable to National/International Standard,
2. The Expanded Uncertainty of associated with measurement at approximate 95.45% confidence level with coverage factor k=2

Calibrated By

Sameer M S
(Calibration Engineer)

Checked By

P. Santhosh Kumar
(Lab- In-Charge)

