

**LIFE CARE DIAGNOSTIC (556)**

**Outlier And Analyte Summary Report**

**Outlier Details For Cycle No 210207 and Sample No 01**

**Report Date : 19/07/2021**

Analyte	Instrument	Result Value	Standard Unit	Z-Score
X Chloride	Robonik Prietest Touch	112.00	mEq/L	3.57
X Sodium	Robonik Prietest Touch	181.00	mEq/L	22.62
X Potassium	Robonik Prietest Touch	6.50	mEq/L	8.79
Legend @ : z score ≤ 2.0 - Acceptable ! : 2.0 < z score < 3.0 - Warning X : z score ≥ 3 - Unacceptable # : Not Evaluated ⌚ : Delayed Result Entry * : Not considered for evaluation.				

**Problem Classification:** \_\_\_\_\_

\_\_\_\_\_

\_\_\_\_\_

\_\_\_\_\_

**Corrective Action:** \_\_\_\_\_

\_\_\_\_\_

\_\_\_\_\_

\_\_\_\_\_

**Reviewed by:** \_\_\_\_\_

**Dated:** \_\_\_\_\_

**LIFE CARE DIAGNOSTIC (556)**

**Outlier And Analyte Summary Report**

**Result Details For Cycle No 210207 and Sample No 01**

**Report Date : 19/07/2021**

**Instrument : Robonik Prietest Touch**

Analyte	StandardUnit	Result Value	Mean	Z-Score	RMZ
@ Bilirubin Total	mg/dl	0.39	--	0.33	0.09
@ Bilirubin Direct	mg/dl	0.12	--	0.33	0.17
@ Total Protein	g/dL	6.50	--	0.32	0.35
@ Albumin	g/dL	3.50	--	-0.94	-0.32
@ Uric Acid	mg/dl	4.00	--	-0.56	-0.23
@ Calcium	mg/dl	7.52	--	-1.00	-0.77
@ Triglycerides	mg/dl	147.20	--	-0.10	-0.22
@ ALT	U/L	27.80	--	0.30	0.90
@ AST	U/L	23.10	--	0.23	0.46
@ ALP	U/L	66.30	--	-0.66	-0.44
@ LDH	U/L	266.70	--	0.03	-0.09
X Chloride	mmol/L	112.00	--	3.57	3.82
X Sodium	mmol/L	181.00	--	22.62	23.33
X Potassium	mmol/L	6.50	--	8.79	7.75

Legend @ : z score ≤ 2.0 - Acceptable  
! : 2.0 < z score < 3.0 - Warning  
X : z score ≥ 3 - Unacceptable  
# : Not Evaluated  
🕒 : Delayed Result Entry  
\* : Not considered for evaluation.

Total Parameters	14
Not Evaluated Parameters	0
Evaluated Parameters	14
Outlier Parameters (X)	3
EQAS Score Biochemistry	78.57 %

*Dr Puneet Kumar Nigam*

**Dr Puneet Kumar Nigam**  
**PT coordinator & Technical Manager, MHL EQAS**  
Unit No. 409-416.  
Commercial Building - 1A,  
Kohinoor Mall, Kiroi Road, Kurla (W),  
Mumbai - 400070