

LIFE CARE DIAGNOSTIC (556)

Outlier And Analyte Summary Report

Outlier Details For Cycle No 210207 and Sample No 02

Report Date : 19/07/2021

Analyte	Instrument	Result Value	Standard Unit	Z-Score
X Lymphocytes Differential.	Zybio Z3	5.49	%	-3.23
X Neutrophils Differential.	Zybio Z3	92.84	%	5.32
<p>Legend @ : z score \leq 2.0 - Acceptable ! : 2.0 < z score < 3.0 - Warning X : z score \geq 3 - Unacceptable # : Not Evaluated 🕒 : Delayed Result Entry * : Not considered for evaluation.</p>				

Problem Classification: _____

Corrective Action: _____

Reviewed by: _____

Dated: _____

LIFE CARE DIAGNOSTIC (556)

Outlier And Analyte Summary Report

Result Details For Cycle No 210207 and Sample No 02


Report Date : 19/07/2021

Instrument : Zybio Z3

Analyte	StandardUnit	Result Value	Mean	Z-Score	RMZ
# Eosinophils Differential.	%				
@ Erythrocyte (RBC) Count.	mill/cu.mm	5.14	--	-0.51	-0.04
@ Haemoglobin.	g/dL	17.60	--	1.37	1.84
X Lymphocytes Differential.	%	5.49	--	-3.23	-1.62
@ MCH (Mean Corpuscular Hb).	pg	34.20	--	1.54	1.39
@ MCHC (Mean Corpuscular Hb Concentration).	g/dL	31.90	--	0.39	0.61
@ MCV (Mean Corpuscular Volume).	fL	107.10	--	0.22	-0.08
X Neutrophils Differential.	%	92.84	--	5.32	4.09
@ Platelet Count.	1000/ μ L	282.00	--	0.67	0.46
@ Total Leucocytes (WBC) count.	/c.mm	6910.00	--	-0.84	-1.13

Legend @ : z score \leq 2.0 - Acceptable
! : 2.0 < z score < 3.0 - Warning
X : z score \geq 3 - Unacceptable
: Not Evaluated
🕒 : Delayed Result Entry
* : Not considered for evaluation.

Total Parameters	10
Not Evaluated Parameters	1
Evaluated Parameters	9
Outlier Parameters (X)	2
EQAS Score Haematology - CBC WB	77.78 %



Dr Puneet Kumar Nigam
PT coordinator & Technical Manager, MHL EQAS
Unit No. 409-416.
Commercial Building - 1A,
Kohinoor Mall, Kiroli Road, Kurla (W),
Mumbai - 400070

End of Report