

**Registered Office:**

**BioSystems Diagnostics Pvt. Ltd.,**  
A4, SIPCOT Industrial Park, Irungattukottai,  
Thandalam Village, Sriperumbudur - 602105.  
Ph: +91 - 44 - 3034 1911 / 12 /13  
E-mail: customersupport@biosystems.in  
Website: www.biosystems.in  
CIN - U29297TN2003PTC051492



Date: 02/05/2021

**CURRENT CALIBRATION CERTIFICATE OF A15**

CUSTOMER NAME – **Suriyaa Diagnostic Centre, Chennai**  
Serial Number = 831055874  
Firmware = A15 User V5.10  
Firmware Integrity Checksum value = 0x5E5A  
Firmware Integrity Checksum size = 439400  
A15 Configuration checksum value = 0xE863  
A15 Configuration checksum size = 986  
Rotor Temperature = 36.99  
Prove Temperature = 39.7  
System Liquid Scale Sensivity = 250  
Waste Scale Sensivity = 250  
Level Detection Improved Rack = 1  
Sensitivity: (Tube Rack) = 196  
Sensitivity: (Pediatric Rack) = 40  
Sensitivity: (Reagent Rack) = 10  
X - Origin = 44  
Y - Origin = 263  
Z - Origin = 430  
X - Washing Station = 372  
Y - Washing Station = 94  
Z - Washing Station = 540  
X - Reactions Rotor = 84  
Y - Reactions Rotor = 1053  
Z - Reactions Rotor = 600  
Tray XYZ- Pediatric Rack  
X1 (Rack1, 2, 3, 4) = 2727, 2177, 1624, 1073  
X2 (Rack1, 2, 3, 4) = 2712, 2162, 1618, 1051  
Y1 (Rack1, 2, 3, 4) = 100, 104, 104, 102  
Y2 (Rack1, 2, 3, 4) = 1457, 1455, 1454, 1457  
Z1 (Rack1, 2, 3, 4) = 722, 722, 722, 722  
Z2 (Rack1, 2, 3, 4) = 695, 695, 695, 695  
Z Relationship Pediatric-Tube = 477  
Tray XYZ- Reagent Rack  
X1 (Rack1, 2, 3, 4) = 2476, 1928, 1371, 818  
X2 (Rack1, 2, 3, 4) = 2463, 1913, 1363, 806  
Y1 (Rack1, 2, 3, 4) = 115, 108, 108, 109  
Y2 (Rack1, 2, 3, 4) = 1444, 1445, 1443, 1441