

Parashar Micro Measurement Pvt. Ltd.



B-59, SECTOR-64, NOIDA (U.P.) 201 307
 PH.: 0120-4252360-65, Email : pmmpl.india@gmail.com
 Website : www.pmmpl.in, www.parashar.info
 (NABL ACCREDITED LABORATORY)



Calibration Certificate

ULR - CC279521000017916F		Page 1 of 1				
Certificate Number : PMM/12789/3		UUC Received on 20.07.2021				
Calibrated For	M/s. Nirvana Diagnostic Centre Medical Mor, PIG Road, Rohtak - 124001	Calibrated on Valid Upto Cert. Issue Date	20.07.2021 19.07.2022 23.07.2021			
Description of Equipments						
Nomenclature	Micro Pipette	Accuracy/Class	Not Mentioned			
Range	2 - 20 µl	Condition of UUC	Physically Ok			
Resolution / Least Count	0.02 µl	Serial No.	LW04048			
Make & Model No.	Agappe	Party ID Mark No.	-----			
Material / Type	-----	Location	Lab			
Master Equipment / Standard Used						
Sr.No.	Nomenclature	Make & Model	Serial No. / I/D Mark	Calibrated From	Certificate No.	Due Date of Calibration
1	Digital Weighing Balance	Sartorius	PMM/WM/06	PMMPL, Noida	PMM/1265	17.05.2022

Standard used for calibration are traceable to National Standards through unbroken chain of calibration.

Calibration Performed at	Laboratory	Environmental	Temperature = 23 ± 2° C
Calibration Procedure No.	CP/MMV/04	Condition(s)	Humidity = 30 to 75 % RH
Reference Standard / Guideline	ISO-8655-6		

Mechanical Calibration

CALIBRATION RESULTS

Sr.No.	Nominal Value in UUC (in µl)	Actual Observed Value Mean Value in Std. (in µl)	Error (in µl)	Mean Average Value at 27°C (in µl)
1	10	10.01	-0.01	10.01
2	15	15.00	0.00	15.00
3	20	19.98	0.02	19.98

UUC:- Unit Under Calibration.

Std. - Standard Instrument

Uncertainty of Measurement (at approx 95% Confidence Level with Coverage factor $k = 2$) = ± 0.17 µl

Calibrated By : Pankaj Yadav
 Designation : Calibration Engineer



Approved By : Jai Prakash
 Designation : Tech. Manager / Auth. Sign.

- Conditions :
1. This certificate refers only to the particular item submitted for calibration.
 2. The calibration result reported in this certificate are valid at the time of and under the stated conditions of measurement.
 3. This particular certificate can not be reproduced except in full, without prior permission of chief executive officer of the lab.