

Parashar Micro Measurement Pvt. Ltd.

B-59, SECTOR-64, NOIDA (U.P.) 201 307
 PH.: 0120-4252360-65, Email : pmmpl.india@gmail.com
 Website : www.pmmpl.in, www.parashar.info
 (NABL ACCREDITED LABORATORY)



Calibration Certificate

ULR - CC279521000017915F			Page 1 of 1			
Certificate Number : PMM/12789/2			Date of Request		20.07.2021	
Calibrated For	M/s. Nirvana Diagnostic Centre		Calibrated on		20.07.2021	
	Medical Mor, PIG Road, Rohtak - 124001		Valid Upto		19.07.2022	
			Cert. Issue Date		23.07.2021	
Description of Equipments						
Nomenclature	Centrifuge	Accuracy/Class	Not Mentioned			
Range	0 - 5250 rpm	Condition of UUC	Physically Ok			
Resolution / Least Count	10 rpm	Serial No.	ZBIN - 24983			
Make & Model No.	Remi	Party ID Mark No.	-----			
Material / Type	-----	Location	Lab			
Master Equipment / Standard Used						
Sr.No.	Nomenclature	Make & Model	Serial No. / I/D Mark	Calibrated From	Certificate No.	Due Date of Calibration
1	Dig. Tachometer	Lutron	PMM/TM/01	PMMPL, Noida	PMM/TM/1201	07.09.2021

Standard used for calibration are traceable to National Standards through unbroken chain of calibration.

Calibration Performed at	Site	Environmental	Temperature = 23 ± 5° C
Calibration Procedure No.	CP/MMV/S/03	Condition(s)	Humidity = 30 to 75 % RH
Reference Standard / Guideline	-----		


Mechanical Calibration

CALIBRATION RESULTS				
Sr.No.	Nominal Value in UUC (in rpm)	Actual Observed Value in Std. (in rpm)	Error (in rpm)	± Uncertainty in Measurement (in rpm)
1	1000	1001	-1	51
2	1400	1405	-5	51
3	3000	3009	-9	51
4	4000	4012	-12	51
5	5000	5015	-15	51

UUC - Unit Under Calibration

Std. - Standard Instrument

Uncertainty of Measurement (at approx 95% Confidence Level with Coverage factor $k = 2$)

Calibrated By : 
 Designation : Calibration Engineer



Approved By : 
 Designation : Tech. Manager / Auth. Sign.

Conditions : 1. This certificate refers only to the particular item submitted for calibration.
 2. The calibration result reported in this certificate are valid at the time of and under the stated conditions of measurement.
 3. This particular certificate can not be reproduced except in full, without prior permission of chief executive officer of the lab.