

TRANSASIA®

Certificate of Calibration

Date-08-04-2021

Customer Name: Family Hospital and Research Center, kanpur up Department of Pathology

Model: EC5+v2(Semi bio chemistry analyzer)

Serial No.: N110517

Calibration done date -.....08-04-2021.....

Next Calibration due Date: On or Before07-04-2022.....

Lab In charge:.....Mr. R. P. OJHA.....

This is to certify that above mentioned product has been verified of calibration according to standard procedures provided by the manufacturer. Transasia Bio Medicals Ltd.

Calibration at site performed by Saurabh Tripathi


(Saurabh Tripathi)
Area Service Manager
Kanpur.



FILTER	GAIN	OFFSET
340	011457	-03548
405	005897	-03373
450	014509	-03512
505	029311	-03483
546	026463	-03257
578	022641	-03414
630	040871	-03483
670	042116	-03319

TRANSASIA BIO-MEDICALS LTD.

www.transasia.co.in



08/04/21 20:00:38
LAB TECH R P OJHA



TRANSASIA BIO-MEDICALS LTD.

TRANSASIA HOUSE, 8 CHANDIVALI STUDIO ROAD, MUMBAI-400 072 TEL : 4030 9000 FAX : (022) 4030 9090 / 2857 3030
Email:transasia@transasia.co.in
DELHI (011) 25732223 CHENNAI (044) 28227149 KOLKATA (033) 22157839 BANGALORE (080) 25568044 AHMEDABAD (079) 26407050