

TRANSASIA®

Certificate of Calibration

Date-08-04-2021

Customer Name: Family Hospital and Research Center, kanpur up Department of Pathology

Model: EC5+v2(Semi bio chemistry analyzer)

Serial No.: N110517

Calibration done date -.....08-04-2021.....

Next Calibration due Date: On or Before07-04-2022.....

Lab In charge:.....Mr. R. P. OJHA.....

This is to certify that above mentioned product has been verified of calibration according to standard procedures provided by the manufacturer. Transasia Bio Medicals Ltd.

Calibration at site performed by Saurabh Tripathi


(Saurabh Tripathi)
Area Service Manager
Kanpur.



FILTER	GAIN	OFFSET
340	011457	-03548
405	005897	-03373
450	014509	-03512
505	029311	-03483
546	026463	-03257
578	022641	-03414
630	040871	-03483
670	042116	-03319

TRANSASIA BIO-MEDICALS LTD.

www.transasia.co.in



08/04/21 20:00:38
LAB TECH R P OJHA



TRANSASIA BIO-MEDICALS LTD.

TRANSASIA HOUSE, 8 CHANDIVALI STUDIO ROAD, MUMBAI-400 072 TEL : 4030 9000 FAX : (022) 4030 9090 / 2857 3030
Email: transasia@transasia.co.in
DELHI (011) 25732223 CHENNAI (044) 28227149 KOLKATA (033) 22157839 BANGALORE (080) 25568044 AHMEDABAD (079) 26407050

2021/7/29 12:55

CR1 DELTA STD U.D. 0.0850
CONC : 2.0 mg/dl
FACTOR: 24.096

001761 CR1 2.31 mg/dl H

29/05/21 19:09:10
LAB TECH R P OJHA

ALB Reagent OD 0.0705
CRITICAL REAGENT ABS.

ALB STD U.D. 0.4755
CONC : 3.6 g/dl
FACTOR: 8.8889

001762 ALB 4.86 g/dl

29/05/21 19:13:13
LAB TECH R P OJHA

PRO Reagent OD 0.1106

PRO STD U.D. 0.2733
CONC : 6 g/dl
FACTOR: 36.878

001763 PRO 7.28 g/dl

29/05/21 19:20:01
LAB TECH R P OJHA

UREA DELTA STD U.D. -0.055
CONC : 50 mg/dl
FACTOR: -909.1

001764 UREA 35.45 mg/dl
RX ABSLIM UNDER

29/05/21 19:33:22
LAB TECH R P OJHA

BIT Reagent OD 0.0000

001765 BIT 1.26 mg/dl H

29/05/2021 CONTROL

29/05/21 18:49:36
LAB TECH R P OJHA

CHO Reagent OD 0.0158

CHO STD O.D. 0.1786
CONC :200 mg/dL
FACTOR:1228.5

001757 CHO 201.6 mg/dL

29/05/21 18:51:41
LAB TECH R P OJHA

TRI Reagent OD 0.0024

TRI STD O.D. 0.0751
CONC :200 mg/dL
FACTOR:2751.0

001758 TRI 159.6 mg/dL

29/05/21 18:54:24
LAB TECH R P OJHA

GLU Reagent OD 0.0626

GLU STD O.D. 0.3522
CONC :100 mg/dL
FACTOR:345.30

001759 GLU 79.94 mg/dL

29/05/21 18:56:49
LAB TECH R P OJHA

UAC Reagent OD 0.0128

UAC STD O.D. 0.0624
CONC :6 mg/dL
FACTOR:120.97

001760 UAC 5.95 mg/dL

29/05/21 19:02:31
LAB TECH R P OJHA