

Parashar Micro Measurement Pvt. Ltd.



The way to Technology
(18 Years of Excellence)

B-59, SECTOR-64, NOIDA (U.P.) 201 307
PH.: 0120-4252360-65, Email : pmmpl.india@gmail.com
Website : www.pmmpl.in, www.parashar.info
(NABL ACCREDITED LABORATORY)



Calibration Certificate

ULR - CC279521000015579F		Page 1 of 1				
Certificate Number : PMM/12654/3		UUC Received on 08.07.2021				
Calibrated For	M/s. Pathology Lab. District Hospital, Mathura	Calibrated on 08.07.2021	Valid Upto 07.07.2022			
		Cert. Issue Date 10.07.2021				
Description of Equipments						
Nomenclature	Micro Pipette	Accuracy/Class	Not Mentioned			
Range	100 - 1000 µl	Condition of UUC	Physically Ok			
Resolution / Least Count	1 µl	Serial No.	ZY63500			
Make & Model No.	Labserve	Party ID Mark No.	-----			
Material / Type	-----	Location	-----			
Master Equipment / Standard Used						
Sr.No.	Nomenclature	Make & Model	Serial No. / I/D Mark	Calibrated From	Certificate No.	Due Date of Calibration
1	Digital Weighing Balance	Sartorius	PMM/WM/06	PMMPL, Noida	PMM/1265	17.05.2022
Standard used for calibration are traceable to National Standards through unbroken chain of calibration.						
Calibration Performed at		Laboratory	Environmental	Temperature = 23 ± 2° C		
Calibration Procedure No.		CP/MMV/04	Condition(s)	Humidity = 30 to 75 % RH		
Reference Standard / Guideline		ISO-8655-6				
Mechanical Calibration						
CALIBRATION RESULTS						
Sr.No.	Nominal Value in UUC (in µl)	Actual Observed Value Mean Value in Std. (in µl)	Error (in µl)	Mean Average Value at 27°C (in µl)		
1	100	100.10	-0.10	100.11		
2	500	499.85	0.15	499.91		
3	1000	1002.72	-2.72	1002.83		

UUC - Unit Under Calibration.

Std. - Standard Instrument

Uncertainty of Measurement (at approx 95% Confidence Level with Coverage factor $k = 2$) = ± 0.65 µl

Calibrated By : Pankaj Yadav
Designation : Calibration Engineer



Approved By : Jai Prakash
Designation : Tech. Manager / Auth. Sign.

Conditions : 1. This certificate refers only to the particular item submitted for calibration.
2. The calibration result reported in this certificate are valid at the time of and under the stated conditions of measurement.
3. This particular certificate can not be reproduced except in full, without prior permission of chief executive officer of the lab.