

Calibration Certificate

Certificate No.:	CCTPL/28-01/40	Field:	Mechanical
SRF No.:	CCTPL/0101/01	Date of Receipt:	03/02/2021
SRF Date:	03/02/2021	Date of Calibration:	04/02/2021
Company Name & Address	M/s. Fatima Hospital	Recommended Due Date.:	03/02/2022
	Mother Taresa Road, Padri Bazar,	As per agreed by customer	
	Gorakhpur. Pin - 273014	Certificate Issue Date:	04/02/2021

DUC* Details

Instrument Name:	Water Bath	Instrument Sr. No.:	----
Make:	----	Instrument Id No.:	----
Model No.:	---	Location:	Lab
Range:	0-300 °C	DUC Condition:	OK
Readability:	1°C	Calibration Performed at:	At Site

Standard Equipments Used(Traceable to National Standard)

Sr. No	Instrument Name	Cert. No	Cal. By	Calibration Due Date
1	RTD (PT 100) with Indicator	CCTPL/R/RTD/01	CCTPL, Noida	17/01/2021

Environment Condition	Temperature	Humidity	Reference Standard	Calibration Procedure No.
	25±5°C	55±10%RH	DKD-R-5-7	CCTPL/TH/CP/05S

Calibration Result

Sr. No.	Nominal Temperature (°C)	Standard Reading (°C)	DUC* Reading (°C)	Deviation (°C)	Expanded Uncertainty ±(°C)
1	10	10.268	10	-0.3	1.0
2	50	50.238	50	-0.4	1.0
3	150	150.433	150	-0.4	1.0
4	250	249.566	249	-0.6	1.0
5	300	300.606	301	0.6	1.0

The above reading are average of 10 measurement reading.

Note : DUC* : Device Under Calibration
 Uncertainty of measurment at approx 95% Confidence Level with coverage factor k = 1.96.



Calibrated By : *Priyank Sharma*
 Priyank Sharma
 Calibration Engineer

Approved By : *Sandeep Sharma*
 Sandeep Sharma
 Director

Calibration Certificate

Certificate No.:	CCTPL/28-01/40	Field:	Mechanical
SRF No.:	CCTPL/0101/01	Date of Receipt:	03/02/2021
SRF Date:	03/02/2021	Date of Calibration:	04/02/2021
Company Name & Address	M/s. Fatima Hospital	Recommended Due Date.:	03/02/2022
	Mother Taresa Road, Padri Bazar,	As per agreed by customer	
	Gorakhpur. Pin - 273014	Certificate Issue Date:	04/02/2021

DUC* Details

Instrument Name:	Water Bath	Instrument Sr. No.:	----
Make:	----	Instrument Id No.:	----
Model No.:	---	Location:	Lab
Range:	0-300 °C	DUC Condition:	OK
Readability:	1°C	Calibration Performed at:	At Site

Standard Equipments Used(Traceable to National Standard)

Sr. No	Instrument Name	Cert. No	Cal. By	Calibration Due Date
1	RTD (PT 100) with Indicator	CCTPL/R/RTD/01	CCTPL, Noida	17/01/2021

Environment Condition	Temperature	Humidity	Reference Standard	Calibration Procedure No.
	25±5°C	55±10%RH	DKD-R-5-7	CCTPL/TH/CP/05S

Calibration Result

Sr. No.	Nominal Temperature (°C)	Standard Reading (°C)	DUC* Reading (°C)	Deviation (°C)	Expanded Uncertainty ±(°C)
1	10	10.268	10	-0.3	1.0
2	50	50.238	50	-0.4	1.0
3	150	150.433	150	-0.4	1.0
4	250	249.566	249	-0.6	1.0
5	300	300.606	301	0.6	1.0

The above reading are average of 10 measurement reading.

Note : DUC* : Device Under Calibration

Uncertainty of measurement at approx 95% Confidence Level with coverage factor k = 1.96.



Calibrated By : *Priyank Sharma*
Priyank Sharma
Calibration Engineer

Approved By : *Sandeep Sharma*
Sandeep Sharma
Director

Note 1.This Certificate only to the particular instrument submitted for calibration.
2.This Calibration Result reported in this certificate valid at the time of and under the stated condition of measurement.
3.This particular certificate can not be reproduce except in full, without prior permission of CCTPL.

--- End of Certificate ---

Page 1 of 1



0120-4105146 | 9810703533 | 9899899577



info@calyss.in

