

## Calibration Certificate

<b>Certificate No.:</b>	CCTPL/05-07/01	<b>Field:</b>	Mechanical
<b>SRF No.:</b>	CCTPL/0107/01	<b>Date of Receipt:</b>	30/07/2021
<b>SRF Date:</b>	30/07/2021	<b>Date of Calibration:</b>	30/07/2021
<b>Company Name &amp; Address</b>	M/s. Fatima Hospital  Mother Taresa Road, Padri Bazar Gorkhpur Pin-273014	<b>Recommended Due Date.:</b> (As per agreed by customer)	<b>29/07/2022</b>
		<b>Certificate Issue Date:</b>	31/07/2021

### DUC\* Details

<b>Instrument Name:</b>	Micro Pipette	<b>Instrument Sr. No.:</b>	----
<b>Make:</b>	-----	<b>Instrument Id No.:</b>	----
<b>Model No.:</b>	-----	<b>Location:</b>	-----
<b>Range:</b>	10 to 100 µl	<b>DUC Condition:</b>	Ok
<b>Readability:</b>	5 µl	<b>Calibration Performed at:</b>	At Lab

### Standard Equipments Used(Traceable to National Standard)

Sr. No	Instrument Name	Cert. No	Cal. By	Calibration Due Date
1	Analytical Balance	PMM/AB/12080/1	PMM, Noida	06/05/2022

Environment Condition	Temperature	Humidity	Reference Standard	Calibration Procedure
	23±2°C	55±10%RH	ISO: 8655-6	CCTPLP/M/MV-02

### CALIBRATION RESULT

Sr. No.	DUC Value (in µl)	Observed Value on Standard (in mg)	Error (in µl)	Mean Value at 27°C (in µl)
1	10	9.84	0.16	9.84
2	50	49.78	0.22	49.78
3	100	99.84	0.15	99.85

### Note : DUC\* : Device Under Calibration

Uncertainty of measurement at approx 95% Confidence Level with coverage factor k = 2, is ± 1.0 µl.

**Calibrated By :** Priyank Sharma  
Calibration Engineer

**Approved By :** Sandeep Sharma  
Director

**Note.** 1.This Certificate only to the particular instrument submitted for calibration.  
2.This calibration Result reported in this certificate valid at the time of and under the stated condition of measurement.  
3.This particular certificate can not be reproduce except in full, without prior permission of CCTPL

~---End of Certificate---

Page 1 of 1

**Calibration Certificate**

Certificate No.:	CCTPL/28-01/08	Field:	Mechanical
SRF No.:	CCTPL/0101/01	Date of Receipt:	03/02/2021
SRF Date:	03/02/2021	Date of Calibration:	04/02/2021
Company Name & Address	M/s. Fatima Hospital Mother Taresa Road, Padri Bazar, Gorakhpur. Pin - 273014	Recommended Due Date.:	03/02/2022
		As per agreed by customer	
		Certificate Issue Date:	04/02/2021

**DUC\* Details**

Instrument Name:	Micro Pipette	Instrument Sr. No.:	---
Make:	----	Instrument Id No.:	----
Model No.:	----	Location:	Lab
Range:	10 - 100 µl	DUC Condition:	Ok
Readability:	5 µl	Calibration Performed at:	At Lab

**Standard Equipments Used(Traceable to National Standard)**

Sr. No	Instrument Name	Cert. No	Cal. By	Calibration Due Date
1	Analytical Balance	PMM/AB/9374/1	PMM, Noida	31/05/2021

Environment Condition	Temperature	Humidity	Reference Standard	Calibration Procedure
	23±2°C	55±10%RH	ISO: 8655-6	CCTPLP/M/MV-02

**CALIBRATION RESULT**


Sr. No.	DUC Value (in µl)	Observed Value on Standard (in mg)	Error (in µl)	Mean Value at 27°C (in µl)
1	10	9.84	0.16	9.84
2	50	24.91	25.09	24.91
3	100	99.84	0.16	99.85

**Note : DUC\* : Device Under Calibration**

Uncertainty of measurement at approx 95% Confidence Level with coverage factor k = 2, is ± 1.1 µl.

Calibrated By :

*P. Sharma*  
Priyank Sharma  
Calibration Engineer



Approved By :

*S. Sharma*  
Sandeep Sharma  
Director

- Note.**
1. This Certificate only to the particular instrument submitted for calibration.
  2. This Calibration Result reported in this certificate valid at the time of and under the stated condition of measurement.
  3. This particular certificate can not be reproduce except in full, without prior permission of CCTPL

—End of Certificate—

Page 1 of 1

