

Calibration Certificate

Certificate No.:	CCTPL/28-01/01	Field:	Mechanical
SRF No.:	CCTPL/0101/01	Date of Receipt:	03/02/2021
SRF Date:	03/02/2021	Date of Calibration:	04/02/2021
Company Name & Address	M/s. Fatima Hospital	Recommended Due Date.:	03/02/2022
	Mother Taresa Road, Padri Bazar, Gorakhpur. Pin - 273014	As per agreed by customer	
		Certificate Issue Date:	04/02/2021

DUC* Details

Instrument Name:	Micro Pipette	Instrument Sr. No.:	----
Make:	----	Instrument Id No.:	----
Model No.:	----	Location:	Lab
Range:	25 µl (Fixed)	DUC Condition:	Ok
Readability:	----	Calibration Performed at:	At Lab

Standard Equipments Used(Traceable to National Standard)

Sr. No	Instrument Name	Cert. No	Cal. By	Calibration Due Date
1	Analytical Balance	PMM/AB/9374/1	PMM, Noida	31/05/2021


Environment Condition	Temperature	Humidity	Reference Standard	Calibration Procedure
	23±2°C	55±10%RH	ISO: 8655-6	CCTPLP/M/MV-02


CALIBRATION RESULT

Sr. No.	DUC Value (in µl)	Observed Value on Standard (in mg)	Error (in µl)	Mean Value at 27°C (in µl)
1	25	24.65	0.35	24.65

Note : DUC* : Device Under Calibration

Uncertainty of measurement at approx 95% Confidence Level with coverage factor k = 2, is ± 1.1 µl.

Calibrated By : 
Priyank Sharma
Calibration Engineer



Approved By : 
Sandeep Sharma
Director

Note.	1.This Certificate only to the particular instrument submitted for calibration. 2.This Calibration Result reported in this certificate valid at the time of and under the stated condition of measurement. 3.This particular certificate can not be reproduce except in full, without prior permission of CCTPL
--------------	---

—End of Certificate—

Page 1 of 1

Calibration Certificate

Certificate No.:	CCTPL/28-01/02	Field:	Mechanical
SRF No.:	CCTPL/0101/01	Date of Receipt:	03/02/2021
SRF Date:	03/02/2021	Date of Calibration:	04/02/2021
Company Name & Address	M/s. Fatima Hospital Mother Taresa Road, Padri Bazar, Gorakhpur. Pin - 273014	Recommended Due Date.:	03/02/2022
		As per agreed by customer	
		Certificate Issue Date:	04/02/2021

DUC* Details

Instrument Name:	Micro Pipette	Instrument Sr. No.:	----
Make:	----	Instrument Id No.:	----
Model No.:	----	Location:	Lab
Range:	50 µl (Fixed)	DUC Condition:	Ok
Readability:	----	Calibration Performed at:	At Lab

Standard Equipments Used(Traceable to National Standard)

Sr. No	Instrument Name	Cert. No	Cal. By	Calibration Due Date
1	Analytical Balance	PMM/AB/9374/1	PMM, Noida	31/05/2021


Environment Condition	Temperature	Humidity	Reference Standard	Calibration Procedure
	23±2°C	55±10%RH	ISO: 8655-6	CCTPLP/M/MV-02

CALIBRATION RESULT

Sr. No.	DUC Value (in µl)	Observed Value on Standard (in mg)	Error (in µl)	Mean Value at 27°C (in µl)
1	50	49.65	0.35	49.65

Note : DUC* : Device Under Calibration

Uncertainty of measurement at approx 95% Confidence Level with coverage factor k = 2, is ± 1.1 µl.



Calibrated By :

Priyank Sharma
Priyank Sharma
Calibration Engineer

Approved By :

Sandeep Sharma
Sandeep Sharma
Director

- Note.**
- 1.This Certificate only to the particular instrument submitted for calibration.
 - 2.This Calibration Result reported in this certificate valid at the time of and under the stated condition of measurement.
 - 3.This particular certificate can not be reproduce except in full, without prior permission of CCTPL

—End of Certificate—

Page 1 of 1



Calibration Certificate

Certificate No.:	CCTPL/28-01/03	Field:	Mechanical
SRF No.:	CCTPL/0101/01	Date of Receipt:	03/02/2021
SRF Date:	03/02/2021	Date of Calibration:	04/02/2021
Company Name & Address	M/s. Fatima Hospital	Recommended Due Date.:	03/02/2022
	Mother Taresa Road, Padri Bazar,	As per agreed by customer	
	Gorakhpur. Pin - 273014	Certificate Issue Date:	04/02/2021

DUC* Details

Instrument Name:	Micro Pipette	Instrument Sr. No.:	----
Make:	----	Instrument Id No.:	----
Model No.:	----	Location:	Lab
Range:	100 µl (Fixed)	DUC Condition:	Ok
Readability:	----	Calibration Performed at:	At Lab

Standard Equipments Used(Traceable to National Standard)

Sr. No	Instrument Name	Cert. No	Cal. By	Calibration Due Date
1	Analytical Balance	PMM/AB/9374/1	PMM, Noida	31/05/2021

Environment Condition	Temperature	Humidity	Reference Standard	Calibration Procedure
		23±2°C	55±10%RH	ISO: 8655-6

CALIBRATION RESULT


Sr. No.	DUC Value (in µl)	Observed Value on Standard (in mg)	Error (in µl)	Mean Value at 27°C (in µl)
1	100	99.75	0.25	99.76

Note : DUC* : Device Under Calibration

Uncertainty of measurement at approx 95% Confidence Level with coverage factor k = 2, is ± 1.1 µl.

Calibrated By :

Priyank Sharma
Priyank Sharma
Calibration Engineer



Approved By :

Sandeep Sharma
Sandeep Sharma
Director

Note. 1.This Certificate only to the particular instrument submitted for calibration.
2.This Calibration Result reported in this certificate valid at the time of and under the stated condition of measurement.
3.This particular certificate can not be reproduce except in full, without prior permission of CCTPL

---End of Certificate---

Page 1 of 1



Calibration Certificate

Certificate No.:	CCTPL/28-01/04	Field:	Mechanical
SRF No.:	CCTPL/0101/01	Date of Receipt:	03/02/2021
SRF Date:	03/02/2021	Date of Calibration:	04/02/2021
Company Name & Address	M/s. Fatima Hospital	Recommended Due Date.:	03/02/2022
	Mother Taresa Road, Padri Bazar, Gorakhpur, Pin - 273014	As per agreed by customer	
		Certificate Issue Date:	04/02/2021

DUC* Details

Instrument Name:	Micro Pipette	Instrument Sr. No.:	----
Make:	----	Instrument Id No.:	----
Model No.:	----	Location:	Lab
Range:	200 µl (Fixed)	DUC Condition:	Ok
Readability:	----	Calibration Performed at:	At Lab

Standard Equipments Used(Traceable to National Standard)

Sr. No	Instrument Name	Cert. No	Cal. By	Calibration Due Date
1	Analytical Balance	PMM/AB/9374/1	PMM, Noida	31/05/2021

Environment Condition	Temperature	Humidity	Reference Standard	Calibration Procedure
	23±2°C	55±10%RH	ISO: 8655-6	CCTPLP/M/MV-02

CALIBRATION RESULT


Sr. No.	DUC Value (in µl)	Observed Value on Standard (in mg)	Error (in µl)	Mean Value at 27°C (in µl)
1	200	199.63	0.37	199.65

Note : DUC* : Device Under Calibration

Uncertainty of measurement at approx 95% Confidence Level with coverage factor k = 2, is ± 1.1 µl.

Calibrated By :

Priyank Sharma
Priyank Sharma
Calibration Engineer



Approved By :

Sandeep Sharma
Sandeep Sharma
Director

Note. 1. This Certificate only to the particular instrument submitted for calibration.
2. This Calibration Result reported in this certificate valid at the time of and under the stated condition of measurement.
3. This particular certificate can not be reproduce except in full, without prior permission of CCTPL

—End of Certificate—

Page 1 of 1

Calibration Certificate

Certificate No.:	CCTPL/28-01/05	Field:	Mechanical
SRF No.:	CCTPL/0101/01	Date of Receipt:	03/02/2021
SRF Date:	03/02/2021	Date of Calibration:	04/02/2021
Company Name & Address	M/s. Fatima Hospital Mother Taresa Road, Padri Bazar, Gorakhpur. Pin - 273014	Recommended Due Date.:	03/02/2022
		As per agreed by customer	
		Certificate Issue Date:	04/02/2021

DUC* Details

Instrument Name:	Micro Pipette	Instrument Sr. No.:	----
Make:	----	Instrument Id No.:	----
Model No.:	----	Location:	Lab
Range:	250 µl (Fixed)	DUC Condition:	Ok
Readability:	----	Calibration Performed at:	At Lab

Standard Equipments Used(Traceable to National Standard)

Sr. No	Instrument Name	Cert. No	Cal. By	Calibration Due Date
1	Analytical Balance	PMM/AB/9374/1	PMM, Noida	31/05/2021

Environment Condition	Temperature	Humidity	Reference Standard	Calibration Procedure
		23±2°C	55±10%RH	ISO: 8655-6

CALIBRATION RESULT

Sr. No.	DUC Value (in µl)	Observed Value on Standard (in mg)	Error (in µl)	Mean Value at 27°C (in µl)
1	250	249.45	0.55	249.47

Note : DUC* : Device Under Calibration

Uncertainty of measurement at approx 95% Confidence Level with coverage factor k = 2, is ± 1.1 µl.

Calibrated By :

P. Sharma
Priyank Sharma
Calibration Engineer



Approved By :

S. Sharma
Sandeep Sharma
Director

- Note.**
- 1.This Certificate only to the particular instrument submitted for calibration.
 - 2.This Calibration Result reported in this certificate valid at the time of and under the stated condition of measurement.
 - 3.This particular certificate can not be reproduce except in full, without prior permission of CCTPL

—End of Certificate—

Page 1 of 1

Calibration Certificate

Certificate No.:	CCTPL/28-01/06	Field:	Mechanical
SRF No.:	CCTPL/0101/01	Date of Receipt:	03/02/2021
SRF Date:	03/02/2021	Date of Calibration:	04/02/2021
Company Name & Address	M/s. Fatima Hospital Mother Taresa Road, Padri Bazar, Gorakhpur. Pin - 273014	Recommended Due Date.:	03/02/2022
		As per agreed by customer	
		Certificate Issue Date:	04/02/2021

DUC* Details

Instrument Name:	Micro Pipette	Instrument Sr. No.:	----
Make:	----	Instrument Id No.:	----
Model No.:	----	Location:	Lab
Range:	5 - 50 µl	DUC Condition:	Ok
Readability:	0.1 µl	Calibration Performed at:	At Lab

Standard Equipments Used(Traceable to National Standard)

Sr. No	Instrument Name	Cert. No	Cal. By	Calibration Due Date
1	Analytical Balance	PMM/AB/9374/1	PMM, Noida	31/05/2021

Environment Condition	Temperature	Humidity	Reference Standard	Calibration Procedure
	23±2°C	55±10%RH	ISO: 8655-6	CCTPLP/M/MV-02

CALIBRATION RESULT

Sr. No.	DUC Value (in µl)	Observed Value on Standard (in mg)	Error (in µl)	Mean Value at 27°C (in µl)
1	10	9.85	0.15	9.85
2	25	24.78	0.22	24.78
3	50	49.68	0.32	49.68

Note : DUC* : Device Under Calibration

Uncertainty of measurement at approx 95% Confidence Level with coverage factor k = 2, is ± 1.1 µl.

Calibrated By :

Priyank Sharma
Priyank Sharma
Calibration Engineer



Approved By :

Sandeep Sharma
Sandeep Sharma
Director

- Note.**
- 1.This Certificate only to the particular instrument submitted for calibration.
 - 2.This Caliibration Result reported in this certificate valid at the time of and under the stated condition of measurement.
 - 3.This particular certificate can not be reproduce except in full, without prior permission of CCTPL

—End of Certificate—

Page 1 of 1



Calibration Certificate

Certificate No.:	CCTPL/28-01/07	Field:	Mechanical
SRF No.:	CCTPL/0101/01	Date of Receipt:	03/02/2021
SRF Date:	03/02/2021	Date of Calibration:	04/02/2021
Company Name & Address	M/s. Fatima Hospital Mother Taresa Road, Padri Bazar, Gorakhpur. Pin - 273014	Recommended Due Date.:	03/02/2022
		As per agreed by customer	
		Certificate Issue Date:	04/02/2021

DUC* Details

Instrument Name:	Micro Pipette	Instrument Sr. No.:	----
Make:	----	Instrument Id No.:	----
Model No.:	----	Location:	Lab
Range:	5 - 50 µl	DUC Condition:	Ok
Readability:	0.1 µl	Calibration Performed at:	At Lab

Standard Equipments Used(Traceable to National Standard)

Sr. No	Instrument Name	Cert. No	Cal. By	Calibration Due Date
1	Analytical Balance	PMM/AB/9374/1	PMM, Noida	31/05/2021

Environment Condition	Temperature	Humidity	Reference Standard	Calibration Procedure
	23±2°C	55±10%RH	ISO: 8655-6	CCTPLP/M/MV-02

CALIBRATION RESULT

Sr. No.	DUC Value (in µl)	Observed Value on Standard (in mg)	Error (in µl)	Mean Value at 27°C (in µl)
1	10	9.85	0.15	9.85
2	25	24.78	0.22	24.78
3	50	49.68	0.32	49.68

Note : DUC* : Device Under Calibration

Uncertainty of measurement at approx 95% Confidence Level with coverage factor k = 2, is ± 1.1 µl.

Calibrated By :

P. Sharma
Priyank Sharma
Calibration Engineer



Approved By :

S. Sharma
Sandeep Sharma
Director

Note. 1.This Certificate only to the particular instrument submitted for calibration.
2.This Calibration Result reported in this certificate valid at the time of and under the stated condition of measurement.
3.This particular certificate can not be reproduce except in full, without prior permission of CCTPL

—End of Certificate—

Page 1 of 1

Calibration Certificate

Certificate No.:	CCTPL/28-01/08	Field:	Mechanical
SRF No.:	CCTPL/0101/01	Date of Receipt:	03/02/2021
SRF Date:	03/02/2021	Date of Calibration:	04/02/2021
Company Name & Address	M/s. Fatima Hospital	Recommended Due Date.:	03/02/2022
	Mother Taresa Road, Padri Bazar,	As per agreed by customer	
	Gorakhpur. Pin - 273014	Certificate Issue Date:	04/02/2021

DUC* Details

Instrument Name:	Micro Pipette	Instrument Sr. No.:	----
Make:	----	Instrument Id No.:	----
Model No.:	----	Location:	Lab
Range:	10 - 100 µl	DUC Condition:	Ok
Readability:	5 µl	Calibration Performed at:	At Lab

Standard Equipments Used(Traceable to National Standard)

Sr. No	Instrument Name	Cert. No	Cal. By	Calibration Due Date
1	Analytical Balance	PMM/AB/9374/1	PMM, Noida	31/05/2021

Environment Condition	Temperature	Humidity	Reference Standard	Calibration Procedure
	23±2°C	55±10%RH	ISO: 8655-6	CCTPLP/M/MV-02

CALIBRATION RESULT


Sr. No.	DUC Value (in µl)	Observed Value on Standard (in mg)	Error (in µl)	Mean Value at 27°C (in µl)
1	10	9.84	0.16	9.84
2	50	24.91	25.09	24.91
3	100	99.84	0.16	99.85

Note : DUC* : Device Under Calibration

Uncertainty of measurement at approx 95% Confidence Level with coverage factor k = 2, is ± 1.1 µl.

Calibrated By :

Priyank Sharma
Priyank Sharma
Calibration Engineer



Approved By :

Sandeep Sharma
Sandeep Sharma
Director

Note. 1.This Certificate only to the particular instrument submitted for calibration.
2.This Caliibration Result reported in this certificate valid at the time of and under the stated condition of measurement.
3.This particular certificate can not be reproduce except in full, without prior permission of CCTPL

—End of Certificate—

Page 1 of 1



Calibration Certificate

Certificate No.:	CCTPL/28-01/09	Field:	Mechanical
SRF No.:	CCTPL/0101/01	Date of Receipt:	03/02/2021
SRF Date:	03/02/2021	Date of Calibration:	04/02/2021
Company Name & Address	M/s. Fatima Hospital Mother Taresa Road, Padri Bazar, Gorakhpur. Pin - 273014	Recommended Due Date.:	03/02/2022
		As per agreed by customer	
		Certificate Issue Date:	04/02/2021

DUC* Details

Instrument Name:	Micro Pipette	Instrument Sr. No.:	----
Make:	----	Instrument Id No.:	----
Model No.:	----	Location:	Lab
Range:	10 - 100 µl	DUC Condition:	Ok
Readability:	5 µl	Calibration Performed at:	At Lab

Standard Equipments Used(Traceable to National Standard)

Sr. No	Instrument Name	Cert. No	Cal. By	Calibration Due Date
1	Analytical Balance	PMM/AB/9374/1	PMM, Noida	31/05/2021

Environment Condition	Temperature	Humidity	Reference Standard	Calibration Procedure
	23±2°C	55±10%RH	ISO: 8655-6	CCTPLP/M/MV-02

CALIBRATION RESULT


Sr. No.	DUC Value (in µl)	Observed Value on Standard (in mg)	Error (in µl)	Mean Value at 27°C (in µl)
1	100	99.62	0.38	99.63
2	500	499.13	0.87	499.17
3	1000	998.73	1.27	998.81

Note : DUC* : Device Under Calibration

Uncertainty of measurement at approx 95% Confidence Level with coverage factor k = 2, is ± 1.1 µl.

Calibrated By :

Priyank Sharma
Priyank Sharma
Calibration Engineer



Approved By :

Sandeep Sharma
Sandeep Sharma
Director

- Note.**
- 1.This Certificate only to the particular instrument submitted for calibration.
 - 2.This Calibration Result reported in this certificate valid at the time of and under the stated condition of measurement.
 - 3.This particular certificate can not be reproduce except in full, without prior permission of CCTPL

—End of Certificate—

Page 1 of 1



Calibration Certificate

Certificate No.:	CCTPL/28-01/24	Field:	Mechanical
SRF No.:	CCTPL/0101/01	Date of Receipt:	03/02/2021
SRF Date:	03/02/2021	Date of Calibration:	04/02/2021
Company Name & Address	M/s. Fatima Hospital	Recommended Due Date.:	03/02/2022
	Mother Taresa Road, Padri Bazar,	As per agreed by customer	
	Gorakhpur. Pin - 273014	Certificate Issue Date:	04/02/2021

DUC* Details

Instrument Name:	Biosystem	Instrument Sr. No.:	-----
Make:	----	Instrument Id No.:	----
Model No.:	BTS - 350	Location:	Lab
Range:	37 °C	DUC Condition:	OK
Readability:	0.1 °C	Calibration Performed at:	At Site

Standard Equipments Used(Traceable to National Standard)

Sr. No	Instrument Name	Cert. No	Cal. By	Calibration Due Date
1	RTD with Indicator	CCTPL/R/RTD/01	CCTPL, Noida	17/01/2021


Environment Condition	Temperature	Humidity	Reference Standard	Calibration Procedure No.
	25±5°C	55±10%RH	DKD-R-5-7	CCTPL/TH/CP/05S


Calibration Result

Sr. No.	Nominal Temperature (°C)	Standard Reading (°C)	DUC* Reading (°C)	Deviation (°C)	Expanded Uncertainty ±(°C)
1	37	36.231	37	0.769	1.3

Note : DUC* : Device Under Calibration

Uncertainty of measurement at approx 95% Confidence Level with coverage factor k = 2.

Calibrated By : 
Priyank Sharma
Calibration Engineer



Approved By : 
Sandeep Sharma
Director

Note. 1.This Certificate only to the particular instrument submitted for calibration.
2.This Calibration Result reported in this certificate valid at the time of and under the stated condition of measurement.
3.This particular certificate can not be reproduce except in full, without prior permission of CCTPL.

--- End of Certificate ---

Page 1 of 1

Calibration Certificate

Certificate No.:	CCTPL/28-01/25	Field:	Mechanical
SRF No.:	CCTPL/0101/01	Date of Receipt:	03/02/2021
SRF Date:	03/02/2021	Date of Calibration:	04/02/2021
Company Name & Address	M/s. Fatima Hospital	Recommended Due Date.:	03/02/2022
	Mother Taresa Road, Padri Bazar, Gorakhpur. Pin - 273014	As per agreed by customer	
		Certificate Issue Date:	04/02/2021

DUC* Details

Instrument Name:	DTC With Sensor (Incubator)	Instrument Sr. No.:	----
Make:	Ambassador	Instrument Id No.:	----
Model No.:	----	Location:	Lab
Range:	0 - 99.9 °C	DUC Condition:	OK
Readability:	0.1 °C	Calibration Performed at:	At Site

Standard Equipments Used(Traceable to National Standard)

Sr. No	Instrument Name	Cert. No	Cal. By	Calibration Due Date
1	RTD with Indicator	CCTPL/R/RTD/01	CCTPL, Noida	17/01/2021


Environment Condition	Temperature	Humidity	Reference Standard	Calibration Procedure No.
	25±5°C	55±10%RH	DKD-R-5-7	CCTPL/TH/CP/05S


Calibration Result


Sr. No.	Nominal Temperature (°C)	Standard Reading (°C)	DUC* Reading (°C)	Deviation (°C)	Expanded Uncertainty ±(°C)
1	37	37.125	37	-0.125	1.3

Note : DUC* : Device Under Calibration

Uncertainty of measurement at approx 95% Confidence Level with coverage factor k = 2.



Calibrated By : 
Priyank Sharma
Calibration Engineer

Approved By : 
Sandeep Sharma
Director

Note. 1.This Certificate only to the particular instrument submitted for calibration.
2.This Calibration Result reported in this certificate valid at the time of and under the stated condition of measurement.
3.This particular certificate can not be reproduce except in full, without prior permission of CCTPL.

--- End of Certificate ---

Page 1 of 1

Calibration Certificate

Certificate No.:	CCTPL/28-01/28	Field:	Mechanical
SRF No.:	CCTPL/0101/01	Date of Receipt:	03/02/2021
SRF Date:	03/02/2021	Date of Calibration:	04/02/2021
Company Name & Address	M/s. Fatima Hospital	Recommended Due Date.:	03/02/2022
	Mother Taresa Road, Padri Bazar,	As per agreed by customer	
	Gorakhpur. Pin - 273014	Certificate Issue Date:	04/02/2021

DUC* Details

Instrument Name:	Water Bath	Instrument Sr. No.:	---
Make:	---	Instrument Id No.:	---
Model No.:	---	Location:	Lab
Range:	0-300 °C	DUC Condition:	OK
Readability:	1°C	Calibration Performed at:	At Site

Standard Equipments Used(Traceable to National Standard)

Sr. No	Instrument Name	Cert. No	Cal. By	Calibration Due Date
1	RTD (PT 100) with Indicator	CCTPL/R/RTD/01	CCTPL, Noida	17/01/2021

Environment Condition	Temperature	Humidity	Reference Standard	Calibration Procedure No.
	25±5°C	55±10%RH	DKD-R-5-7	CCTPL/TH/CP/05S


Calibration Result

Sr. No.	Nominal Temperature (°C)	Standard Reading (°C)	DUC* Reading (°C)	Deviation (°C)	Expanded Uncertainty ±(°C)
1	10	10.268	10	-0.3	1.0
2	50	50.238	50	-0.4	1.0
3	150	150.433	150	-0.4	1.0
4	250	249.566	249	-0.6	1.0
5	300	300.606	301	0.6	1.0

The above reading are average of 10 measurement reading.

Note : DUC* : Device Under Calibration

Uncertainty of measurement at approx 95% Confidence Level with coverage factor k = 1.96.



Calibrated By :

Priyank Sharma
Priyank Sharma
Calibration Engineer

Approved By :

Sandeep Sharma
Sandeep Sharma
Director

- Note**
- 1.This Certificate only to the particular instrument submitted for calibration.
 - 2.This Calibration Result reported in this certificate valid at the time of and under the stated condition of measurement.
 - 3.This particular certificate can not be reproduce except in full, without prior permission of CCTPL.

--- End of Certificate ---

Page 1 of 1



Calibration Certificate

Certificate No.:	CCTPL/28-01/40	Field:	Mechanical
SRF No.:	CCTPL/0101/01	Date of Receipt:	03/02/2021
SRF Date:	03/02/2021	Date of Calibration:	04/02/2021
Company Name & Address	M/s. Fatima Hospital	Recommended Due Date.:	03/02/2022
	Mother Taresa Road, Padri Bazar,	As per agreed by customer	
	Gorakhpur. Pin - 273014	Certificate Issue Date:	04/02/2021

DUC* Details

Instrument Name:	Water Bath	Instrument Sr. No.:	---
Make:	---	Instrument Id No.:	---
Model No.:	---	Location:	Lab
Range:	0-300 °C	DUC Condition:	OK
Readability:	1°C	Calibration Performed at:	At Site

Standard Equipments Used(Traceable to National Standard)

Sr. No	Instrument Name	Cert. No	Cal. By	Calibration Due Date
1	RTD (PT 100) with Indicator	CCTPL/R/RTD/01	CCTPL, Noida	17/01/2021

Environment Condition	Temperature	Humidity	Reference Standard	Calibration Procedure No.
	25±5°C	55±10%RH	DKD-R-5-7	CCTPL/TH/CP/05S


Calibration Result

Sr. No.	Nominal Temperature (°C)	Standard Reading (°C)	DUC* Reading (°C)	Deviation (°C)	Expanded Uncertainty ±(°C)
1	10	10.268	10	-0.3	1.0
2	50	50.238	50	-0.4	1.0
3	150	150.433	150	-0.4	1.0
4	250	249.566	249	-0.6	1.0
5	300	300.606	301	0.6	1.0

The above reading are average of 10 measurement reading.

Note : DUC* : Device Under Calibration

Uncertainty of measurement at approx 95% Confidence Level with coverage factor k = 1.96.



Calibrated By :

Priyank Sharma
Priyank Sharma
Calibration Engineer

Approved By :

Sandeep Sharma
Sandeep Sharma
Director

- Note**
- 1.This Certificate only to the particular instrument submitted for calibration.
 - 2.This Calibration Result reported in this certificate valid at the time of and under the stated condition of measurement.
 - 3.This particular certificate can not be reproduce except in full, without prior permission of CCTPL.

--- End of Certificate ---

Page 1 of 1

Calibration Certificate

Certificate No.:	CCTPL/28-01/26	Field:	Mechanical
SRF No.:	CCTPL/0101/01	Date of Receipt:	03/02/2021
SRF Date:	03/02/2021	Date of Calibration:	04/02/2021
Company Name & Address	M/s. Fatima Hospital	Recommended Due Date.:	03/02/2022
	Mother Taresa Road, Padri Bazar,	As per agreed by customer	
	Gorakhpur. Pin - 273014	Certificate Issue Date:	04/02/2021

DUC* Details

Instrument Name:	Pressure Gauge (Autoclave)	Instrument Sr. No.:	----
Make:	Electron	Instrument Id No.:	----
Model No.:	----	Location:	Lab
Range:	0-150 lb/ln2	DUC Condition:	Ok
Readability	5 lb/ln2	Calibration Performed at:	At Site

Standard Equipments Used(Traceable to National Standard)

Sr. No	Instrument Name	Cert. No	Cal. By	Calibration Due Date
1	Digital Pressure Gauge	CC20552000000825F	Testncal	01/12/2021


Environment Condition	Temperature	Humidity	Reference Standard	Calibration Procedure No.
	23±5°C	55±10%RH	Comparison Method	CCTPL/ME/CP/31


Calibration Result


Sr. No.	DUC Reading (lb/ln ²)	Std. Reading				Std. Reading (lb/ln ²)	Error (lb/ln ²)
		Ascending (M1) (bar)	Decending (M2) (bar)	Ascending (M3) (bar)	Average ((M1+M3)/2+M2)/2 (bar)		
1	0	0.000	0.000	0.000	0.000	0.000	0.000
2	50	3.503	3.504	3.509	3.505	50.808	-0.808
3	100	7.023	7.019	7.029	7.023	101.849	-1.849
4	150	10.505	10.508	10.509	10.508	152.349	-2.349

Note : DUC* : Device Under Calibration

Uncertainty of measurment at approx 95% Confidence Level with coverage factor k = 2, is ± 2.1 bar.

Calibrated By : 
Priyank Sharma
Calibration Engineer



Approved By : 
Sandeep Sharma
Director

- Note.** 1.This Certificate only to the particular instrument submitted for calibration.
2.This Calibration Result reported in this certificate valid at the time of and under the stated condition of measurement.
3.This particular certificate can not be reproduce except in full, without prior permission of CCTPL.

--- End of Certificate ---

Page 1 of 1

Calibration Certificate

Certificate No.:	CCTPL/28-01/27	Field:	Mechanical
SRF No.:	CCTPL/0101/01	Date of Receipt:	03/02/2021
SRF Date:	03/02/2021	Date of Calibration:	04/02/2021
Company Name & Address	M/s. Fatima Hospital	Recommended Due Date.:	03/02/2022
	Mother Taresa Road, Padri Bazar,	As per agreed by customer	
	Gorakhpur. Pin - 273014	Certificate Issue Date:	04/02/2021

DUC* Details

Instrument Name:	Pressure Gauge (Autoclave)	Instrument Sr. No.:	----
Make:	Electron	Instrument Id No.:	----
Model No.:	----	Location:	Lab
Range:	0-150 lb/in ²	DUC Condition:	Ok
Readability	5 lb/in ²	Calibration Performed at:	At Site

Standard Equipments Used(Traceable to National Standard)

Sr. No	Instrument Name	Cert. No	Cal. By	Calibration Due Date
1	Digital Pressure Gauge	CC20552000000825F	Testncal	01/12/2021

Environment Condition	Temperature	Humidity	Reference Standard	Calibration Procedure No.
	23±5°C	55±10%RH	Comparison Method	CCTPL/ME/CP/31

Calibration Result


Sr. No.	DUC Reading (lb/in ²)	Std. Reading				Average ((M1+M3)/2+M2)/2 (bar)	Std. Reading (lb/in ²)	Error (lb/in ²)
		Ascending (M1) (bar)	Decending (M2) (bar)	Ascending (M3) (bar)				
1	0	0.000	0.000	0.000	0.000	0.000	0.000	
2	50	3.501	3.500	3.503	3.501	50.808	-0.808	
3	100	7.021	7.020	7.024	7.021	101.849	-1.849	
4	150	10.508	10.506	10.509	10.507	152.349	-2.349	

Note : DUC* : Device Under Calibration

Uncertainty of measurment at approx 95% Confidence Level with coverage factor k = 2, is ± 2.1 bar.

Calibrated By :

Priyank Sharma
Priyank Sharma
Calibration Engineer



Approved By :

Sandeep Sharma
Sandeep Sharma
Director

- Note. 1. This Certificate only to the particular instrument submitted for calibration.
2. This Calibration Result reported in this certificate valid at the time of and under the stated condition of measurement.
3. This particular certificate can not be reproduce except in full, without prior permission of CCTPL.

--- End of Certificate ---

Page 1 of 1

REMI SALES & ENGINEERING LIMITED**CALIBRATION**

CUSTOMER / BRANCH: M/s. M/s. Blood Bank, Fatima Hospital, Gorakhpur

DISCRIPTION OF MODEL: LAB CENTRIFUGE

MODEL: NEYA 4

SR.NO : ZEBN-07662

CERTIFICATE NO: REMI/0316/2019-20

CALIBRATION DATE: 08/02/2020

CALIBRATION VALIDITY: 07/08/2020


DETAILS OF EQUIPMENT USED FOR CALIBRATION

Description	Certified by	Certificate No.	Calibration Date	Validity
Digital Tachometer	Classic Instrumentation Pvt. Ltd.	CI/2K19-2K20/05/ME006	02/05/2019	01/05/2020

OBSERVATIONS

Sr. No.	Instrument speed SET (RPM)	Speed reading on Tachometer (RPM)	Diff. in Actual & Tachometer (RPM)	Calibration Date	Remark
1.	2500(RPM)	2517(RPM)	17(RPM)	08/02/2020	O.K
2.	3000(RPM)	3019(RPM)	19(RPM)	08/02/2020	O.K

- Remark:**
1. Calibration Traceable to national standards.
 2. Overall Deviation is within specified accuracy.
 3. Results are related only to the item calibrated.
 4. Maximum tolerance permissible is $\pm 1^{\circ}\text{C}$ & $\pm 5^{\circ}\text{C}$ whichever is applicable & RPM 1.5% - 2% FSD. & $\pm 5\text{g.}$ to set weight.



Calibrated by: Mr. Vineet Kumar

SALES & SERVICE BRANCHES :

Ahmedabad, Bengaluru, Kolkata, Chennai, Hyderabad, Nagpur, New Delhi, Kochi, Kanpur, Indore, Goa


CALIBRATION**CUSTOMER / BRANCH:** M/s. Blood Bank, Fatima Hospital, Gorakhpur**DISCRIPTION OF MODEL:** LAB REFRIGERATOR**MODEL:** SEROCOOL 400 **SR.NO:** ZFGC-23704 **CERTIFICATE NO:** REMI/ 0313/2019-20**CALIBRATION DATE:** 02/01/2020**CALIBRATION VALIDITY:** 01/06/2020**DETAILS OF EQUIPMENT USED FOR CALIBRATION**

Description	Certified by	Certificate No.	Calibration Date	Validity
Digital Temperature Indicator	Classic Instrumentation Pvt. Ltd.	CI/2K19-2K20/05/TH005	02/05/2019	01/05/2020

OBSERVATIONS

Sr. No.	Instrument Temp. Displayed (°C)	Indicator Temp. (°C)	Temperature Deviation	Calibration Date	Remark
1.	4°C	4.2°C	0.2°C	02/01/2020	O.K
2.					

- Remark:**
1. Calibration Traceable to national standards.
 2. Overall Deviation is within specified accuracy.
 3. Results are related only to the item calibrated.
 4. Maximum tolerance permissible is $\pm 1^\circ\text{C}$ & $\pm 5^\circ\text{C}$ whichever is applicable & RPM 1.5% - 2% FSD. & $\pm 5\text{g.}$ to set weight.



Calibrated by: Mr. Vineet Kumar

SALES & SERVICE BRANCHES :

Ahmedabad, Bengaluru, Kolkata, Chennai, Hyderabad, Nagpur, New Delhi, Kochi, Kanpur, Indore, Goa



Date: 01/07/2019

Certificate of Calibration

Customer Name: FATIMA HOSPITAL

Model: Fully Automated Bio-Chemistry Analyzer EM-200

Serial No.: B2001646

Calibration done date: 01/07/2019

Next Calibration due Date: 01/07/2020

Lab In charge: Dr. Aditi Kumar

*This is to certify that above mentioned product has been verified of calibration
By
Checking Lamp Voltage 11.9V*

Checking gain of Photo meter for all wavelength values obtained are as under:


Filter	340	405	505	546	578	600	660	700
Gain	435	482	485	494	415	426	476	467

Checked calibration of probe on all locations.

*Checked temperature at Reagent tray at 8.0 degree (Range 4 to 11 degree) and
reaction tray 37.2 degree (Range 36.5 to 37.5 Degree)*

*Followed by calibration of all the three types of chemistries (picked up
CHOL, CREAT and SGPT) and QC run with satisfactory values of QC*

Calibration at site performed by Mr. Shivarchan Pandey.


Authorized Signatory
(Mr. Shivarchan Pandey)
Application Specialist
For Transasia Bio-Medicals Ltd.



Date: 01/07/2019

Certificate of Calibration

Customer Name: FATIMA HOSPITAL

Model: Fully Automated Bio-Chemistry Analyzer EM-200

Serial No.: B2001646

Calibration done date: 01/07/2019

Next Calibration due Date: 01/07/2020

Lab In charge: Dr. Aditi Kumar

This is to certify that above mentioned product has been verified of calibration

By

Checking Lamp Voltage 11.9V

Checking gain of Photo meter for all wavelength values obtained are as under:

Filter	340	405	505	546	578	600	660	700
Gain	435	482	485	494	415	426	476	467

Checked calibration of probe on all locations.

Checked temperature at Reagent tray at 8.0 degree (Range 4 to 11 degree) and reaction tray 37.2 degree (Range 36.5 to 37.5 Degree)

Followed by calibration of all the three types of chemistries (picked up CHOL, CREAT and SGPT) and QC run with satisfactory values of QC

Calibration at site performed by Mr. Shivarchan Pandey.

Signature
Authorized Signatory,
(Mr. Shivarchan Pandey)
Application Specialist
For Transasia Bio-Medicals Ltd.

Parashar Micro Measurement Pvt. Ltd.

2018



B-59, SECTOR-64, NOIDA (U.P.) 201 307
 PH.: 0120-4252360-65, Email : pmmpl.india@gmail.com
 Website : www.pmmpl.in, www.parashar.info
 (NABL ACCREDITED LABORATORY)



— Calibration Certificate —

Page 1 of 1

Certificate Number : CAL/PG/S/466356		Date of Request	15.12.2018
Calibrated For	M/s. Fatima Hospital	Calibrated on	19.12.2018
	Mother Teresa Road, Padri Bazar PO, Gorakhpur, Uttar Pradesh - 273014	Valid Upto	18.12.2019
		Cert. Issue Date	21.12.2018

Description of Sample

Nomenclature	Pressure Gauge	Accuracy/Class	Not Mentioned
Range	0 - 2 kg/cm ²	Condition of UUC	Physically Ok
Resolution	0.05 kg/cm ²	Calibration Performed at	At Site
Make & Model No.	-----	Calibration Procedure No.	CP/M/P/S/01
Serial No.	-----	Reference IS Standard	IS: DKD - R6 - 1
Party ID Mark	-----	Environmental Condition	Temperature = 23 ± 5° C Humidity = 40 to 70%RH
Location	Laboratory		

Master Equipment / Standard Used

Sr.No.	Nomenclature	Make	Serial No. / I/D Mark	Calibrated From	Certificate No.	Due Date of Calibration
1	Dig. Pressure Gauge	VE	PMM/PG/07	PMMPL, Noida	PMM/PG/1082	09.10.2019

Standard used for calibration are traceable to National Standards through unbroken chain of calibration.

CALIBRATION RESULTS

Sr.No.	Nominal Value in UUC (in kg/cm ²)	Actual Observed Value in Std.		Error (in kg/cm ²)
		(in bar)	(in kg/cm ²)	
1	0.0	0.000	0.000	0.000
2	0.5	0.502	0.512	-0.012
3	1.0	0.998	1.018	-0.018
4	1.5	1.497	1.527	-0.027
5	2.0	1.997	2.036	-0.036
6	1.5	1.491	1.520	-0.020
7	1.0	0.994	1.014	-0.014
8	0.5	0.499	0.509	-0.009
9	0.0	0.000	0.000	0.000


UUC - Unit Under Calibration

Std. - Standard Instrument

Uncertainty of Measurement (at approx 95% Confidence Level with Coverage factor $k = 2$) = ± 0.15 kg/cm²

Calibrated By : 
 Designation : Calibration Engineer



Approved By : 
 Designation : Tech Manager / Auth. Sign.

- Conditions : 1. This certificate refers only to the particular item submitted for calibration.
 2. The calibration result reported in this certificate are valid at the time of and under the stated conditions of measurement.
 3. This particular certificate can not be reproduced except in full, without prior permission of chief executive officer of the lab.

Parashar Micro Measurement Pvt. Ltd.



The way to Technology
(15 Years of Excellence)

B-59, SECTOR-64, NOIDA (U.P.) 201 307
PH.: 0120-4252360-65, Email : pmmpl.india@gmail.com
Website : www.pmmpl.in, www.parashar.info
(NABL ACCREDITED LABORATORY)



Calibration Certificate

Page 1 of 1

Certificate Number : CAL/WB/S/466355		Date of Request	15.12.2018
Calibrated For	M/s. Fatima Hospital	Calibrated on	19.12.2018
	Mother Teresa Road, Padri Bazar PO,	Valid Upto	18.12.2019
	Gorakhpur, Uttar Pradesh - 273014	Cert. Issue Date	21.12.2018

Description of Sample

Nomenclature	Water Bath (Thermostat)	Accuracy/Class	Not Mentioned
Range	0 to 110 ° C	Condition of UUC	Physically Ok
Resolution	5° C	Calibration Performed at	At Site
Make & Model No.	Ambassador	Calibration Procedure No.	CP/T - 03
Serial No.	-----	Reference IS Standard	IS:7358 - 1984
Party ID Mark	-----	Environmental Condition	Temperature = 25 ± 5° C
Location	Laboratory		Humidity = 40 to 70% RH

Master Equipment / Standard Used

Sr.N.	Nomenclature	Make	Serial No. / I/D Mark	Calibrated From	Certificate No.	Due Date of Calibration
1	RTD With Dig. Indicator	Tempsens Fluke	PMM/RTD/03 PMM/CTR/01	Tempsens Udaipur	TL/218/109.1.1	06.05.2019

Standard used for calibration are traceable to National Standards through unbroken chain of calibration.

CALIBRATION RESULTS


S.N.	Nominal Value in UUC (in ° C)	Actual Observed Value in Std. (in ° C)	Error (in ° C)
1	35	34.62	0.38
2	40	39.48	0.52
3	50	49.32	0.68
4	60	59.10	0.90
5	80	78.86	1.14

UUC - Unit Under Calibration

Std. - Standard Instrument

Uncertainty of Measurement (at approx 95% Confidence Level with Coverage factor $k = 2$) = ± 3.3 ° C

Calibrated By : Rajesh Gupta
Designation : Calibration Engineer



Approved By : Moh. Iqbal
Designation : Tech Manager / Auth. Sign.

- Conditions :
1. This certificate refers only to the particular item submitted for calibration.
 2. The calibration result reported in this certificate are valid at the time of and under the stated conditions of measurement.
 3. This particular certificate can not be reproduced except in full, without prior permission of chief executive officer of the lab.

Parashar Micro Measurement Pvt. Ltd.

2014



B-59, SECTOR-64, NOIDA (U.P.) 201 307
 PH.: 0120-4252360-65, Email : pmmpl.india@gmail.com
 Website : www.pmmpl.in, www.parashar.info
 (NABL ACCREDITED LABORATORY)



Calibration Certificate

Page 1 of 1

Certificate Number : CAL/CF/S/466353		Date of Request	15.12.2018
Calibrated For	M/s. Fatima Hospital	Calibrated on	19.12.2018
	Mother Teresa Road, Padri Bazar PO,	Valid Upto	18.12.2019
	Gorakhpur, Uttar Pradesh - 273014	Cert. Issue Date	21.12.2018

Description of Sample

Nomenclature	Centrifuge	Accuracy/Class	Not Mentioned
Range	0 - 6000 rpm	Condition of UUC	Physically Ok
Resolution	10 rpm	Calibration Performed at	At Site
Make & Model No.	Remi	Calibration Procedure No.	CP/MMV/S/03
Serial No.	ZEBN - 07662	Reference IS Standard	-----
Party ID Mark	-----	Environmental Condition	Temperature = 23 ± 5° C
Location	Laboratory		Humidity = 40 to 70%RH

Master Equipment / Standard Used

Sr.No.	Nomenclature	Make	Serial No. / I/D Mark	Calibrated From	Certificate No.	Due Date of Calibration
1	Dig. Tachometer	Lutron	PMM/TM/01	PMPL	PMM/TM/1075	11.09.2019

Standard used for calibration are traceable to National Standards through unbroken chain of calibration.

CALIBRATION RESULTS

Sr.No.	Nominal Value in UUC (in rpm)	Actual Observed Value in Std. (in rpm)	± Uncertainty in Measurement (in %)
1	1000	998.9	5.8
2	2000	1997	2.9
3	3000	2995	1.9
4	4000	3992	1.2
5	5000	4990	1.2
6	6000	5988	1.2

UUC - Unit Under Calibration

Std. - Standard Instrument

Uncertainty of Measurement (at approx 95% Confidence Level with Coverage factor $k = 2$)

Calibrated By : Rajesh Gupta
 Designation : Calibration Engineer



Approved By : Moh. Iqbal
 Designation : Tech Manager / Auth. Sign.

- Conditions :
1. This certificate refers only to the particular item submitted for calibration.
 2. The calibration result reported in this certificate are valid at the time of and under the stated conditions of measurement.
 3. This particular certificate can not be reproduced except in full, without prior permission of chief executive officer of the lab.

Parashar Micro Measurement Pvt. Ltd.



The way to Technology
(15 Years of Excellence)

B-59, SECTOR-64, NOIDA (U.P.) 201 307
PH.: 0120-4252360-65, Email : pmmpl.india@gmail.com
Website : www.pmmpl.in, www.parashar.info
(NABL ACCREDITED LABORATORY)



— Calibration Certificate —

Page 1 of 1

Certificate Number : CAL/WB/S/466355		Date of Request	15.12.2018
Calibrated For	M/s. Fatima Hospital Mother Teresa Road, Padri Bazar PO, Gorakhpur, Uttar Pradesh - 273014	Calibrated on	19.12.2018
		Valid Upto	18.12.2019
		Cert. Issue Date	21.12.2018

Description of Sample

Nomenclature	Water Bath (Thermostat)	Accuracy/Class	Not Mentioned
Range	0 to 110 ° C	Condition of UUC	Physically Ok
Resolution	5° C	Calibration Performed at	At Site
Make & Model No.	Ambassador	Calibration Procedure No.	CP/T - 03
Serial No.	_____	Reference IS Standard	IS:7358 - 1984
Party ID Mark	_____	Environmental Condition	Temperature = 25 ± 5° C Humidity = 40 to 70% RH
Location	Laboratory		

Master Equipment / Standard Used

Sr.N.	Nomenclature	Make	Serial No. / I/D Mark	Calibrated From	Certificate No.	Due Date of Calibration
1	RTD With Dig. Indicator	Tempsens Fluke	PMM/RTD/03 PMM/CTR/01	Tempsens Udaipur	TL/218/109.1.1	06.05.2019

Standard used for calibration are traceable to National Standards through unbroken chain of calibration.

CALIBRATION RESULTS


S.N.	Nominal Value in UUC (in ° C)	Actual Observed Value in Std. (in ° C)	Error (in ° C)
1	35	34.62	0.38
2	40	39.48	0.52
3	50	49.32	0.68
4	60	59.10	0.90
5	80	78.86	1.14

UUC - Unit Under Calibration

Std. - Standard Instrument

Uncertainty of Measurement (at approx 95% Confidence Level with Coverage factor $k = 2$) = ± 3.3 ° C

Calibrated By : Rajesh Gupta
Designation : Calibration Engineer



Approved By : Moh. Iqbal
Designation : Tech Manager / Auth. Sign.

- Conditions :
1. This certificate refers only to the particular item submitted for calibration.
 2. The calibration result reported in this certificate are valid at the time of and under the stated conditions of measurement.
 3. This particular certificate can not be reproduced except in full, without prior permission of chief executive officer of the lab.

Parashar Micro Measurement Pvt. Ltd.



B-59, SECTOR-64, NOIDA (U.P.) 201 307
 PH.: 0120-4252360-65, Email : pmmpl.india@gmail.com
 Website : www.pmmpl.in, www.parashar.info
 (NABL ACCREDITED LABORATORY)



— Calibration Certificate —

Page 1 of 1

Certificate Number : CAL/INS/466351		Date of Request	15.12.2018
Calibrated For	M/s. Fatima Hospital	Calibrated on	19.12.2018
	Mother Teresa Road, Padri Bazar PO,	Valid Upto	18.12.2019
	Gorakhpur, Uttar Pradesh - 273014	Cert. Issue Date	21.12.2018

Description of Sample

Nomenclature	Incubator (Thermostat)	Accuracy/Class	Not Mentioned
Range	0 to 110 ° C	Condition of UUC	Physically Ok
Resolution	5° C	Calibration Performed at	At Site
Make & Model No.	Ambassder	Calibration Procedure No.	CP/T - 03
Serial No.	_____	Reference IS Standard	IS:7358 - 1984
Party ID Mark	_____	Environmental Condition	Temperature = 25 ± 5° C Humidity = 40 to 70% RH
Location	Laboratory		

Master Equipment / Standard Used.

Sr.N.	Nomenclature	Make	Serial No. / I/D Mark	Calibrated From	Certificate No.	Due Date of Calibration
1	RTD With Dig. Indicator	Tempsens Fluke	PMM/RTD/03- PMM/CTR/01	Tempsens Udaipur	TL/218/109.1.1	06.05.2019

Standard used for calibration are traceable to National Standards through unbroken chain of calibration.

CALIBRATION RESULTS

S.N.	Nominal Value in UUC (in ° C)	Actual Observed Value in Std. (in ° C)	Error (in ° C)
1	35	35.23	-0.23
2	40	40.30	-0.30
3	50	50.46	-0.46
4	60	60.62	-0.62
5	80	80.89	-0.89

UUC - Unit Under Calibration

Std. - Standard Instrument

Uncertainty of Measurement (at approx 95% Confidence Level with Coverage factor $k = 2$) = ± 3.3 ° C

Calibrated By : Rajesh Gupta
 Designation : Calibration Engineer



Approved By : Moh. Iqbal
 Designation : Tech Manager / Auth. Sign.

- Conditions :
1. This certificate refers only to the particular item submitted for calibration.
 2. The calibration result reported in this certificate are valid at the time of and under the stated conditions of measurement.
 3. This particular certificate can not be reproduced except in full, without prior permission of chief executive officer of the lab.

Parashar Micro Measurement Pvt. Ltd.

2018



The way to Technology
(15 Years of Excellence)

B-59, SECTOR-64, NOIDA (U.P.) 201 307
 PH.: 0120-4252360-65, Email : pmmpl.india@gmail.com
 Website : www.pmmpl.in, www.parashar.info
 (NABL ACCREDITED LABORATORY)



— Calibration Certificate —

Page 1 of 1

Certificate Number : CAL/BM/S/466352		Date of Request	15.12.2018
Calibrated For	M/s. Fatima Hospital	Calibrated on	19.12.2018
	Mother Teresa Road, Padri Bazar PO, Gorakhpur, Uttar Pradesh - 273014	Valid Upto	18.12.2019
		Cert. Issue Date	21.12.2018

Description of Sample

Nomenclature	Blood Mixer	Accuracy/Class	Not Mentioned
Range	Refer to Result	Condition of UUC	Physically Ok
Resolution	_____	Calibration Performed at	At Site
Make & Model No.	Lab line Equipment	Calibration Procedure No.	CP/M/MV/S/03
Serial No.	_____	Reference IS Standard	_____
Party ID Mark	_____	Environmental Condition	Temperature = 23 ± 5° C
Location	Laboratory		Humidity = 40 to 70%RH

Master Equipment / Standard Used

Sr.No.	Nomenclature	Make	Serial No. / I/D Mark	Calibrated From	Certificate No.	Due Date of Calibration
1	Dig. Tachometer	Lutron	PMM/TM/01	PMPL	PMM/TM/1075	11.09.2019

Standard used for calibration are traceable to National Standards through unbroken chain of calibration.

CALIBRATION RESULTS


Sr.No.	Actual Observed Value in Std. (in rpm)	± Uncertainty in Measurement (in %)
1	29.8	1.7


UUC - Unit Under Calibration

Std. - Standard Instrument

Uncertainty of Measurement (at approx 95% Confidence Level with Coverage factor $k = 2$)

Calibrated By : 
 Designation : Calibration Engineer



Approved By : 
 Designation : Tech Manager / Auth. Sign.

- Conditions :
1. This certificate refers only to the particular item submitted for calibration.
 2. The calibration result reported in this certificate are valid at the time of and under the stated conditions of measurement.
 3. This particular certificate can not be reproduced except in full, without prior permission of chief executive officer of the lab.

Parashar Micro Measurement Pvt. Ltd.



The way to Technology
(15 Years of Excellence)

B-59, SECTOR-64, NOIDA (U.P.) 201 307
PH.: 0120-4252360-65, Email : pmmpl.india@gmail.com
Website : www.pmmpl.in, www.parashar.info
(NABL ACCREDITED LABORATORY)



Calibration Certificate

Page 1 of 1

Certificate Number : CAL/WB/S/466350		Date of Request	15.12.2018
Calibrated For	M/s. Fatima Hospital	Calibrated on	19.12.2018
	Mother Teresa Road, Padri Bazar PO,	Valid Upto	18.12.2019
	Gorakhpur, Uttar Pradesh - 273014	Cert. Issue Date	21.12.2018

Description of Sample

Nomenclature	Water Bath (Thermostat)	Accuracy/Class	Not Mentioned
Range	0 to 110 ° C	Condition of UUC	Physically Ok
Resolution	5° C	Calibration Performed at	At Site
Make & Model No.	Tempstar	Calibration Procedure No.	CP/T - 03
Serial No.	-----	Reference IS Standard	IS:7358 - 1984
Party ID Mark	-----	Environmental Condition	Temperature = 25 ± 5° C
Location	Laboratory		Humidity = 40 to 70% RH

Master Equipment / Standard Used

Sr.N.	Nomenclature	Make	Serial No. / I/D Mark	Calibrated From	Certificate No.	Due Date of Calibration
1	RTD With Dig. Indicator	Tempsens Fluke	PMM/RTD/03 PMM/CTR/01	Tempsens Udaipur	TL/218/109.1.1	06.05.2019

Standard used for calibration are traceable to National Standards through unbroken chain of calibration.

CALIBRATION RESULTS

S.N.	Nominal Value in UUC (in ° C)	Actual Observed Value in Std. (in ° C)	Error (in ° C)
1	40	40.56	-0.56
2	50	50.79	-0.79
3	60	61.23	-1.23
4	80	81.65	-1.65


UUC - Unit Under Calibration

Std. - Standard Instrument

Uncertainty of Measurement (at approx 95% Confidence Level with Coverage factor $k = 2$) = ± 3.3 ° C

Calibrated By : Rajesh Gupta
Designation : Calibration Engineer

Approved By : Moh. Iqbal
Designation : Tech Manager / Auth. Sign.



- Conditions :
1. This certificate refers only to the particular item submitted for calibration.
 2. The calibration result reported in this certificate are valid at the time of and under the stated conditions of measurement.
 3. This particular certificate can not be reproduced except in full, without prior permission of chief executive officer of the lab.

Parashar Micro Measurement Pvt. Ltd.



B-59, SECTOR-64, NOIDA (U.P.) 201 307
 PH.: 0120-4252360-65, Email : pmmpi.india@gmail.com
 Website : www.pmmpi.in, www.parashar.info
 (NABL ACCREDITED LABORATORY)



— Calibration Certificate —

Page 1 of 1

Certificate Number : CAL/MP/466349		UUC Received on	15.12.2018
Calibrated For	M/s. Fatima Hospital	Calibrated on	19.12.2018
	Mother Teresa Road, Padri Bazar PO, Gorakhpur, Uttar Pradesh - 273014	Valid Upto	18.12.2019
		Cert. Issue Date	21.12.2018

Description of Sample

Nomenclature	Micro Pipette	Accuracy/Class	Not Mentioned
Range	250 μ l	Condition of UUC	Physically Ok
Resolution	-----	Calibration Performed at	Laboratory
Make & Model No.	Biotronics	Calibration Procedure No.	CP/M/MV/04
Serial No.	-----	Reference IS Standard	ISO-8655-6, ISO/FDIS-4787-2010 (E)
Party ID Mark	-----	Environmental Condition	Temperature = 23 \pm 2° C Humidity = 50 \pm 10% RH
Location	Laboratory		

Master Equipment / Standard Used

Sr.No.	Nomenclature	Make	Serial No. / I/D Mark	Calibrated From	Certificate No.	Due Date of Calibration
1	Weighing Balance	Sartorius	PMM/WM/06	PMMPL, Noida	PMM/WB/1066	23.05.2019

Standard used for calibration are traceable to National Standards through unbroken chain of calibration.

CALIBRATION RESULTS

Sr.No.	Nominal Value in UUC (in μ l)	Actual Observed Value Mean Value in Std. (in μ l)	Error (in μ l)	Mean Average Value at 27°C (in μ l)
1	250	250.76	-0.76	250.78


UUC:- Unit Under Calibration.

Std. - Standard Instrument

Uncertainty of Measurement (at approx 95% Confidence Level with Coverage factor $k = 2$) = $\pm 0.6 \mu$ l

Priyank Sharma

Calibrated By : Priyank Sharma
 Designation : Calibration Engineer



Moh. Iqbal

Approved By : Moh. Iqbal
 Designation : Tech Manager / Auth. Sign.

- Conditions :
1. This certificate refers only to the particular item submitted for calibration.
 2. The calibration result reported in this certificate are valid at the time of and under the stated conditions of measurement.
 3. This particular certificate can not be reproduced except in full, without prior permission of chief executive officer of the lab.

Parashar Micro Measurement Pvt. Ltd.

2018



B-59, SECTOR-64, NOIDA (U.P.) 201 307
 PH.: 0120-4252360-65, Email : pmmpl.india@gmail.com
 Website : www.pmmpl.in, www.parashar.info
 (NABL ACCREDITED LABORATORY)



Calibration Certificate

Page 1 of 1

Certificate Number : CAL/MP/466342		UUC Received on	15.12.2018
Calibrated For	M/s. Fatima Hospital Mother Teresa Road, Padri Bazar PO, Gorakhpur, Uttar Pradesh - 273014	Calibrated on	19.12.2018
		Valid Upto	18.12.2019
		Cert. Issue Date	21.12.2018

Description of Sample

Nomenclature	Micro Pipette	Accuracy/Class	Not Mentioned
Range	200 µl	Condition of UUC	Physically Ok
Resolution	-----	Calibration Performed at	Laboratory
Make & Model No.	-----	Calibration Procedure No.	CP/M/MV/04
Serial No.	323132	Reference IS Standard	ISO-8655-6, ISO/FDIS-4787-2010 (E)
Party ID Mark	-----	Environmental Condition	Temperature = 23 ± 2° C Humidity = 50 ± 10% RH
Location	Laboratory		

Master Equipment / Standard Used

Sr.No.	Nomenclature	Make	Serial No. / I/D Mark	Calibrated From	Certificate No.	Due Date of Calibration
1	Weighing Balance	Sartorius	PMM/WM/06	PMMPL, Noida	PMM/WB/1066	23.05.2019

Standard used for calibration are traceable to National Standards through unbroken chain of calibration.

CALIBRATION RESULTS

Sr.No.	Nominal Value in UUC (in µl)	Actual Observed Value Mean Value in Std. (in µl)	Error (in µl)	Mean Average Value at 27°C (in µl)
1	200	201.34	-1.34	201.36

UUC:- Unit Under Calibration.

Std. - Standard Instrument

Uncertainty of Measurement (at approx 95% Confidence Level with Coverage factor $k = 2$) = ± 0.6 µl

Priyank

Calibrated By : Priyank Sharma
 Designation : Calibration Engineer



Moh. Iqbal

Approved By : Moh. Iqbal
 Designation : Tech Manager / Auth. Sign.

- Conditions :
1. This certificate refers only to the particular item submitted for calibration.
 2. The calibration result reported in this certificate are valid at the time of and under the stated conditions of measurement.
 3. This particular certificate can not be reproduced except in full, without prior permission of chief executive officer of the lab.

Parashar Micro Measurement Pvt. Ltd.



B-59, SECTOR-64, NOIDA (U.P.) 201 307
 PH.: 0120-4252360-65, Email : pmmpl.india@gmail.com
 Website : www.pmmpl.in, www.parashar.info
 (NABL ACCREDITED LABORATORY)



— Calibration Certificate —

Page 1 of 1

Certificate Number : CAL/MP/466348		UUC Received on	15.12.2018
Calibrated For	M/s. Falima Hospital	Calibrated on	19.12.2018
	Mother Teresa Road, Padri Bazar PO,	Valid Upto	18.12.2019
	Gorakhpur, Uttar Pradesh - 273014	Cert. Issue Date	21.12.2018

Description of Sample

Nomenclature	Micro Pipette	Accuracy/Class	Not Mentioned
Range	25 µl	Condition of UUC	Physically Ok
Resolution	-----	Calibration Performed at	Laboratory
Make & Model No.	Biohit	Calibration Procedure No.	CP/MMV/04
Serial No.	DV15404	Reference IS Standard	ISO-8655-6, ISO/FDIS-4787-2010 (E)
Party ID Mark	-----	Environmental Condition	Temperature = 23 ± 2° C Humidity = 50 ± 10% RH
Location	Laboratory		

Master Equipment / Standard Used

Sr.No.	Nomenclature	Make	Serial No. / I/D Mark	Calibrated From	Certificate No.	Due Date of Calibration
1	Weighing Balance	Sartorius	PMM/WM/06	PMMPL, Noida	PMM/WB/1066	23.05.2019

Standard used for calibration are traceable to National Standards through unbroken chain of calibration.

CALIBRATION RESULTS


Sr.No.	Nominal Value in UUC (in µl)	Actual Observed Value Mean Value in Std. (in µl)	Error (in µl)	Mean Average Value at 27°C (in µl)
1	25	24.93	0.07	24.93

UUC:- Unit Under Calibration.

Std. - Standard Instrument

Uncertainty of Measurement (at approx 95% Confidence Level with Coverage factor $k = 2$) = ± 0.15 µl

Calibrated By : *Priyank Sharma*
 Designation : Calibration Engineer



Approved By : *Moh. Iqbal*
 Designation : Tech Manager / Auth. Sign.

- Conditions :
1. This certificate refers only to the particular item submitted for calibration.
 2. The calibration result reported in this certificate are valid at the time of and under the stated conditions of measurement.
 3. This particular certificate can not be reproduced except in full, without prior permission of chief executive officer of the lab.

Parashar Micro Measurement Pvt. Ltd.

2018



B-59, SECTOR-64, NOIDA (U.P.) 201 307
 PH.: 0120-4252360-65, Email : pmmpl.india@gmail.com
 Website : www.pmmpl.in, www.parashar.info
 (NABL ACCREDITED LABORATORY)



Calibration Certificate

Page 1 of 1

Certificate Number : CAL/MP/466347		UUC Received on	15.12.2018
Calibrated For	M/s. Fatima Hospital Mother Teresa Road, Padri Bazar PO, Gorakhpur, Uttar Pradesh - 273014	Calibrated on	19.12.2018
		Valid Upto	18.12.2019
		Cert. Issue Date	21.12.2018

Description of Sample

Nomenclature	Micro Pipette	Accuracy/Class	Not Mentioned
Range	5 to 50 μ l	Condition of UUC	Physically Ok
Resolution	0.5 μ l	Calibration Performed at	Laboratory
Make & Model No.	Thermo Scientific	Calibration Procedure No.	CP/MMV/04
Serial No.	CH31778	Reference IS Standard	ISO-8655-6, ISO/FDIS-4787-2010 (E)
Party ID Mark	---	Environmental Condition	Temperature = $23 \pm 2^\circ$ C Humidity = $50 \pm 10\%$ RH
Location	Laboratory		

Master Equipment / Standard Used

Sr.No.	Nomenclature	Make	Serial No. / I/D Mark	Calibrated From	Certificate No.	Due Date of Calibration
1	Weighing Balance	Sartorius	PMM/MM/06	PMMPL, Noida	PMM/WB/1066	23.05.2019

Standard used for calibration are traceable to National Standards through unbroken chain of calibration.

CALIBRATION RESULTS

Sr.No.	Nominal Value in UUC (in μ l)	Actual Observed Value Mean Value in Std. (in μ l)	Error (in μ l)	Mean Average Value at 27° C (in μ l)
1	5*	5.05	-0.05	5.05
2	25	25.11	-0.11	25.11
3	50	50.23	-0.23	50.24

UUC:- Unit Under Calibration.

Std. - Standard Instrument

Uncertainty of Measurement (at approx 95% Confidence Level with Coverage factor $k = 2$) = $\pm 0.3 \mu$ l

Sharma

Calibrated By : Priyank Sharma
 Designation : Calibration Engineer



Moh. Iqbal

Approved By : Moh. Iqbal
 Designation : Tech Manager / Auth. Sign.

- Conditions :
1. This certificate refers only to the particular item submitted for calibration.
 2. The calibration result reported in this certificate are valid at the time of and under the stated conditions of measurement.
 3. This particular certificate can not be reproduced except in full, without prior permission of chief executive officer of the lab.

Parashar Micro Measurement Pvt. Ltd.



The way to Technology
(15 Years of Excellence)

B-59, SECTOR-64, NOIDA (U.P.) 201 307
PH.: 0120-4252360-65, Email : pmmpl.india@gmail.com
Website : www.pmmpl.in, www.parashar.info
(NABL ACCREDITED LABORATORY)



— Calibration Certificate —

Page 1 of 1

Certificate Number : CAL/MP/466346		UUC Received on	15.12.2018
Calibrated For	M/s. Fatima Hospital	Calibrated on	19.12.2018
	Mother Teresa Road, Padri Bazar PO,	Valid Upto	18.12.2019
	Gorakhpur, Uttar Pradesh - 273014	Cert. Issue Date	21.12.2018

Description of Sample

Nomenclature	Micro Pipette	Accuracy/Class	Not Mentioned
Range	50 µl	Condition of UUC	Physically Ok
Resolution	-----	Calibration Performed at	Laboratory
Make & Model No.	Thermo Scientific	Calibration Procedure No.	CP/M/MV/04
Serial No.	CH16532	Reference IS Standard	ISO-8655-6, ISO/FDIS-4787-2010 (E)
Party ID Mark	-----	Environmental Condition	Temperature = 23 ± 2° C Humidity = 50 ± 10% RH
Location	Laboratory		

Master Equipment / Standard Used

Sr.No.	Nomenclature	Make	Serial No. / I/D Mark	Calibrated From	Certificate No.	Due Date of Calibration
1	Weighing Balance	Sartorius	PMM/WM/06	PMMPL, Noida	PMM/WB/1066	23.05.2019

Standard used for calibration are traceable to National Standards through unbroken chain of calibration.

CALIBRATION RESULTS

Sr.No.	Nominal Value in UUC (in µl)	Actual Observed Value Mean Value in Std. (in µl)	Error (in µl)	Mean Average Value at 27°C (in µl)
1	50	50.10	-0.10	50.11

UUC:- Unit Under Calibration.

Std. - Standard Instrument

Uncertainty of Measurement (at approx 95% Confidence Level with Coverage factor $k = 2$) = ± 0.15 µl

Sharma

Calibrated By : Priyank Sharma
Designation : Calibration Engineer



Moh. Iqbal

Approved By : Moh. Iqbal
Designation : Tech Manager / Auth. Sign.

- Conditions :
1. This certificate refers only to the particular item submitted for calibration.
 2. The calibration result reported in this certificate are valid at the time of and under the stated conditions of measurement.
 3. This particular certificate can not be reproduced except in full, without prior permission of chief executive officer of the lab.

Parashar Micro Measurement Pvt. Ltd.

2018



The way to Technology
(15 Years of Excellence)

B-59, SECTOR-64, NOIDA (U.P.) 201 307
 PH.: 0120-4252360-65, Email : pmmpl.india@gmail.com
 Website : www.pmmpl.in, www.parashar.info
 (NABL ACCREDITED LABORATORY)



Calibration Certificate

Page 1 of 1

Certificate Number : CAL/MP/466345		UUC Received on	15.12.2018
Calibrated For	M/s. Fatima Hospital	Calibrated on	19.12.2018
	Mother Teresa Road, Padri Bazar PO,	Valid Upto	18.12.2019
	Gorakhpur, Uttar Pradesh - 273014	Cert. Issue Date	21.12.2018

Description of Sample

Nomenclature	Micro Pipette	Accuracy/Class	Not Mentioned
Range	500 µl	Condition of UUC	Physically Ok
Resolution	-----	Calibration Performed at	Laboratory
Make & Model No.	Biotech	Calibration Procedure No.	CP/M/MV/04
Serial No.	913074	Reference IS Standard	ISO-8655-6, ISO/FDIS-4787-2010 (E)
Party ID Mark	-----	Environmental Condition	Temperature = 23 ± 2° C Humidity = 50 ± 10% RH
Location	Laboratory		

Master Equipment / Standard Used

Sr.No.	Nomenclature	Make	Serial No. / I/D Mark	Calibrated From	Certificate No.	Due Date of Calibration
1	Weighing Balance	Sartorius	PMM/WM/06	PMMPL, Noida	PMM/WB/1066	23.05.2019

Standard used for calibration are traceable to National Standards through unbroken chain of calibration.

CALIBRATION RESULTS

Sr.No.	Nominal Value in UUC (in µl)	Actual Observed Value Mean Value in Std. (in µl)	Error (in µl)	Mean Average Value at 27°C (in µl)
1.	500	498.53	1.47	498.58

UUC:- Unit Under Calibration.

Std. - Standard Instrument

Uncertainty of Measurement (at approx 95% Confidence Level with Coverage factor $k = 2$) = ± 0.6 µl

P. Sharma

Calibrated By : Priyank Sharma
 Designation : Calibration Engineer



Moh. Iqbal

Approved By : Moh. Iqbal
 Designation : Tech Manager / Auth. Sign.

- Conditions :
1. This certificate refers only to the particular item submitted for calibration.
 2. The calibration result reported in this certificate are valid at the time of and under the stated conditions of measurement.
 3. This particular certificate can not be reproduced except in full, without prior permission of chief executive officer of the lab.

Parashar Micro Measurement Pvt. Ltd.

2018



The way to Technology
(15 Years of Excellence)

B-59, SECTOR-64, NOIDA (U.P.) 201 307
 PH.: 0120-4252360-65, Email : pmmpl.india@gmail.com
 Website : www.pmmpl.in, www.parashar.info
 (NABL ACCREDITED LABORATORY)



Calibration Certificate

Page 1 of 1

Certificate Number : CAL/MP/ 466344		UUC Received on	15.12.2018
Calibrated For	M/s. Fatima Hospital	Calibrated on	19.12.2018
	Mother Teresa Road, Padri Bazar PO,	Valid Upto	18.12.2019
	Gorakhpur, Uttar Pradesh - 273014	Cert. Issue Date	21.12.2018

Description of Sample

Nomenclature	Micro Pipette	Accuracy/Class	Not Mentioned
Range	100 µl	Condition of UUC	Physically Ok
Resolution	-----	Calibration Performed at	Laboratory
Make & Model No.	Finnpipette	Calibration Procedure No.	CP/MMV/04
Serial No.	S343754	Reference IS Standard	ISO-8655-6, ISO/FDIS-4787-2010 (E)
Party ID Mark	-----	Environmental Condition	Temperature = 23 ± 2° C Humidity = 50 ± 10% RH
Location	Laboratory		

Master Equipment / Standard Used

Sr.No.	Nomenclature	Make	Serial No. / I/D Mark	Callibrated From	Certificate No.	Due Date of Calibration
1	Weighing Balance	Sartorius	PMM/WM/06	PMMPL, Noida	PMM/WB/1066	23.05.2019

Standard used for calibration are traceable to National Standards through unbroken chain of calibration.

CALIBRATION RESULTS

Sr.No.	Nominal Value in UUC (in µl)	Actual Observed Value Mean Value in Std. (in µl)	Error (in µl)	Mean Average Value at 27°C (in µl)
1	100	99.79	0.21	99.80

UUC:- Unit Under Calibration.

Std. - Standard Instrument

Uncertainty of Measurement (at approx 95% Confidence Level with Coverage factor $k = 2$) = ± 0.15 µl

Sharma

Calibrated By : Priyank Sharma
 Designation : Calibration Engineer



Moh. Iqbal

Approved By : Moh. Iqbal
 Designation : Tech Manager / Auth. Sign.

Conditions : 1. This certificate refers only to the particular item submitted for calibration.

2. The calibration result reported in this certificate are valid at the time of and under the stated conditions of measurement.

3. This particular certificate can not be reproduced except in full, without prior permission of chief executive officer of the lab.

Parashar Micro Measurement Pvt. Ltd.



The way to Technology
(15 Years of Excellence)

B-59, SECTOR-64, NOIDA (U.P.) 201 307
PH.: 0120-4252360-65, Email : pmmpl.india@gmail.com
Website : www.pmmpl.in, www.parashar.info
(NABL ACCREDITED LABORATORY)



Calibration Certificate

Page 1 of 1

Certificate Number : CAL/MP/466343		UUC Received on	15.12.2018
Calibrated For	M/s. Fatima Hospital	Calibrated on	19.12.2018
	Mother Teresa Road, Padri Bazar PO,	Valid Upto	18.12.2019
	Gorakhpur, Uttar Pradesh - 273014	Cert. Issue Date	21.12.2018

Description of Sample

Nomenclature	Micro Pipette	Accuracy/Class	Not Mentioned
Range	100 to 1000 μ l	Condition of UUC	Physically Ok
Resolution	5 μ l	Calibration Performed at	Laboratory
Make & Model No.	-----	Calibration Procedure No.	CP/M/MV/04
Serial No.	NC852306	Reference IS Standard	ISO-8655-6, ISO/FDIS-4787-2010 (E)
Party ID Mark	-----	Environmental Condition	Temperature = $23 \pm 2^\circ$ C Humidity = $50 \pm 10\%$ RH
Location	Laboratory		

Master Equipment / Standard Used

Sr.No.	Nomenclature	Make	Serial No. / I/D Mark	Calibrated From	Certificate No.	Due Date of Calibration
1	Weighing Balance	Sartorius	PMM/WM/06	PMMPL, Noida	PMM/WB/1066	23.05.2019

Standard used for calibration are traceable to National Standards through unbroken chain of calibration.

CALIBRATION RESULTS

Sr.No.	Nominal Value in UUC (in μ l)	Actual Observed Value Mean Value in Std. (in μ l)	Error (in μ l)	Mean Average Value at 27° C (in μ l)
1	100	100.60	-0.60	100.61
2	500	501.32	-1.32	501.37
3	1000	1002.64	-2.64	1002.75

UUC:- Unit Under Calibration.

Std. - Standard Instrument

Uncertainty of Measurement (at approx 95% Confidence Level with Coverage factor $k = 2$) = $\pm 3 \mu$ l

Priyank

Calibrated By : Priyank Sharma
Designation : Calibration Engineer



Mohd

Approved By : Moh. Iqbal
Designation : Tech Manager / Auth. Sign.

- Conditions : 1. This certificate refers only to the particular item submitted for calibration.
2. The calibration result reported in this certificate are valid at the time of and under the stated conditions of measurement.
3. This particular certificate can not be reproduced except in full, without prior permission of chief executive officer of the lab.

CALIBRATION REPORT

Equipment	Coagulation Analyzer
Model	KC1 Delta
Serial number	17G63129

CALIBRATION Validity

Calibration performed date	12-12-2018
Calibration validity	Six Months
Next Calibration Due on	June 2019

Please check the following points before calibration the instrument		Range	Yes/No
1	The power supply is 220 Volt 10% inputs Grounding is 0 + or -10 Output Voltage 12V + or -1V	0 - 10	Yes
2	Magnetic strength by FCM meter. Range (90 + or - 2)	88 - 92 uA	Yes
3	Fresh Cuvette used	New cuvette	Yes
4	Cuvette rotation / minute	38 - 42 rpm	Yes
5	Cuvette Temperature	37 °C + 0.1	Yes
6	Reagent and control material are store at	2 °C to 8 °C	Yes

After Calibration controls were run as per following details:

STA-Coagulation Control Reagents 1 :: Lot No.252633 :: Exp : 09/2019				
Test Name	Range	Target	Results Obtained	Remarks
PT	12.5-15.5	13.5	13.7	Ok
APTT	28.0-36.0	32.0	33.7	Ok

Control Date :12.12.2018

Q C Department

M. K. Singh

CALIBRATION REPORT

1. Instrument : Semi Automated Bio Chemistry Analyzer
2. Model : BTS-350
3. Sr.No. : 801757459
4. Date of Calibration : 04.12.2018

TEST PARAMETERS

Test Parameters	Actual Parameters			OK/Not OK
SMPS Voltage 15.0 V	15.05 V			OK
Flow Cell Temperature	25°C	30°C	37°C	OK
Pump Calibration Value (12000 to 45000)	21000			OK
Positioning	1100			OK
Pump Delay	2			OK

Sr.No.	Filters	Gain		Range
		PH1	PH2	
1	340 nm	979959	248534	200000-1000000
2	560 nm	967069	370210	200000-1000000
3	670 nm	966624	369922	200000-1000000
4	600 nm	978499	430719	200000-1000000
5	535 nm	961981	477213	200000-1000000
6	505 nm	976119	643852	200000-1000000
7	635 nm	970104	516798	200000-1000000
8	405 nm	979938	667987	200000-1000000
9	Filter Offset	4099	4105	2500-4500

Photometric Accuracy :

The reading of the absorbance are taken of following standard filters at the frequency of 505 nm. The results are compared with its range.

Range of the Instrument: Optical Density Linear From 0 to 2.5

Std Absorbance	Measured Absorbance	Range
0.3	0.299	0.3000-0.3060
0.6	0.600	0.6000-0.6050
1.0	1.002	1.0000-1.0900
2.0	2.003	2.0000-2.0800

Precision Control :

The absorbance of two standard filters of 0.4 and 2.0 is taken. The replicates of each is made and mentioned as CV(%).

Std Absorbance	CV (%)
0.4	0.04 L%
2.0	0.006 %

Clinical Chemistry Control :

CHEMISTRY	Level -1			Level -2		
	Measured	Mean	Range	Measured	Mean	Range
GLU	93.1	93	74.4-111.6	243.4	251	213.4-288.7
SGPT	49.3	47	37.6-56.4	112.5	128	102.4-153.6
CRTN	0.10	0.93	0.74-1.11	3.31	3.8	3.23-4.37

Date: 04.12.18


Q.C.

CALIBRATION CERTIFICATE

Customer Name	FATIMA HOSPITAL
Adress	Mother Teresa Road Padri Bazar Gorakhpur

CALIBRATION Report for

Equipment	Heamatology Analyzer
Model	MEK-6420P
Serial number	01252735

CALIBRATION Validity

Calibration performed date	27-12-2018
Calibration validity	One Year
Next Calibration Due on	December 2019

CHECKS BEFORE CALIBRATION

PHYSICA CHECKS:

Outer Casing/Chassis O.K.(No Damage)	Yes : <input checked="" type="checkbox"/> / No: <input type="checkbox"/>
UPS Power Supply OK	Yes : <input checked="" type="checkbox"/> / No: <input type="checkbox"/>
Grounding OK	Yes : <input checked="" type="checkbox"/> / No: <input type="checkbox"/>

PERFORMANCE CHECKS AFTER CALIBRATION

HGB Voltage	1.930
WBC Voltage	18.39
RBC Voltage	18.15
Diluent Voltage	1.51 V
Lysing Reagent	0.99 V
HGB Unit Temp.(Degree Cent.)	26.7
MC Unit.(Degree Cent.)	25.0
Power Board.(Degree Cent.)	26.7

Calibrator Details :-

- SC-CAL PLUS
- LOT NO-PLUS 0119
- EXP- 05-02.2019

After calibration Q.C. Test for all level passed successfully.

Test results is coming ok	✓	Remarks	Done
---------------------------	---	---------	------

Service Engineer





LABCARE

Care for quality • Care for customer

Labcare Bio-Medicals

CALIBRATION REPORT

1. Instrument : Semi Automated Bio Chemistry Analyzer
2. Model : BTS-350
3. Sr.No. : 801757459
4. Date of Calibration : 04.12.2018

TEST PARAMETERS

Test Parameters	Actual Parameters			OK/Not OK
SMPS Voltage 15.0 V	15.05 V			OK
Flow Cell Temperature	25°C	30°C	37°C	OK
Pump Calibration Value (12000 to 45000)	21000			OK
Positioning	1100			OK
Pump Delay	2			OK

Sr.No.	Filters	Gain		Range
		PH1	PH2	
1	340 nm	979959	248534	200000-1000000
2	560 nm	967069	370210	200000-1000000
3	670 nm	966624	369922	200000-1000000
4	600 nm	978499	430719	200000-1000000
5	535 nm	961981	477213	200000-1000000
6	505 nm	976119	643852	200000-1000000
7	635 nm	970104	516798	200000-1000000
8	405 nm	979938	667987	200000-1000000
9	Filter Offset	4099	4105	2500-4500



Care for quality • Care for customer

Photometric Accuracy: **labcare Bio-Medicals**

The reading of the absorbance are taken of following standard filters at the frequency of 505 nm. The results are compared with its range.

Range of the Instrument: Optical Density Linear From 0 to 2.5

Std Absorbance	Measured Absorbance	Range
0.3	0.299	0.3000-0.3060
0.6	0.600	0.6000-0.6050
1.0	1.002	1.0000-1.0900
2.0	2.003	2.0000-2.0800

Precision Control :

The absorbance of two standard filters of 0.4 and 2.0 is taken. The replicates of each is made and mentioned as CV (%).

Std Absorbance	CV (%)
0.4	0.04 L%
2.0	0.006 %

Clinical Chemistry Control:

CHEMISTRY	Level -1			Level -2		
	Measured	Mean	Range	Measured	Mean	Range
GLU	93.1	93	74.4-111.6	243.4	251	213.4-288.7
SGPT	49.3	47	37.6-56.4	112.5	128	102.4-153.6
CRTN	0.10	0.93	0.74-1.11	3.31	3.8	3.23-4.37

Date: 04.12.18

Q.C.

Shankar



LABCARE

Care for quality • Care for customer

Labcare Bio-Medicals

CALIBRATION CERTIFICATE

Customer Name	FATIMA HOSPITAL
Adress	Mother Teresa Road Padri Bazar Gorakhpur

CALIBRATION Report for

Equipment	Heamatology Analyzer
Model	MEK-6420P
Serial number	01252735

CALIBRATION Validity

Calibration performed date	27-12-2018
Calibration validity	One Year
Next Calibration Due on	December 2019

CHECKS BEFORE CALIBRATION

PHYSICA CHECKS:

Outer Casing/Chassis O.K.(No Damage)	Yes : <input checked="" type="checkbox"/> / No: <input type="checkbox"/>
UPS Power Supply OK	Yes : <input checked="" type="checkbox"/> / No: <input type="checkbox"/>
Grounding OK	Yes : <input checked="" type="checkbox"/> / No: <input type="checkbox"/>

214 - DIWAN BAZAR, GORAKHPUR-273001 (U.P.) INDIA

Mobile : 9935072701 Phone : (+91) 0551-6545680

E-mail : labcares@yahoo.co.in



LABCARE

Care for quality • Care for customer

Labcare Bio-Medicals

PERFORMANCE CHECKS AFTER CALIBRATION

HGB Voltage	1.930
WBC Voltage	18.39
RBC Voltage	18.15
Diluent Voltage	1.51 V
Lysing Reagent	0.99 V
HGB Unit Temp. (Degree Cent.)	26.7
MC Unit. (Degree Cent.)	25.0
Power Board. (Degree Cent.)	26.7

Calibrator Details:-

- SC-CAL PLUS
- LOT NO-PLUS 0119
- EXP- 05-02.2019

After calibration Q.C. Test for all level passed successfully.

Test results is coming ok	✓	Remarks	Done
---------------------------	---	---------	------

Service Engineer

Chauhan

214 - DIWAN BAZAR, GORAKHPUR-273001 (U.P.) INDIA

Mobile : 9935072701 Phone : (+91) 0551-6545680

E-mail : labcares@yahoo.co.in

Calibration Certificate

Certificate No.:	CCTPL/05-07/01	Field:	Mechanical
SRF No.:	CCTPL/0107/01	Date of Receipt:	30/07/2021
SRF Date:	30/07/2021	Date of Calibration:	30/07/2021
Company Name & Address	M/s. Fatima Hospital Mother Taresa Road, Padri Bazar Gorkhpur Pin-273014	Recommended Due Date.: (As per agreed by customer)	29/07/2022
		Certificate Issue Date:	31/07/2021

DUC* Details

Instrument Name:	Micro Pipette	Instrument Sr. No.:	----
Make:	-----	Instrument Id No.:	----
Model No.:	-----	Location:	-----
Range:	10 to 100 µl	DUC Condition:	Ok
Readability:	5 µl	Calibration Performed at:	At Lab

Standard Equipments Used(Traceable to National Standard)

Sr. No	Instrument Name	Cert. No	Cal. By	Calibration Due Date
1	Analytical Balance	PMM/AB/12080/1	PMM, Noida	06/05/2022

Environment Condition	Temperature	Humidity	Reference Standard	Calibration Procedure
	23±2°C	55±10%RH	ISO: 8655-6	CCTPLP/M/MV-02

CALIBRATION RESULT

Sr. No.	DUC Value (in µl)	Observed Value on Standard (in mg)	Error (in µl)	Mean Value at 27°C (in µl)
1	10	9.84	0.16	9.84
2	50	49.78	0.22	49.78
3	100	99.84	0.15	99.85

Note : DUC* : Device Under Calibration

Uncertainty of measurement at approx 95% Confidence Level with coverage factor k = 2, is ± 1.0 µl.

Calibrated By : Priyank Sharma
Calibration Engineer

Approved By : Sandeep Sharma
Director

Note. 1.This Certificate only to the particular instrument submitted for calibration.
2.This calibration Result reported in this certificate valid at the time of and under the stated condition of measurement.
3.This particular certificate can not be reproduce except in full, without prior permission of CCTPL

~---End of Certificate---

Page 1 of 1

Calibration Certificate

Certificate No.:	CCTPL/28-01/08	Field:	Mechanical
SRF No.:	CCTPL/0101/01	Date of Receipt:	03/02/2021
SRF Date:	03/02/2021	Date of Calibration:	04/02/2021
Company Name & Address	M/s. Fatima Hospital Mother Taresa Road, Padri Bazar, Gorakhpur. Pin - 273014	Recommended Due Date.:	03/02/2022
		As per agreed by customer	
		Certificate Issue Date:	04/02/2021

DUC* Details

Instrument Name:	Micro Pipette	Instrument Sr. No.:	---
Make:	----	Instrument Id No.:	----
Model No.:	----	Location:	Lab
Range:	10 - 100 µl	DUC Condition:	Ok
Readability:	5 µl	Calibration Performed at:	At Lab

Standard Equipments Used(Traceable to National Standard)

Sr. No	Instrument Name	Cert. No	Cal. By	Calibration Due Date
1	Analytical Balance	PMM/AB/9374/1	PMM, Noida	31/05/2021

Environment Condition	Temperature	Humidity	Reference Standard	Calibration Procedure
	23±2°C	55±10%RH	ISO: 8655-6	CCTPLP/M/MV-02

CALIBRATION RESULT


Sr. No.	DUC Value (in µl)	Observed Value on Standard (in mg)	Error (in µl)	Mean Value at 27°C (in µl)
1	10	9.84	0.16	9.84
2	50	24.91	25.09	24.91
3	100	99.84	0.16	99.85

Note : DUC* : Device Under Calibration

Uncertainty of measurement at approx 95% Confidence Level with coverage factor k = 2, is ± 1.1 µl.

Calibrated By :

P. Sharma
Priyank Sharma
Calibration Engineer



Approved By :

S. Sharma
Sandeep Sharma
Director

- Note.**
1. This Certificate only to the particular instrument submitted for calibration.
 2. This Calibration Result reported in this certificate valid at the time of and under the stated condition of measurement.
 3. This particular certificate can not be reproduce except in full, without prior permission of CCTPL

—End of Certificate—

Page 1 of 1