




Vaidyanatheshwara

CERTIFICATE OF CALIBRATION



No. 50/1A, 9th Main Road, 3rd Cross, Rajiv Gandhi Nagar, J.S. Kaveri, Nandhini Layout Post, Bangalore - 560 096
 Contact : 080-23377266, Mob : 9986586733 / 9446080177 / 9964300319 | Email : info@viplgroup.com Web : www.viplgroup.com

NABL Accredited Calibration Lab as per ISO/IEC 17025 : 2017 With vide Certificate No: CC-2473

1 Name and Address of the Customer

M/s.: NU Hospitals,
 CA-P1, Machenahalli Industrial Area
 Bhadravathi Taluk, Shivamogga,
 Karnataka 577222.

2 Customer Reference

- 2.1 ULR No. : CC247321300002901F
- 2.2 Format No. : VI-FRM-TH-004
- 2.3 SRF.No : 1780
- 2.4 DC No / Date : 414 / 21-06-2021
- 2.5 Certificate No. : VI/21-22/1780-10
- 2.6 Date of Issue : 22-06-2021

3 Details Of Device Under Calibration(DUC)

- 3.1 Nomenclature : Digital Thermometer
- 3.2 Make / Model : Mextech / DT-9
- 3.3 SI.No / Id. No : --- / DTM-01
- 3.4 DUC Condition : Satisfactory
- 3.5 Calibration Ref. Std : SOP-38-04
- 3.6 No. of Pages : 2
- 3.7 Calibration Date : 22-06-2021
- 3.8 Calibration Due : 21-06-2022
- 3.9 Calibration done at : VI Thermal Lab
- 3.10 Discipline : Thermal

4 Environmental Condition

Temperature 25.3 °C Humidity 53 %Rh

5 Standards Used for Temperature Calibration

Sl. No.	Nomenclature	Make / Model	ID.No.	Certificate. No.	Validity
1	Platinum Resistance Thermometer With Indicator	FLUKE / 5609 (Sensor) & 1524 (Indicator)	03520 (Sensor) & 3028237 (Indicator)	CR/NCAL/51263	27-11-2021
2	Dry Block Calibrator is used as source.				

6 Note:

- 6.1 The Calibration Certificate relates only to the above DUC
- 6.2 Publication or reproduction of this Certificate in any form other than by complete set of the whole report & in the language, written, is not permitted without the written consent of VI Lab.
- 6.3 Corrections/erasing, invalidate the Calibration Certificate.
- 6.4 Calibration of the DUC are traceable to National standards/International Standards
- 6.5 Any error in this Certificate should be brought to our knowledge within 30 days from the date of this Cert.
- 6.6 Results Reported are valid at the time of and under the stated conditions of measurements.
- 6.7 The usage of NABL Symbol is as per NABL guidelines given in NABL-133

Calibrated By


Kiran R
 (Calibration Engineer)

Checked By

Umesh.D
 (Lab In-Charge)

Authorised By

Gangadhar.C.K.





COMMITTED TO THE CUSTOMER SINCE 1998

VAIDYANATHESHWARA INSTRUMENTS

CERTIFICATE OF CALIBRATION



No. 301/A, 9th Main Road, 3rd Cross, Rajiv Gandhi Nagar, J.B. Kaval, Nandhini Layout Post, Bangalore - 560 096
Contact : 080-23377266, Mob : 9980530733 / 3448080177 / 9964303118 | Email : info@vigroup.com Web : www.viplgroup.com

NABL Accredited Calibration Lab as per ISO/IEC 17025 : 2017 With vide Certificate No: CC-2473

Certificate No. VI/21-22/1780-10

Page: 2 of 2

Results:

Range / Least Count : -50 to 300 °C / 0.1°C

Table with 7 columns: Sl No, Set Point, STD Reading (°C), DUC Reading (°C), Error Observed (°C), Measurement Uncertainty ± (°C), k Factor. Contains 8 rows of calibration data.

Conclusion/Remarks:-

1 Measurement uncertainty reported is at 95.45 % confidence level with k=2.

Calibrated By

Kiran. R

(Calibration Engineer)

Checked By

Umesh.D

(Lab In-Charge)

Authorised By

Gangadhar C.K

