

## LAB MONTHLY SUMMARY



Lab Name **THRIVENI SAINIK MINING P LTD** Lab No **8369**  
 Month **June** Year **2021**  
 Constituent Group **Chemistry II**



Date of Result Entered : 19/06/2021

Date of Report Published : 03/07/2021

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE II	GOD-POD II	Any Analyser (Automation / Semi Automation)	940	193.80	10.52	20.38	202 mg/dl	0.40	1.33
2	UREA II	Urease UV / GLDH II	Any Analyser (Automation / Semi Automation)	580	74.92	14.27	10.69	72 mg/dl	-0.27	0.89
3	CREATININE II	Jaffes Kinetic-Alkaline Picrate II	Any Analyser (Automation / Semi Automation)	675	4.42	11.68	0.52	4.8 mg/dl	0.74	0.04
4	T.BILIRUBIN II	Diazonium Salt (Colorimetric) / Jendrassik II	Any Analyser (Automation / Semi Automation)	800	3.86	16.46	0.64	4.57 mg/dl	1.12	0.04
5	T-PROTEIN II	Biuret - Colorimetric II	Any Analyser (Automation / Semi Automation)	729	5.51	11.11	0.61	5.55 g/dl	0.07	0.05
6	ALBUMIN II	BCG - Colorimetric II	Any Analyser (Automation / Semi Automation)	703	3.28	10.79	0.35	3.19 g/dl	-0.25	0.03
7	URIC ACID II	Enzymatic / Uricase Colorimetric II	Any Analyser (Automation / Semi Automation)	756	5.04	22.70	1.14	7.41 mg/dl	2.08	0.08
8	CHOLESTROL II	CHOD-PAP II	Any Analyser (Automation / Semi Automation)	906	109.30	12.80	13.99	119 mg/dl	0.69	0.93
9	TRIGLYCERIDE II	GPO-PAP / Enzymatic Colorimetric / End Point II	Any Analyser (Automation / Semi Automation)	774	139.56	14.86	20.74	131 mg/dl	-0.41	1.49

SDI Range	Interpretation
Within -1.0 to +1.0	Excellent.
Between $\pm 1.0$ to $\pm 2.0$	Good.
Between $\pm 2.0$ to $\pm 3.0$	Accept with caution. Warning Signal.
Beyond $\pm 3.0$	Unacceptable performance. Action Signal.

Homogeneity and Stability of the sample is passed.

Data in CMC EQAS reports is confidential

Contact details:

Email: clinqc@cmcvellore.ac.in

Contact Number: 0416-2283102

Dr. Pamela Christudoss

CMC EQAS Co-Ordinator

Christian Medical College, Vellore

\*\*\*\*\* End of Report \*\*\*\*\*