



# Multitech Technofab Enterprises

475, First Floor, F.I.E., Patparganj Industrial Area  
Delhi - 110092

Phone No. : 09810943334 011-41764006, Email:  
multitech.technofab@gmail.com, mtedelhi2020@gmail.com,  
Web. : www.multitechtechnofab.com



## CALIBRATION CERTIFICATE

Calibration Certificate No.: MTE/204/CC/1601		Certificate Issued on: 02-08-2021				
Job / Identification No. : MTE/CAL/JOB/2464		Date of Calibration : 31-07-2021				
Item Receipt Date: 30-07-2021		Recommended Date of Next Calibration: 31-07-2022				
Calibration Certificate for : Micropipette		Item Location :				
CALIBRATED FOR	YASH DIAGNOSTIC CENTRE 702,VISHESHARPUR,PACHHATIA, Azamgarh Rd, , JAUNPUR, Uttar Pradesh, 222001		Reference: Order thru Telephonic			
DISCIPLINE	Mechanical	GROUP	Volume			
Item's Description						
Name of the Item	Make / Model	Serial / I.D. No.	Working / Inst. Range	Least Count	Unit	Item Condition
Micropipette	DR PERFECT / NA	NA / YASH/MP/02	10 to 100	1	µL	Working
ENVIRONMENTAL CONDITION	Temperature (in °C):	25.4 ± 2 °C	Humidity (in % RH):	52 ± 10% RH	Atm. Pressure in (hPa)	992
STANDARD REFERENCE DETAILS Traceable to National / International Standard						
Name of the Standard	Make / Model	Certificate No.	Calibration Agency / Traceability	Date of Calibration	Valid up to	
Semi Micro Balance	DANWER / DWA265	BBT/091/JAN/21	BLUEBOX TECHNO	07-01-2021	07-01-2022	
PRINCIPLE/METHODOLOGY OF CALIBRATION:		Lab Procedure No.: MTE-2/M/2.9		St. Procedure No.: ISO 8655-6		

RESULTS: In-Lab Calibration  
All Readings are in µL, Specified Otherwise

U.U.C. READING	Volume @ 27°(ml) Chanel 1	UNCERTAINTY <sup>^</sup>
10	0.010054	± 0.06
50	0.05018	± 0.06
100	0.100431	± 0.06

<sup>^</sup>The reported uncertainty in measurement is stated as the standard uncertainty multiplied by the coverage factor k = 2, which for a normal distribution corresponds to a coverage probability of approximately 95% confidence level.

### REMARKS:

- The recommended date for next calibration is mentioned, as asked by the customer.

### Note

- This report is not to be reported wholly or in part and cannot be used as evidence in any court of law and not to be used in any advertising media without our special permission in writing.
- The result listed refers only to the calibrated samples and applicable parameters Endorsement of products is neither inferred nor implied.
- Total liability of our organization is limited to the invoiced amount.
- Samples will be destroyed after one month from the date of issue of calibration Certificate unless otherwise specified.
- In case any reconfirmation of contents of Calibration Certificate is required, please contact our office

\*\*\*\*\* END OF REPORT \*\*\*\*\*

ULR : CC322521000001054F

Calibrate By  
(Parveen Kumar) , parveen  
Deputy Technical Manager

For and on behalf of  
**Multitech Technofab Enterprises**

Laboratory Head