

VIEW LAB MONTHLY SUMMARY

Lab Name	SRI KRISHNA CLINICAL LABORATORY	Details About Robust Analysis
Lab No	14431	Detail About Monthly Summary
Month	February	Detail about SDI
Year	2021	
Constituent Group	Chemistry II	

Click on the analyte to view Graphical Data

Date of Result Entered : 09/02/2021

Date of Report Published : 06/03/2021

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE II	GOD-POD II	Transasia / Erba	747	251.19	10.64	26.72	250 mg/dl	-0.04	1.95
2	UREA II	Urease UV / GLDH II	Transasia / Erba	584	54.62	11.45	6.25	56 mg/dl	0.22	0.52
3	CREATININE II	Jaffes Kinetic-Alkaline Picrate II	Transasia / Erba	452	2.80	10.81	0.30	2.9 mg/dl	0.33	0.03
4	T.BILIRUBIN II	Diazonium Salt (Colorimetric) / Jendrassik II	Transasia / Erba	532	1.22	19.52	0.24	1.55 mg/dl	1.39	0.02
5	T-PROTEIN II	Biuret - Colorimetric II	Transasia / Erba	580	5.56	10.15	0.56	6 g/dl	0.78	0.05
6	ALBUMIN II	BCG - Colorimetric II	Transasia / Erba	570	3.31	9.68	0.32	3.8 g/dl	1.53	0.03
7	URIC ACID II	Enzymatic / Uricase Colorimetric II	Transasia / Erba	576	3.83	16.03	0.61	4.3 mg/dl	0.77	0.05
8	CHOLESTROL II	CHOD-PAP II	Transasia / Erba	655	110.72	11.41	12.63	127 mg/dl	1.29	0.99
9	TRIGLYCERIDE II	GPO-PAP / Enzymatic Colorimetric / End Point II	Agape	172	94.70	13.24	12.54	105 mg/dl	0.82	1.91

SDI Range	Interpretation
Within -1.0 to +1.0	Excellent.
Between ± 1.0 to ± 2.0	Good.
Between ± 2.0 to ± 3.0	Accept with caution. Warning Signal.
Beyond ± 3.0	Unacceptable performance. Action Signal.

Links

Modify Address	Modify Methods	Submit Lab Results	View Lab Monthly Summary
Yearly Summary	Change Password	Home	Sign-out

LAB MONTHLY SUMMARY



Lab Name

SRI KRISHNA CLINICAL LABORATORY

Lab No

14431

Month

March

Year

2021

Constituent Group Chemistry II

Certificate No. PC/1924
NABL PT Accredited Laboratory

Date of Result Entered : 21/03/2021

Date of Report Published : 01/04/2021

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants CV	SD	Your Value	SDI	U
1	GLUCOSE II	GOD-POD II	Transasia / Erba	791	160.91	10.47	16.84	158 mg/dl	-0.17	1.20
2	UREA II	Urease UV / GLDH II	Transasia / Erba	605	42.26	12.78	5.40	42 mg/dl	-0.05	0.44
3	CREATININE II	Jaffes Kinetic- Alkaline Picrate II	Transasia / Erba	491	1.67	11.16	0.19	1.5 mg/dl	-0.91	0.02
4	T.BILIRUBIN II	Diazonium Salt (Colorimetric) / Jendrassik II	Transasia / Erba	658	3.60	13.31	0.48	3.3 mg/dl	-0.63	0.04
5	T-PROTEIN II	Biuret - Colorimetric II	Transasia / Erba	595	5.89	9.22	0.54	5.9 g/dl	0.02	0.04
6	ALBUMIN II	BCG - Colorimetric II	Transasia / Erba	599	3.52	9.28	0.33	3.4 g/dl	-0.37	0.03
7	URIC ACID II	Enzymatic / Urlicase Colorimetric II	Transasia / Erba	668	7.46	13.58	1.01	7.2 mg/dl	-0.26	0.08
8	CHOLESTROL II	CHOD-PAP II	Transasia / Erba	717	118.97	12.88	15.32	120 mg/dl	0.07	1.14
9	TRIGLYCERIDE II	GPO-PAP / Enzymatic Colorimetric / End Point II	Agape	182	112.18	16.15	18.12	105 mg/dl	-0.40	2.69

SDI Range	Interpretation
Within -1.0 to +1.0	Excellent.
Between ±1.0 to ±2.0	Good.
Between ±2.0 to ±3.0	Accept with caution. Warning Signal.
Beyond ±3.0	Unacceptable performance. Action Signal.

Homogeneity and Stability of the sample is passed.

Data in CMC EQAS reports is confidential

Contact details:

Email:clinqc@cmcvellore.ac.in

Contact Number: 0416-2283102

LAB MONTHLY SUMMARY



Lab Name

SRI KRISHNA CLINICAL LABORATORY

Lab No

14431

Month

April

Year

2021

Constituent Group

Chemistry II



National Reference Laboratory

Date of Result Entered : 21/04/2021

Date of Report Published : 01/05/2021

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants CV	SD	Your Value	SDI	U
1	GLUCOSE II	GOD-POD II	Transasia / Erba	713	318.19	10.10	32.15	323 mg/dl	0.15	2.41
2	UREA II	Urease UV / GLDH II	Transasia / Erba	522	27.05	13.69	3.70	26 mg/dl	-0.28	0.32
3	CREATININE II	Jaffes Kinetic- Alkaline Picrate II	Transasia / Erba	468	6.55	11.06	0.72	7 mg/dl	0.62	0.07
4	T.BILIRUBIN II	Diazonium Salt (Colorimetric) / Jendrassik II	Transasia / Erba	676	3.28	15.32	0.50	3.4 mg/dl	0.24	0.04
5	T-PROTEIN II	Biuret - Colorimetric II	Transasia / Erba	537	5.61	8.16	0.46	5.9 g/dl	0.63	0.04
6	ALBUMIN II	BCG - Colorimetric II	Transasia / Erba	550	3.32	9.16	0.30	3.36 g/dl	0.13	0.03
7	URIC ACID II	Enzymatic / Uricase Colorimetric II	Transasia / Erba	601	4.45	17.19	0.76	4.5 mg/dl	0.07	0.06
8	CHOLESTROL II	CHOD-PAP II	Transasia / Erba	677	113.01	11.66	13.18	108 mg/dl	-0.38	1.01
9	TRIGLYCERIDE II	GPO-PAP / Enzymatic Colorimetric / End Point II	Agape	181	225.08	10.11	22.75	222 mg/dl	-0.14	3.38

SDI Range	Interpretation
Within -1.0 to +1.0	Excellent.
Between ±1.0 to ±2.0	Good.
Between ±2.0 to ±3.0	Accept with caution. Warning Signal.
Beyond ±3.0	Unacceptable performance. Action Signal.

Homogeneity and Stability of the sample is passed.

Data in CMC EQAS reports is confidential

Contact details:

Email: clinqc@cmcvellore.ac.in

Contact Number: 0416-2283102

LAB MONTHLY SUMMARY



Lab Name **SRI KRISHNA CLINICAL LABORATORY** Lab No **14431**
 Month **May** Year **2021**
 Constituent Group **Chemistry II**

Certificate No. PC-1724
 NABL PT Accredited Laboratory

Date of Result Entered : 18/05/2021

Date of Report Published : 05/06/2021

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants CV	SD	Your Value	SDI	U
1	GLUCOSE II	GOD-POD II	Transasia / Erba	670	132.95	10.22	13.59	129 mg/dl	-0.29	1.05
2	UREA II	Urease UV / GLDH II	Transasia / Erba	562	63.66	12.61	8.03	63 mg/dl	-0.08	0.68
3	CREATININE II	Jaffes Kinetic- Alkaline Picrate II	Transasia / Erba	457	1.89	11.86	0.22	1.9 mg/dl	0.04	0.02
4	T.BILIRUBIN II	Diazonium Salt (Colorimetric) / Jendrassik II	Transasia / Erba	634	2.47	17.92	0.44	2.7 mg/dl	0.52	0.04
5	T-PROTEIN II	Biuret - Colorimetric II	Transasia / Erba	588	5.74	9.09	0.52	5.8 g/dl	0.11	0.04
6	ALBUMIN II	BCG - Colorimetric II	Transasia / Erba	567	3.39	9.40	0.32	3.54 g/dl	0.47	0.03
7	URIC ACID II	Enzymatic / Uricase Colorimetric II	Transasia / Erba	636	7.08	13.17	0.93	6.6 mg/dl	-0.52	0.07
8	CHOLESTROL II	CHOD-PAP II	Transasia / Erba	671	109.08	12.80	13.96	123 mg/dl	1.00	1.08
9	TRIGLYCERIDE II	GPO-PAP / Enzymatic Colorimetric / End Point II	Transasia / Erba	615	208.59	12.05	25.13	216 mg/dl	0.29	2.03

SDI Range	Interpretation
Within -1.0 to +1.0	Excellent.
Between ±1.0 to ±2.0	Good.
Between ±2.0 to ±3.0	Accept with caution. Warning Signal.
Beyond ±3.0	Unacceptable performance. Action Signal.

Homogeneity and Stability of the sample is passed.

Data in CMC EQAS reports is confidential
 Contact details:
 Email: clinqc@cmcvellore.ac.in
 Contact Number: 0416-2283102

VIEW LAB MONTHLY SUMMARY

Lab Name	SRI KRISHNA CLINICAL LABORATORY	Details About Robust Analysis
Lab No	14431	Detail About Monthly Summary
Month	June	Detail about SDI
Year	2021	
Constituent Group	Chemistry II	

[Click on the analyte to view Graphical Data](#)

[All Analyser Result](#) [Print](#)

Date of Result Entered : 22/06/2021

Date of Report Published : 03/07/2021

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants CV	Participants SD	Your Value	SDI	U
1	GLUCOSE II	GOD-POD II	Transasia / Erba	725	197.93	10.20	20.20	175 mg/dl	-1.14	1.50
2	UREA II	Urease UV / GLDH II	Transasia / Erba	579	76.33	12.48	9.52	73 mg/dl	-0.35	0.79
3	CREATININE II	Jaffes Kinetic-Alkaline Picrate II	Transasia / Erba	460	4.43	12.53	0.56	4.7 mg/dl	0.49	0.05
4	T.BILIRUBIN II	Diazonium Salt (Colorimetric) / Jendrassik II	Transasia / Erba	649	3.70	13.01	0.48	3.9 mg/dl	0.41	0.04
5	T-PROTEIN II	Biuret - Colorimetric II	Transasia / Erba	595	5.61	10.40	0.58	5.9 g/dl	0.50	0.05
6	ALBUMIN II	BCG - Colorimetric II	Transasia / Erba	597	3.32	9.58	0.32	3.4 g/dl	0.25	0.03
7	URIC ACID II	Enzymatic / Uricase Colorimetric II	Transasia / Erba	621	4.70	17.49	0.82	5 mg/dl	0.36	0.07
8	CHOLESTROL II	CHOD-PAP II	Transasia / Erba	692	105.54	13.56	14.31	110 mg/dl	0.31	1.09
9	TRIGLYCERIDE II	GPO-PAP / Enzymatic Colorimetric / End Point II	Transasia / Erba	649	123.44	17.80	21.98	132 mg/dl	0.39	1.73

SDI Range	Interpretation
Within -1.0 to +1.0	Excellent.
Between ± 1.0 to ± 2.0	Good.
Between ± 2.0 to ± 3.0	Accept with caution. Warning Signal.
Beyond ± 3.0	Unacceptable performance. Action Signal.

Links

Modify Address	Modify Methods	Submit Lab Results	View Lab Monthly Summary
Yearly Summary	Change Password	Home	Sign-out

VIEW LAB MONTHLY SUMMARY

Lab Name	SRI KRISHNA CLINICAL LABORATORY	Details About Robust Analysis									
Lab No	14431	Detail About Monthly Summary									
Month	July	Detail about SDI									
Year	2021										
Constituent Group	Chemistry II										
Click on the analyte to view Graphical Data										All Analyser Result	Print
										Date of Result Entered : 21/07/2021	
										Date of Report Published : 02/08/2021	
Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants CV	SD	Your Value	SDI	U	
1	GLUCOSE II	GOD-POD II	Transasia / Erba	751	148.33	10.85	16.09	142 mg/dl	-0.39	1.17	
2	UREA II	Urease UV / GLDH II	Transasia / Erba	570	32.47	13.36	4.34	36 mg/dl	0.81	0.36	
3	CREATININE II	Jaffes Kinetic-Alkaline Picrate II	Transasia / Erba	479	2.77	11.56	0.32	2.9 mg/dl	0.41	0.03	
4	T.BILIRUBIN II	Diazonium Salt (Colorimetric) / Jendrassik II	Transasia / Erba	655	2.84	12.80	0.36	2.8 mg/dl	-0.11	0.03	
5	T-PROTEIN II	Biuret - Colorimetric II	Transasia / Erba	609	5.55	8.57	0.48	5.8 g/dl	0.53	0.04	
6	ALBUMIN II	BCG - Colorimetric II	Transasia / Erba	607	3.32	8.78	0.29	3.47 g/dl	0.52	0.02	
7	URIC ACID II	Enzymatic / Uricase Colorimetric II	Transasia / Erba	666	4.95	17.29	0.86	4.8 mg/dl	-0.18	0.07	
8	CHOLESTROL II	CHOD-PAP II	Transasia / Erba	783	112.10	13.65	15.31	116 mg/dl	0.25	1.09	
9	TRIGLYCERIDE II	GPO-PAP / Enzymatic Colorimetric / End Point II	Transasia / Erba	666	167.18	13.26	22.17	178 mg/dl	0.49	1.72	
SDI Range		Interpretation									
Within -1.0 to +1.0		Excellent.									
Between ±1.0 to ±2.0		Good.									
Between ±2.0 to ±3.0		Accept with caution. Warning Signal.									
Beyond ±3.0		Unacceptable performance. Action Signal.									

Links

Modify Address	Modify Methods	Submit Lab Results	View Lab Monthly Summary
Yearly Summary	Change Password	Home	Sign-out