

LAB MONTHLY SUMMARY



Lab Name **YOU TAKE CARE WELLNESS CLINIC & DIAGNOSTIC CENTRE (P) LTD.** Lab No **5953**
 Month **July** Year **2021**
 Constituent Group **Chemistry II**



Certificate No. PL-1024
 NABL PT Accredited Laboratory

Date of Result Entered : 19/07/2021

Date of Report Published : 02/08/2021

Sl No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE II	HEXOKINASE II	Any Analyser (Automation / Semi Automation)	39	151.71	9.61	14.59	164 mg/dl	0.84	4.67
2	UREA II	Urease UV / GLDH II	CPC (Turbochem 100 / 240 Prime	57	32.82	13.34	4.38	34.6 mg/dl	0.41	1.16
3	CREATININE II	Jaffes Kinetic- Alkaline Picrate II	CPC (Turbochem 100 / 240 Prime	57	2.64	12.35	0.33	2.84 mg/dl	0.61	0.09
4	T.BILIRUBIN II	Diazonium Salt (Colorimetric) / Jendrassik II	CPC (Turbochem 100 / 240 Prime	66	2.90	10.40	0.30	2.92 mg/dl	0.07	0.07
5	ALBUMIN II	BCG - Colorimetric II	CPC (Turbochem 100 / 240 Prime	68	3.32	9.06	0.30	3.22 g/dl	-0.33	0.07
6	CHOLESTROL II	CHOD-PAP II	CPC (Turbochem 100 / 240 Prime	75	111.94	13.22	14.80	95.2 mg/dl	-1.13	3.42
7	TRIGLYCERIDE II	GPO-PAP / Enzymatic Colorimetric / End Point II	CPC (Turbochem 100 / 240 Prime	67	176.62	8.30	14.66	180.4 mg/dl	0.26	3.58

SDI Range	Interpretation
Within -1.0 to +1.0	Excellent.
Between ±1.0 to ±2.0	Good.
Between ±2.0 to ±3.0	Accept with caution. Warning Signal.
Beyond ±3.0	Unacceptable performance. Action Signal.

Homogeneity and Stability of the sample is passed.

Data in CMC EQAS reports is confidential

Contact details:

Email: clinqc@cmcvellore.ac.in

Contact Number: 0416-2283102

Pamela Christudoss

Dr. Pamela Christudoss

CMC EQAS Co-Ordinator

Christian Medical College, Vellore

***** End of Report *****

*July
31/8/2021*