



## Calibration Certificate, Semi Automatic Urine Strip Reader

|          |   |
|----------|---|
| Name     | URIPLUS – 200 ( SEMIAUTO URINE STRIP READER )                     |
| Hospital | YASH DIAGNOSTIC CENTER , VISHESHARPUR PACHAHATIYA U.P. – (222001) |
| Date     | 20 July 2021  |

This is certify that URIPLUS-200 (SEMIAUTO URINE STRIP READER) have calibrated & produced following data..


### Action Details:-

- INTERNAL QC STRIP - PASS
- BLANK STRIP READING - NIL
- Water Blank Readings - Negative
- Set Scaller Value - Within Range

| Instrument identification |             |                         |            | QC Passed |     |
|---------------------------|-------------|-------------------------|------------|-----------|-----|
| Serial no.                |             | Date of Calibration     | 20.07.2021 | Yes       | Yes |
| Model                     | URIPLUS-200 | Due Date of Calibration | 19.07.2022 | Yes       | Yes |

After calibration Q.C. Test passed successfully.

|                           |   |         |      |
|---------------------------|---|---------|------|
| Test results is coming ok | ✓ | Remarks | Done |
|---------------------------|---|---------|------|

  
 Service Engineer  
 20/07/21  
 Area Manager  
 Aspen Diagnostics Pvt. Ltd.  
 Delhi-110033