

LAB MONTHLY SUMMARY



Lab Name **ANNAI DIAGNOSTIC LAB** Lab No **13696**
 Month **April** Year **2021**
 Constituent Group **Chemistry II**



Date of Result Entered : 18/04/2021

Date of Report Published : 01/05/2021

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE II	GOD-POD II	Any Analyser (Automation / Semi Automation)	879	310.68	9.27	28.80	294.62 mg/dl	-0.56	1.94
2	UREA II	Others (Urease Berthelot / Ned Dye) II	Any Analyser (Automation / Semi Automation)	847	30.12	15.52	4.68	24.06 mg/dl	-1.30	0.32
3	CREATININE II	Jaffes Kinetic-Alkaline Picrate II	Any Analyser (Automation / Semi Automation)	718	6.46	12.06	0.78	6.23 mg/dl	-0.29	0.06
4	T.BILIRUBIN II	Diazonium Salt (Colorimetric) / Jendrassik II	Any Analyser (Automation / Semi Automation)	791	3.42	17.30	0.59	3.52 mg/dl	0.17	0.04
5	T-PROTEIN II	Biuret - Colorimetric II	Any Analyser (Automation / Semi Automation)	680	5.50	8.70	0.48	4.41 g/dl	-2.27	0.04
6	ALBUMIN II	BCG - Colorimetric II	Any Analyser (Automation / Semi Automation)	685	3.30	9.80	0.32	3.21 g/dl	-0.28	0.02
7	URIC ACID II	Enzymatic / Uricase Colorimetric II	Any Analyser (Automation / Semi Automation)	690	4.59	19.39	0.89	4.04 mg/dl	-0.62	0.07
8	CHOLESTROL II	CHOD-PAP II	Any Analyser (Automation / Semi Automation)	944	117.93	11.98	14.13	102.48 mg/dl	-1.09	0.92
9	TRIGLYCERIDE II	GPO-PAP / Enzymatic Colorimetric / End Point II	Any Analyser (Automation / Semi Automation)	790	224.07	12.54	28.10	176 mg/dl	-1.71	2.00

SDI Range	Interpretation
Within -1.0 to +1.0	Excellent.
Between ± 1.0 to ± 2.0	Good.
Between ± 2.0 to ± 3.0	Accept with caution. Warning Signal.
Beyond ± 3.0	Unacceptable performance. Action Signal.

Homogeneity and Stability of the sample is passed.

Data in CMC EQAS reports is confidential

Contact details:

Email: clinqc@cmcvellore.ac.in

Contact Number: 0416-2283102

Dr. Pamela Christudoss
 CMC EQAS Co-Ordinator
 Christian Medical College, Vellore

***** End of Report *****