

## LAB MONTHLY SUMMARY



Lab Name **ANNAI DIAGNOSTIC LAB** Lab No **13696**  
 Month **February** Year **2021**  
 Constituent Group **Chemistry II**



Date of Result Entered : 23/02/2021

Date of Report Published : 06/03/2021

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE II	GOD-POD II	Any Analyser (Automation / Semi Automation)	940	245.52	10.28	25.24	218.96 mg/dl	-1.05	1.65
2	UREA II	Others ( Urease Berthelot / Ned Dye ) II	Any Analyser (Automation / Semi Automation)	856	53.59	14.85	7.96	58.98 mg/dl	0.68	0.54
3	CREATININE II	Jaffes Kinetic-Alkaline Picrate II	Any Analyser (Automation / Semi Automation)	671	2.79	11.59	0.32	2.83 mg/dl	0.12	0.02
4	T.BILIRUBIN II	Diazonium Salt ( Colorimetric ) / Jendrassik II	Any Analyser (Automation / Semi Automation)	666	1.32	19.65	0.26	0.9 mg/dl	-1.62	0.02
5	T-PROTEIN II	Biuret - Colorimetric II	Any Analyser (Automation / Semi Automation)	732	5.47	10.41	0.57	4.59 g/dl	-1.54	0.04
6	ALBUMIN II	BCG - Colorimetric II	Any Analyser (Automation / Semi Automation)	689	3.25	9.72	0.32	3.24 g/dl	-0.03	0.02
7	URIC ACID II	Enzymatic / Uricase Colorimetric II	Any Analyser (Automation / Semi Automation)	582	3.94	17.52	0.69	3.39 mg/dl	-0.80	0.06
8	CHOLESTROL II	CHOD-PAP II	Any Analyser (Automation / Semi Automation)	884	114.21	11.24	12.84	111.3 mg/dl	-0.23	0.86
9	TRIGLYCERIDE II	GPO-PAP / Enzymatic Colorimetric / End Point II	Any Analyser (Automation / Semi Automation)	754	100.19	15.39	15.42	101.69 mg/dl	0.10	1.12

SDI Range	Interpretation
Within -1.0 to +1.0	Excellent.
Between $\pm 1.0$ to $\pm 2.0$	Good.
Between $\pm 2.0$ to $\pm 3.0$	Accept with caution. Warning Signal.
Beyond $\pm 3.0$	Unacceptable performance. Action Signal.

Homogeneity and Stability of the sample is passed.

Data in CMC EQAS reports is confidential

Contact details:

Email: clinqc@cmcvellore.ac.in

Contact Number: 0416-2283102

Dr. Pamela Christudoss  
 CMC EQAS Co-Ordinator  
 Christian Medical College, Vellore

\*\*\*\*\* End of Report \*\*\*\*\*