

Calibration Certificate

Page No. - 1 of 1

Certificate No.	MACL-2021-H-79-22	Calibration Certificate of HEMATOLOGY ANALYZER
Issue Date	3-05-2021	
Service Request No.	MACL-2021-H-79	
Service Request Date	29-05-2021	
Calibration Date	1-05-2021	
Calibration Due Date (As per Customer request)	30-04-2022	

Name & Address of Customer	Identification on UUC (Unit Under Calibration)	
Vijay Nursing Home F-24/25, Sector-3, Rohini, North West, Delhi-110085	Range/ Size	As Per Instrument
	Make / Model No.	ACCUREX/ ACCU LAB
	Instrument I.D. No.	VNH/HEMATOLOGY/LAB/01
	Instrument Sr. No.	2900pet02683
	Least Count	N/A
	Location	---
	Visual Inspection	OK

Calibration Procedure	MACL/H/CP-01 to 25	Condition of Instrument	OK
Reference Standard	-----	Calibration Performed at	Site
Environmental Condition	Temperature	25 ± 5 °C	Relative Humidity
			50 ± 20 % RH

Standard Equipment Used for Calibration					
Instrument Name	Make	I.D No./Sr.No.	Calibrated By	Certificate No.	Due Date
Digital Process Calibrator	TIPL	MACL/ET/PC-01	Acumen Measurement Delhi	AMC/00001233	25-06-2021


Calibration Results

Visual Inspection :- OK
Result :- The Calibration Result are found Satisfactory & Instrument is fit for use.

Uncertainty of Measurement	The reported uncertainty stated is the expanded uncertainty of measurement obtained by multiplying the standard uncertainty by the coverage factor k=2 correspond to confidence level of 95 %
± 1 %	


- Note :-**
- This Certificate refers only to the particular items submitted for calibration.
 - This Certificate shall not be reproduced, except in full, without the written permission of C E O Micro Advance Calibration Lab Ghaziabad.
 - Result Reported are Valid at the of and under the stated conditions of measurement.
 - Laboratory Standards are traceable to National Standards.
 - Calibration Certificate issue for weight & Measure parameters i.e. Mass, Balance, Volumetric equipment, Measuring Scale/Tapes etc. for scientific purpose only and should not be used for Trade / Commercial use.

CALIBRATED BY


Uma Shanker Sharma
Calibration Engineer



APPROVED BY


Mukesh Kumar Shakya
Authorized Signatory