

Calibration Certificate

Page No - 1 of 1

Certificate No.	MACL-2021-H-79-18	Calibration Certificate of Biochemistry Analyzer
Issue Date	3-05-2021	
Service Request No.	MACL-2021-H-79	
Service Request Date	29-05-2021	
Calibration Date	1-05-2021	
Calibration Due Date (As per Customer request)	30-04-2022	

Name & Address of Customer	Identification on UUC (Unit Under Calibration)	
Vijay Nursing Home F-24/25, Sector-3, Rohini, North West, Delhi-110085	Range/ Size Make / Model No. Instrument I.D. No. Instrument Sr. No. Least Count Location Visual Inspection	As Per Instrument Micro Lab Inshine/RX 50V VNH/BIO=CHEM/LAB/01 1711963 N/A --- OK

Calibration Procedure	MACL/H/CP-01 to 25	Condition of Instrument	OK
Reference Standard	-----	Calibration Performed at	Site
Environmental Condition	Temperature	25 ± 5°C	Relative Humidity 50 ± 20 % RH

Standard Equipment Used for Calibration					
Instrument Name	Make	I.D No./Sr.No.	Calibrated By	Certificate No.	Due Date
Digital Process Calibrator	TIPL	MACL/ET/PC-01	Acumen Measurement Delhi	AMC/00001233	25-06-2021

Calibration Results


Visual Inspection :- OK
 Result :- The Calibration Result are found Satisfactory & Instrument is fit for use.


Uncertainty of Measurement ± 1 % The reported uncertainty stated is the expanded uncertainty of measurement obtained by multiplying the standard uncertainty by the coverage factor k=2 correspond to confidence level of 95 %

Note :-


1. This Certificate refers only to the particular items submitted for calibration.
2. This Certificate shall not be reproduced, except in full, without the written permission of C E O Micro Advance Calibration Lab Ghaziabad.
3. Result Reported are Valid at the of and under the stated conditions of measurement
4. Laboratory Standards are traceable to National Standards.
5. Calibration Certificate issue for weight & Measure parameters i.e Mass. Balance. Volumetric equipment, Measuring Scale/Tapes etc. for scientific purpose only and should not be used for Trade / Commercial use.

CALIBRATED BY


 Uma Shanker Sharma
 Calibration Engineer



APPROVED BY


 Mukesh Kumar Shakya
 Authorized Signatory