



Date : 25/01/2021

## CALIBRATION CERTIFICATE

SRI KRISHNA CLINICAL LABORATORY  
NO 1/1, Karungalpatti, 2<sup>nd</sup> street  
GUGAI  
SALEM-636006

Dear Sir/ Madam ,

As desired by you please find the enclosed herewith the calibration report.

1. Serial No : S201032
2. Range of the instrument : Optical density Liner from 0-3.5 A
3. Precision & Accuracy study on CHEM-7 done on 25-01-2021.

This instrument is manufactured in accordance with the European law in force, in particular to "Electricity Safety" EN 610101 and Electromagnetic Compatibility " EN 61326 Standards.

WAVE LENGTH	GAIN
340	1,68000
405	1,53000
505	1,59600
546	1,60000
578	1,56000
600	1,80000
660	1,86000
700	1,76343



	IDEAL	ACTUAL
Lamp Voltage	12 V	12.1 V
Temperature	37 C	37 C

Photometer Calibration Check	OK
Cuvette Blank Check	OK

Calibration done on ; 25-01-2021  
Calibration due on ; 24-01-2022

The instrument does photometer check before each batch of tests is run. It will give an error message if the photometer test fails and it will not run any samples.

Thanking you and assuring of our prompt attention at all times.

Your faithfully,  
For Transasia Bio Medicals Ltd,

M. Shanmugasundaram  
Sr. Service Engineer