

HIRA CLINICAL LAB (441)

Outlier And Analyte Summary Report
Result Details For Cycle No 210207 and Sample No 03

Report Date : 19/07/2021

Instrument : Roche cobas c 311

Analyte	StandardUnit	Result Value	Mean	Z-Score	RMZ
@ Glucose	mg/dl	71.00	62.21	1.58	0.82
@ Urea	mg/dl	26.00	23.75	1.06	0.84
@ Creatinine	mg/dl	0.66	0.74	-1.05	-0.14
! Phosphorous Inorganic	mg/dl	4.30	3.75	2.87	1.21
@ Cholesterol	mg/dl	162.00	157.26	0.48	0.32
@ HDL	mg/dl	31.00	32.77	-0.78	-0.78

Legend @ : z score ≤ 2.0 - Acceptable

! : 2.0 < z score < 3.0 - Warning

X : z score ≥ 3 - Unacceptable

: Not Evaluated

⌚ : Delayed Result Entry

* : Not considered for evaluation.

Total Parameters	6
Not Evaluated Parameters	0
Evaluated Parameters	6
Outlier Parameters (X)	0
EQAS Score Biochemistry	100.00 %

PK Nigam

Dr Puneet Kumar Nigam
PT coordinator & Technical Manager, MHL EQAS

Unit No. 409-416,
Commercial Building - 1A,
Kohinoor Mall, Kiroli Road, Kuria (W),
Mumbai - 400070