

HIRA CLINICAL LAB (441)

Outlier And Analyte Summary Report

Result Details For Cycle No 210207 and Sample No 01

Report Date : 19/07/2021

Instrument : Roche cobas c 311

Analyte	StandardUnit	Result Value	Mean	Z-Score	RMZ
@ Bilirubin Total	mg/dl	0.18	0.19	-0.22	1.14
@ Bilirubin Direct	mg/dl	0.14	0.10	1.68	1.20
@ Total Protein	g/dL	6.70	6.50	0.89	1.36
@ Albumin	g/dL	4.30	3.88	1.93	1.26
@ Uric Acid	mg/dl	3.90	4.23	-0.80	0.14
@ Calcium	mg/dl	7.70	7.54	0.40	0.90
@ Triglycerides	mg/dl	175.00	155.13	1.46	0.85
@ ALT	U/L	11.00	13.30	-0.98	0.87
@ AST	U/L	24.00	23.54	0.50	0.71
@ ALP	U/L	78.00	77.82	0.20	-0.05
@ GGT	U/L	21.00	22.15	-0.31	-0.31
@ Amylase	U/L	75.00	68.50	0.78	0.63
@ Chloride	mmol/L	89.00	89.37	0.10	0.33
@ Sodium	mmol/L	126.00	128.12	0.25	0.32
@ Potassium	mmol/L	4.30	4.35	0.16	0.31

Legend @ : z score ≤ 2.0 - Acceptable

! : 2.0 < z score < 3.0 - Warning

X : z score ≥ 3 - Unacceptable

: Not Evaluated

⊙ : Delayed Result Entry

* : Not considered for evaluation.

Total Parameters	15
Not Evaluated Parameters	0
Evaluated Parameters	15
Outlier Parameters (X)	0
EQAS Score Biochemistry	100.00 %

PK Nigam

Dr Puneet Kumar Nigam
PT coordinator & Technical Manager, MHL EQAS
Unit No. 409-416,
Commercial Building - 1A,
Kohinoor Mall, Kiroli Road, Kurla (W),
Mumbai - 400070