

CERTIFICATE OF CALIBRATION

Ref. No.: MSM001
Applicant : MEDICITY SPECIALITY MEDICAL LABORATORY
Device Calibrated : REFRIGERATOR
Make/Model: SAMSUNG
Sl.No: ...

Condition at receipt : Functional
Calibrated at : Onsite
Date of Calibration : 10-01-2021
Calibration due date : 09-01-2022

Certified that the above Instrument has been calibrated by trained technical personnel of this centre using a calibration system traceable to National Standards. The calibration details attached with the certificate are authentic quantitative analysis of the instrument's credibility on the date of calibration and relate only to the Instrument calibrated.

Date of Issue : 10-01-2021

Approved Signatory



Cyrix Healthcare Pvt. Ltd.
30/641 B Petta Junction, Poonithura, Kochi-38
Ph : 97455 55073
Email: sales.kl@cyrix.in
service.kl@cyrix.in

CYRIXTM
HEALTH CARE PVT LTD

www.cyrixhealthcare.com

ISO Certified Company

AERB Approved
QA Agency

NABL Accredited
Calibration Agency

CERTIFICATE OF CALIBRATION

Customer Name & Address	MEDICITY SPECIALITY MEDICAL LABORATORY, THALIPARAMBA	Certificate No.	MSM001
		Date of Calibration	10-01-2021
		Valid Upto	09-01-2022
		Date of Issue	10-01-2021

Instruments Details			
Instrument Name	REFRIGERATOR	S No.	..
Make/Model No.	SAMSUNG	Location	LAB
ID No.	...	Visual Inspection	OK

Details of Reference Standards & Major Equipments Used			
Equipment Name	ELECTRICAL SAFETY ANALYSER	THERMO HYGROMETER	INFRARED THERMOMETER
Make	FLUKE	HTC-2	HTC
Model/SR No.	ESA612/2959012	...	IRX-64/20170804051
Calibration Validity	17/01/2021	22/09/2021	13/01/2021
Calibration By	STQC	NENO TECHNICAL SERVICES	PERFECT CALLIBRATION

Environmental Condition			
Temperature	27	Relative Humidity	54

Calibration Results				
SL NO	SET TEMPREATURE (°C)	MEASURED TEMPREATURE (°C)	TOLERANCE(°C)	STATUS
1	-23	-22.7	±5%	P
2	-19	-18.8	±5%	P
3	-15	-14.7	±5%	P
4	1	1.8	±5%	P
5	3	2.6	±5%	P
6	7	6.5	±5%	P

REMARKS:-

- ❖ This result of calibration refers to only to the particular items submitted for calibration.
- ❖ The above result is valid at the time of and under the stated conditions measurement.
- ❖ Calibration certificate issued for weight & measure parameters like mass, weighting balance, volumetric equipment, measuring scales/tapes etc. Are scientific purpose only and should not used for commercial.

Calibrated By
(SE)



Approved By
(QM)



Cyrix Healthcare Pvt. Ltd.,
30/641 B Petta Junction, Poonithura, Kochi-38.
Ph: 97455 55073
Email: sales.kl@cyrix.in
service.kl@cyrix.in

CYRUXTM
HEALTH CARE PVT LTD

www.cyrixhealthcare.com