

CERTIFICATE OF CALIBRATION

Customer Name & Address	MEDICITY SPECIALITY MEDICAL LABORATORY, THALIPARAMBA	Certificate No.	CHPL/PC/MSL/002
		Date of Calibration	10-01-2021
		Valid Upto	09-06-2021
		Date of Issue	10-01-2021

Instruments Details			
Instrument Name	MICRO PIPETTE	S No.	..
Make/Model No.	AVANTOR	Visual Inspection	OK

Details of Reference Standards & Major Equipments Used			
Equipment Name	ANALYTICAL BALANCE	THERMO HYGROMETER	
Make	WENSAR	FANEM	
Model/SR No.	MAB182/39736	MOD3620/RAH85394	
Calibration Validity	13-02-2021	25/09/2021	
Calibration By	STIC	FCRI PALAKKAD	

Environmental Condition			
Temperature	25.6 ^o C	Relative Humidity	40.1%

Calibration Results							
SL NO	SET READING (µl)	MEASURED READING (µl)					STATUS
		I	II	III	IV	V	
1	100	99.96	100.08	99.98	100.10	99.98	P
2	500	500.25	499.95	499.88	500.24	500.18	P
3	1000	1000.01	1000.00	1000.18	1000.48	1000.45	P

REMARKS:-

- ❖ This result of calibration refers to only to the particular items submitted for calibration.
- ❖ The above result is valid at the time of and under the stated conditions measurement.
- ❖ Calibration certificate issued for weight & measure parameters like mass, weighting balance, volumetric equipment, measuring scales/tapes etc. Are scientific purpose only and should not used for commercial.

Calibrated By
(SE)




Approved By
(QM)
SARATH K B



Cyrix Healthcare Pvt. Ltd.,
30/641 B Petta Junction, Poonithura, Kochi-38.
Ph: 97455 55073
Email: sales.kl@cyrix.in
service.kl@cyrix.in

CYRIX™

HEALTH CARE PVT LTD

www.cyrixhealthcare.com