

CERTIFICATE OF CALIBRATION

Customer Name & Address	MEDICITY SPECIALITY MEDICAL LABORATORY, THALIPARAMBA	Certificate No.	CHPL/PC/MSL/001
		Date of Calibration	10-01-2021
		Valid Upto	09-06-2021
		Date of Issue	10-01-2021

Instruments Details			
Instrument Name	MICRO PIPETTE	S No.	..
Make/Model No.	AVANTOR	Visual Inspection	OK

Details of Reference Standards & Major Equipments Used		
Equipment Name	ANALYTICAL BALANCE	THERMO HYGROMETER
Make	WENSAR	FANEM
Model/SR No.	MAB182/39736	MOD3620/RAH85394
Calibration Validity	13-02-2021	25/09/2021
Calibration By	STIC	FCRI PALAKKAD

Environmental Condition			
Temperature	25.6°C	Relative Humidity	40.1%

Calibration Results							
SL NO	SET READING (µl)	MEASURED READING (µl)					STATUS
		I	II	III	IV	V	
1	10	9.99	10.02	9.99	10.08	9.98	P
2	50	50.15	49.95	49.89	50.25	50.08	P
3	100	100.01	100.01	99.99	100.00	99.98	P

REMARKS:-

- ❖ This result of calibration refers to only to the particular items submitted for calibration.
- ❖ The above result is valid at the time of and under the stated conditions measurement.
- ❖ Calibration certificate issued for weight & measure parameters like mass, weighting balance, volumetric equipment, measuring scales/tapes etc. Are scientific purpose only and should not used for commercial.

Calibrated By




Approved By (QM)

SARATH K B



Cyrix Healthcare Pvt. Ltd.,
 30/641 B Petta Junction, Poonithura, Kochi-38.
 Ph: 97455 55073
 Email: sales.kl@cyrix.in
 service.kl@cyrix.in

CYRIX™

HEALTH CARE PVT LTD

www.cyrixhealthcare.com