

LAB MONTHLY SUMMARY



Lab Name **DR B LAL CLINICAL LABORATORY PVT LTD** Lab No **13557**
 Month **January** Year **2021**
 Constituent Group **HbA1c**



Date of Result Entered : 18/01/2021

Date of Report Published : 05/02/2021

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	HbA1c	OTHERS (Any Other Principles / Methods)	Any Analyser (Automation / Semi Automation)	110	5.84	10.06	0.59	4.7 %	-1.94	0.11

SDI Range	Interpretation
Within -1.0 to +1.0	Excellent.
Between ± 1.0 to ± 2.0	Good.
Between ± 2.0 to ± 3.0	Accept with caution. Warning Signal.
Beyond ± 3.0	Unacceptable performance. Action Signal.

Homogeneity and Stability of the sample is passed.

Data in CMC EQAS reports is confidential

Contact details:

Email: clinqc@cmcvellore.ac.in

Contact Number: 0416-2283102

Dr. Pamela Christudoss
 CMC EQAS Co-Ordinator
 Christian Medical College, Vellore

***** End of Report *****