

To Whom It May Concern:

CALIBRATION CERTIFICATE

This is to certify that the system maintenance & calibration was performed on date 08 Feb 2021 for the machine, VITROS-250 Chemistry Analyzer, installed at Dr B Lal Clinical Laboratory, Alwar bearing Serial No.27001297

The calibration process includes optics calibration and the performance verification tests as per the manufacturer's procedures.
Validity: Six months for Optics Calibration.

RAVI RANJAN

Service engineer

I
S
Ravi Ranjan

Ortho Clinical Diagnostics

Unit/Plot No - 810, ILD Trade Centre,
Sector - 47, Sohna Road, Gurgaon -122001

T: +91 124 4014 832

CALIBRATION REPORT

Name of Instrument : VITROS 250
Serial Number : 27001297
Installed at : apex Hospital
Date of Report : 08/02/2021

SECTION / TEST

RESULT

SAMPLE METERING

a. Leak Test : PASS / FAIL

SPECIFICATIONS	OBSERVATION
The Dispense Leak Rate must < 0.210 V dc	Pass
The read-to-read leak rate should be < 0.030 V dc	Pass

b. Hysteresis Test : PASS / FAIL

SPECIFICATIONS	OBSERVATION
The Maximum Hysteresis Value must be < 5.0 steps	Pass
Combined Hysteresis values of any 3 consecutive readings must be < 9.0 steps	Pass

Ravi Ranjan

Ortho Clinical Diagnostics India Private Limited

Registered Office: 403, Leela Business Park, AndheriKurla Road, Andheri East, Mumbai - 400059 | T: +91 22 6787 9300 | CIN: U51397MH2015FTC262650

AndheriKurla Road, Andheri East, Mumbai - 400059 | T: +91 22 6787 9300 | CIN: U51397MH2015FTC262650

REFERENCE METERING

a. Leak Test : PASS / FAIL

SPECIFICATIONS	OBSERVATION
The Aspirate Value < 0.175 V dc	Pass
The Ambient Pressure must in range of 2.25–2.75 V dc	Pass
Consecutive reading should not fall > 0.030 V dc	Pass

b. Hysteresis Test : PASS / FAIL

SPECIFICATIONS	OBSERVATION
Mean Value of Hysteresis should be ≤ 5.0 V dc	Pass
Maximum Value of Hysteresis should be ≤ 10.0 V dc	Pass

IMMUNO WASH METERING

a. Leak Test : PASS / FAIL

SPECIFICATIONS	OBSERVATION
The Aspirate Value < 0.175 V dc	Pass
The Ambient Pressure must in range of 2.25–2.75 V dc	Pass
Consecutive reading should not fall > 0.030 V dc	Pass

Ravi Ranjan

Ortho Clinical Diagnostics

Unit/Plot No - 810, ILD Trade Centre,
Sector - 47, Sohna Road, Gurgaon - 122001

T: +91 124 4014 832

b. Hysteresis Test : PASS / FAIL

Mean Value of Hysteresis should be ≤ 5.0 V dc	Pass
Maximum Value of Hysteresis should be ≤ 11.0 V dc	Pass

INCUBATOR

a. Pad Reflectance Test : PASS / FAIL

SPECIFICATIONS	OBSERVATION
Grand Mean Value must be in range of 0.75 – 0.99	Pass
Range must be ≤ 0.03	Pass

ELECTROMETER

a. Basic Mode Verification Test: PASS / FAIL

Parameters	Test 1	Test 2	Test 3	Test 4
Input	Test	Test	Test	Test
Source	Zero 0	Zero 0	Ref 1	Ref 1
Polarity	Neg 0	Neg 0	Neg 0	Neg 0
Ohm -	Low 0	High 1	Low 0	High 1
Ohm +	Low 0	High 1	Low 0	High 1
Read Type	One 0	One 0	One 0	One 0
Specifications	0.0 mv \pm 2.4mv	0.0 mv \pm 2.4mv	-200 mv \pm 8mv	-200 mv \pm 8mv

Ravi Ranjan

REFLECTOMETER

a. IRIS Adjustment : PASS / FAIL

SPECIFICATIONS	OBSERVATION
Lamp Output range 3.5 to 9.0 across all Filters	Pass

b. Static Test : PASS / FAIL

SPECIFICATIONS	OBSERVATION
All “%CV*K” must be ≤ 18.0	Pass
All “SD Dark” must be ≤ 1.3	Pass

c. Interleaved Static Test: PASS / FAIL

SPECIFICATIONS	OBSERVATION
All “%CV*K” must be ≤ 18.0	Pass
All “SD Dark” must be ≤ 1.3	Pass

d. Dynamic Test : PASS / FAIL

SPECIFICATIONS	OBSERVATION
All “%CV*K” must be ≤ 18.0	Pass
All “SD Dark” must be ≤ 1.3	Pass

Ravi Ranjan