

## LAB MONTHLY SUMMARY



Lab Name **DR REETU NAGE** Lab No **9324**  
 Month **February** Year **2021**  
 Constituent Group **Chemistry II**



Date of Result Entered : 15/02/2021

Date of Report Published : 06/03/2021

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE II	GOD-POD II	Transasia / Erba	747	251.19	10.54	26.72	274.3 mg/dl	0.33	1.55
2	UREA II	Urease UV / GLDH II	Transasia / Erba	584	54.62	11.45	6.25	50 mg/dl	-0.74	0.52
3	CREATININE II	Jaffes Kinetic-Alkaline Picrate II	Transasia / Erba	452	2.80	10.81	0.30	2.7 mg/dl	-0.33	0.03
4	T.BILIRUBIN II	Diazonium Salt (Colorimetric) / Jendrassik II	Transasia / Erba	532	1.22	19.52	0.24	1.18 mg/dl	-0.17	0.02
5	T-PROTEIN II	Biuret - Colorimetric II	Transasia / Erba	580	5.56	10.15	0.56	6.2 g/dl	1.13	0.05
6	ALBUMIN II	BCG - Colorimetric II	Transasia / Erba	570	3.31	9.68	0.32	3.1 g/dl	-0.66	0.03
7	URIC ACID II	Enzymatic / Uricase Colorimetric II	Transasia / Erba	576	3.83	16.03	0.61	3.5 mg/dl	-0.54	0.05
8	CHOLESTROL II	CHOD-PAP II	Transasia / Erba	655	110.72	11.41	12.63	92.4 mg/dl	-1.45	0.99
SDI Range		Interpretation								
Within -1.0 to +1.0		Excellent.								
Between ±1.0 to ±2.0		Good.								
Between ±2.0 to ±3.0		Accept with caution. Warning Signal.								
Beyond ±3.0		Unacceptable performance. Action Signal.								

Homogeneity and Stability of the sample is passed.

Data in CMC EQAS reports is confidential

Contact details:

Email:clinq@cmcvellore.ac.in

Contact Number: 0416-2283102

Dr. Pamela Christudoss  
 CMC EQAS Co-Ordinator  
 Christian Medical College, Vellore

\*\*\*\*\* End of Report \*\*\*\*\*

## LAB MONTHLY SUMMARY



Lab Name **DR REETU NAGE** Lab No **9324**  
 Month **March** Year **2021**  
 Constituent Group **Chemistry II**



Date of Result Entered : 23/03/2021

Date of Report Published : 01/04/2021

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE II	GOD-POD II	Transasia / Erba	791	160.91	10.47	16.84	161.7 mg/dl	0.05	1.20
2	UREA II	Urease UV / GLDH II	Transasia / Erba	605	42.26	12.78	5.40	45.4 mg/dl	0.58	0.44
3	CREATININE II	Jaffes Kinetic-Alkaline Picrate II	Transasia / Erba	491	1.67	11.16	0.19	1.72 mg/dl	0.72	0.02
4	T.BILIRUBIN II	Diazonium Salt ( Colorimetric ) / Jendrassik II	Transasia / Erba	658	3.60	13.31	0.48	4.3 mg/dl	1.46	0.04
5	T-PROTEIN II	Biuret - Colorimetric II	Transasia / Erba	595	5.89	9.22	0.54	5.5 g/dl	-0.72	0.04
6	ALBUMIN II	BCG - Colorimetric II	Transasia / Erba	599	3.52	9.28	0.33	3.5 g/dl	-0.06	0.03
7	URIC ACID II	Enzymatic / Uricase Colorimetric II	Transasia / Erba	668	7.46	13.58	1.01	6.9 mg/dl	-0.54	0.08
8	CHOLESTROL II	CHOD-PAP II	Transasia / Erba	717	118.97	12.88	15.32	110 mg/dl	-0.59	1.14
SDI Range		Interpretation								
Within -1.0 to +1.0		Excellent.								
Between $\pm 1.0$ to $\pm 2.0$		Good.								
Between $\pm 2.0$ to $\pm 3.0$		Accept with caution. Warning Signal.								
Beyond $\pm 3.0$		Unacceptable performance. Action Signal.								

Homogeneity and Stability of the sample is passed.

Data in CMC EQAS reports is confidential

Contact details:

Email: clinqc@cmcvellore.ac.in

Contact Number: 0416-2283102

Dr. Pamela Christudoss  
 CMC EQAS Co-Ordinator  
 Christian Medical College, Vellore

\*\*\*\*\* End of Report \*\*\*\*\*

## VIEW LAB MONTHLY SUMMARY

Lab Name	DR REETU NAGE	<a href="#">Details About Robust Analysis</a>
Lab No	9324	<a href="#">Detail About Monthly Summary</a>
Month	April	<a href="#">Detail about SDI</a>
Year	2021	
Constituent Group	Chemistry II	

[Click on the analyte to view Graphical Data](#)

Date of Result Entered : 17/04/2021

Date of Report Published : 01/05/2021

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	<a href="#">GLUCOSE II</a>	GOD-POD II	Transasia / Erba	713	318.19	10.10	32.15	334.6 mg/dl	0.51	2.41
2	<a href="#">UREA II</a>	Urease UV / GLDH II	Transasia / Erba	522	27.05	13.69	3.70	28.4 mg/dl	0.36	0.32
3	<a href="#">CREATININE II</a>	Jaffes Kinetic-Alkaline Picrate II	Transasia / Erba	468	6.55	11.06	0.72	6.52 mg/dl	-0.04	0.07
4	<a href="#">T.BILIRUBIN II</a>	Diazonium Salt ( Colorimetric ) / Jendrassik II	Transasia / Erba	676	3.28	15.32	0.50	3.76 mg/dl	0.96	0.04
5	<a href="#">T-PROTEIN II</a>	Biuret - Colorimetric II	Transasia / Erba	537	5.61	8.16	0.46	5.7 g/dl	0.20	0.04
6	<a href="#">ALBUMIN II</a>	BCG - Colorimetric II	Transasia / Erba	550	3.32	9.16	0.30	3.1 g/dl	-0.72	0.03
7	<a href="#">URIC ACID II</a>	Enzymatic / Uricase Colorimetric II	Transasia / Erba	601	4.45	17.19	0.76	4.2 mg/dl	-0.33	0.06
8	<a href="#">CHOLESTROL II</a>	CHOD-PAP II	Transasia / Erba	677	113.01	11.66	13.18	107 mg/dl	-0.46	1.01
<b>SDI Range</b>		<b>Interpretation</b>								
Within -1.0 to +1.0		Excellent.								
Between ±1.0 to ±2.0		Good.								
Between ±2.0 to ±3.0		Accept with caution. Warning Signal.								
Beyond ±3.0		Unacceptable performance. Action Signal.								

## Links

<a href="#">Modify Address</a>	<a href="#">Modify Methods</a>	<a href="#">Submit Lab Results</a>	<a href="#">View Lab Monthly Summary</a>
--------------------------------	--------------------------------	------------------------------------	--

<a href="#">Yearly Summary</a>	<a href="#">Change Password</a>	<a href="#">Home</a>	<a href="#">Sign-out</a>
--------------------------------	---------------------------------	----------------------	--------------------------



## VIEW LAB MONTHLY SUMMARY

Lab Name	DR REETU NAGE	<a href="#">Details About Robust Analysis</a>
Lab No	9324	<a href="#">Detail About Monthly Summary</a>
Month	May	<a href="#">Detail about SDI</a>
Year	2021	
Constituent Group	Chemistry II	

[Click on the analyte to view Graphical Data](#)

Date of Result Entered : 26/05/2021

Date of Report Published : 05/06/2021

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	<a href="#">GLUCOSE II</a>	GOD-POD II	Transasia / Erba	670	132.95	10.22	13.59	136.1 mg/dl	0.23	1.05
2	<a href="#">UREA II</a>	Urease UV / GLDH II	Transasia / Erba	562	63.66	12.61	8.03	67.3 mg/dl	0.45	0.68
3	<a href="#">CREATININE II</a>	Jaffes Kinetic-Alkaline Picrate II	Transasia / Erba	457	1.89	11.86	0.22	1.6 mg/dl	-1.29	0.02
4	<a href="#">T.BILIRUBIN II</a>	Diazonium Salt ( Colorimetric ) / Jendrassik II	Transasia / Erba	634	2.47	17.92	0.44	2.4 mg/dl	-0.16	0.04
5	<a href="#">T-PROTEIN II</a>	Biuret - Colorimetric II	Transasia / Erba	588	5.74	9.09	0.52	5.4 g/dl	-0.65	0.04
6	<a href="#">ALBUMIN II</a>	BCG - Colorimetric II	Transasia / Erba	567	3.39	9.40	0.32	3.5 g/dl	0.34	0.03
7	<a href="#">URIC ACID II</a>	Enzymatic / Uricase Colorimetric II	Transasia / Erba	636	7.08	13.17	0.93	7.2 mg/dl	0.13	0.07
8	<a href="#">CHOLESTROL II</a>	CHOD-PAP II	Transasia / Erba	671	109.08	12.80	13.96	101 mg/dl	-0.58	1.08
<b>SDI Range</b>		<b>Interpretation</b>								
Within -1.0 to +1.0		Excellent.								
Between ±1.0 to ±2.0		Good.								
Between ±2.0 to ±3.0		Accept with caution. Warning Signal.								
Beyond ±3.0		Unacceptable performance. Action Signal.								

## Links

<a href="#">Modify Address</a>	<a href="#">Modify Methods</a>	<a href="#">Submit Lab Results</a>	<a href="#">View Lab Monthly Summary</a>
--------------------------------	--------------------------------	------------------------------------	--

<a href="#">Yearly Summary</a>	<a href="#">Change Password</a>	<a href="#">Home</a>	<a href="#">Sign-out</a>
--------------------------------	---------------------------------	----------------------	--------------------------



## VIEW LAB MONTHLY SUMMARY

Lab Name	DR REETU NAGE	<a href="#">Details About Robust Analysis</a>
Lab No	9324	<a href="#">Detail About Monthly Summary</a>
Month	June	<a href="#">Detail about SDI</a>
Year	2021	
Constituent Group	Chemistry II	

[Click on the analyte to view Graphical Data](#)

Date of Result Entered : 20/06/2021

Date of Report Published : 03/07/2021

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	<a href="#">GLUCOSE II</a>	GOD-POD II	Transasia / Erba	725	197.93	10.20	20.20	200.4 mg/dl	0.12	1.50
2	<a href="#">UREA II</a>	Urease UV / GLDH II	Transasia / Erba	579	76.33	12.48	9.52	72.04 mg/dl	-0.45	0.79
3	<a href="#">CREATININE II</a>	Jaffes Kinetic-Alkaline Picrate II	Transasia / Erba	460	4.43	12.53	0.56	5.1 mg/dl	1.21	0.05
4	<a href="#">T.BILIRUBIN II</a>	Diazonium Salt ( Colorimetric ) / Jendrassik II	Transasia / Erba	649	3.70	13.01	0.48	4.4 mg/dl	1.45	0.04
5	<a href="#">T-PROTEIN II</a>	Biuret - Colorimetric II	Transasia / Erba	595	5.61	10.40	0.58	5.8 g/dl	0.33	0.05
6	<a href="#">ALBUMIN II</a>	BCG - Colorimetric II	Transasia / Erba	597	3.32	9.58	0.32	3.34 g/dl	0.04	0.03
7	<a href="#">URIC ACID II</a>	Enzymatic / Uricase Colorimetric II	Transasia / Erba	621	4.70	17.49	0.82	4 mg/dl	-0.85	0.07
8	<a href="#">CHOLESTROL II</a>	CHOD-PAP II	Transasia / Erba	692	105.54	13.56	14.31	123.1 mg/dl	1.23	1.09
<b>SDI Range</b>		<b>Interpretation</b>								
Within -1.0 to +1.0		Excellent.								
Between ±1.0 to ±2.0		Good.								
Between ±2.0 to ±3.0		Accept with caution. Warning Signal.								
Beyond ±3.0		Unacceptable performance. Action Signal.								

## Links

<a href="#">Modify Address</a>	<a href="#">Modify Methods</a>	<a href="#">Submit Lab Results</a>	<a href="#">View Lab Monthly Summary</a>
<a href="#">Yearly Summary</a>	<a href="#">Change Password</a>	<a href="#">Home</a>	<a href="#">Sign-out</a>





## LAB MONTHLY SUMMARY



Lab Name **DR REETU NAGE** Lab No **9324**  
 Month **July** Year **2021**  
 Constituent Group **Chemistry II**



Date of Result Entered : 17/07/2021

Date of Report Published : 02/08/2021

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE II	GOD-POD II	Transasia / Erba	751	148.33	10.85	16.09	137 mg/dl	-0.70	1.17
2	UREA II	Urease UV / GLDH II	Transasia / Erba	570	32.47	13.36	4.34	32 mg/dl	-0.11	0.36
3	CREATININE II	Jaffes Kinetic-Alkaline Picrate II	Transasia / Erba	479	2.77	11.56	0.32	2.49 mg/dl	-0.88	0.03
4	T.BILIRUBIN II	Diazonium Salt ( Colorimetric ) / Jendrassik II	Transasia / Erba	655	2.84	12.80	0.36	2.05 mg/dl	0.31	0.03
5	T-PROTEIN II	Biuret - Colorimetric II	Transasia / Erba	609	5.55	8.57	0.48	5.45 g/dl	-0.21	0.04
6	ALBUMIN II	BCG - Colorimetric II	Transasia / Erba	607	3.32	8.78	0.29	3.38 g/dl	0.21	0.02
7	URIC ACID II	Enzymatic / Uricase Colorimetric II	Transasia / Erba	686	4.95	17.29	0.86	6.33 mg/dl	1.61	0.07
8	CHOLESTROL II	CHOD-PAP II	Transasia / Erba	783	112.10	13.65	15.31	121.4 mg/dl	0.61	1.09
SDI Range		Interpretation								
Within -1.0 to +1.0		Excellent.								
Between ±1.0 to ±2.0		Good.								
Between ±2.0 to ±3.0		Accept with caution. Warning Signal.								
Beyond ±3.0		Unacceptable performance. Action Signal.								

Homogeneity and Stability of the sample is passed.

Data in CMC EQAS reports is confidential

Contact details:

Email:clinqc@cmcvellore.ac.in

Contact Number: 0416-2283102

Dr. Pamela Christudoss

CMC EQAS Co-Ordinator

Christian Medical College, Vellore

\*\*\*\*\* End of Report \*\*\*\*\*