

TEJAS DIAGNOSTICS

SHOP NO-1, NAVNATH SEVA SANGH,
 MAHARASHTRA NAGAR, MANKHURD (E)
 MANKHURD (E), MUMBAI - 400088, MAHARASHTRA - 27
 PHONE : 8689922720, 9769665565/8692055159
 Email : tejasdiagnostics2012@gmail.com
 GSTIN : 27ALGPG1915RIZR
 DL NO : 20B MH-MZ3-63649.21B MH-MZ3-63650 W.E.F.12.11.12

GANGA DIAGNOSTIC CENTER

304 - A OMPRAKASH ARCADE 3RD FLOOR, AMBEDKAR CRDE
 CHEMBUR, MUMBAI - 400071, MAHARASHTRA - 27
 PHONE : 25201159 / 25201157, 9029794923
 GSTIN : PAN :
 DL NO :
 FSSAI :
 SMan : 1 - COUNTER SALE [SUN]
 PLACE OF SUPPLY : MAHARASHTRA - 27

TAX INVOICE

PCode : 5
 Invoice No : S/175
 Inv Date : 26/06/2021 [SAT]
 Due Date : 03/07/2021 CR
 eWay No :
 Tax is Payable On Reverse Charge : No



HSN	LOC	MFG PACK	PRODUCT DESCRIPTION	QTY	FREE	BATCH NO	EXP.	CGST%	SGST%	M.R.P.	RATE	AMOUN
903118		PER 1PCS	MICROPIPETTE - V (10-100UL)	1				9.0	9.0	2000.00	868.64	868
9031		PER 1 BOX	MICRO PIPETTE 100 TO 1000UL PE	1				9.0	9.0	2000.00	868.64	868
9013		PER 1 BOX	MICRO PIPETTE 5 TO 50UL PERFEC	1				9.0	9.0	.01	868.64	868

Nett Done on 12/07/21

S

AMOUNT	SCH AMT	DISC AMT	TAXABLE	GCESS	CGST	CGST AMT	SGST	SGST AMT	TOTAL AM
0.00	0.00	0.00	0.00	0.00	0.0%	0.00	0.0%	0.00	0
0.00	0.00	0.00	0.00	0.00	2.5%	0.00	2.5%	0.00	0
2605.92	0.00	0.00	2605.92	0.00	6.0%	0.00	6.0%	0.00	0
0.00	0.00	0.00	0.00	0.00	9.0%	234.54	9.0%	234.54	3075
2605.92	0.00	0.00	2605.92	0.00	14.0%	0.00	14.0%	0.00	0
						234.54		234.54	3075

WARRANTY : I/We hereby certify that goods mentioned in this invoice is/are warranted to be of the nature and quality which it/these purpot/purpots to be.
 ToPay In Words : THREE THOUSAND SEVENTY-FIVE ONLY.

Certified that the particulars given above are true & correct and amount indicated represents the price actually charged & that there is no flow Additional consideration directly or indirectly from the buyer.

For TEJAS DIAGNOSTICS

S.N. Ghore

CN Ref :
 Cust Ref :
 DN Ref :
 Total Prod : 3 Total QTY : 3

Total Amt : 3075.
 AD/LS Amt : 0.
 Round Off : 0.
 TO PAY ₹ 3075.

1 [1:43:22 PM]
 TRIPLICATE FOR SUPPLIER Chkd By Pckd By Competent Signatory

TEJAS DIAGNOSTICS

DEALERS IN LABORATORY, HOSPITAL REQUISITES & PHARMACEUTICALS

Shop No. 1, Navnath Seva Sangh, Maharashtra Nagar, Mankhurd (East), Mumbai - 400 088.

Mob. 8692055159 / 9769665565 / 8689922720

E-mail. vikas.ghare83@gmail.com

DATE - 26-06-2021

CERTIFICATE OF INSTALLATION

Client Name : Ganga Diagnostic Centre

Product Name : Micropipette V 10-100ul

Micropipette V 100-1000ul

Micropipette V 5-50ul

Make : Perfect

Installation Date : 26-06-2021

Ghare



Perfect

STATUS PASSED

CALIBRATION REPORT

DEVICE ID MICROPIPETTE
METHOD ID VV1000

DESCRIPTION VARY VOL 100-1000ul
DATE 11/06/2021

ENVIRONMENTAL FACTORS

TEMP. 20.00C BARO. PRESSURE 750.00 mmHg REL. HUMIDITY 50.00%
AIR DENSITY 0.001200g/cm³ Z FACTOR 1.002899cm³/g CUBIC EXP. 0.0000cm³/degC

SAMPLE

CALIBRATION STATISTICS

Vol	No.	Wt.	Vol	Vol	Mean	Sd	Actual	Target	Status	Imprecision CV%
100.0	1	99.80	100.10	100.0	100.23	0.094	0.23	0.50	PASSED	<0.60
100.0	2	199.80	100.30	500.0	499.14	1.056	0.17	0.60	PASSED	<0.20
100.0	3	299.80	100.30	1000.0	1000.96	1.861	0.09	0.50	PASSED	<0.20
500.0	1	497.80	499.24							
500.0	2	996.50	500.15							
500.0	3	1493.10	498.14							
1000.0	1	1000.00	1002.90							
1000.0	2	1997.90	1000.79							
1000.0	3	2994.20	999.19							

Perfect

STATUS PASSED

CALIBRATION REPORT

DEVICE ID MICROPIPETTE

DESCRIPTION VARY VOL 5-50ul

METHOD ID VV 50

DATE

16/03/2021

ENVIRONMENTAL FACTORS

TEMP. 20.00 C

BARO. PRESSURE 750.00 mmHg

REL. HUMIDITY 50.00%

AIR DESITY 0.001200g/cm3

Z FACTOR 1.002899 cm3/g

CUBIC EXP.0.0000cm3/deg C

SAMPLE

CALIBRATION STATISTICS

Vol	No.	Wt.	Vol	Vol	Inaccuracy E%		Actual	Imprecision CV%	
					Mean	Sd		Target	Status
5.0	1	4.90	4.91	5.0	5.11	0.200	2.29	2.50	PASSED <1.50
5.0	2	10.00	5.11	25.0	24.97	0.173	0.11	0.70	PASSED <0.30
5.0	3	15.30	5.32	50.0	50.07	0.058	0.15	0.70	PASSED <0.30
25.0	1	25.00	25.07						
25.0	2	49.70	24.77						
25.0	3	74.70	25.07						
50.0	1	49.90	50.04						
50.0	2	99.90	50.15						
50.0	3	149.80	50.04						

PerFect

STATUS PASSED

CALIBRATION REPORT

DEVICE ID MICROPIPETTE

DESCRIPTION VARY VOL 10-100ul

METHOD ID VV 100

DATE 14/04/2021

ENVIRONMENTAL FACTORS

TEMP.20.00C

BARO.PRESSURE 750.00 mmHg

REL.HUMIDITY 50.00%

AIRDENSITY 0.001200g/cm3

FACTOR 1.002899 cm3/g

CUBIC EXP. 0.0000cm3/degC

SAMPLE

CALIBRATION STATISTICS

VOL	NO.	Wt.	VOL.	VOL	Mean	Sd	Inaccuracy E%		Imprecision CV%	
							Actual	Target	Status	
10.0	1	9.90	9.92	10.0	9.95	0.052	0.5	0.20	PASSED	<0.6
10.0	2	19.90	10.03	50.0	49.84	0.081	0.8	0.60	PASSED	<0.3
10.0	3	29.80	9.92	100.0	100.03	0.313	0.9	0.50	PASSED	<0.3
50.0	1	49.80	49.94							
50.0	2	99.80	49.84							
50.0	3	149.10	49.74							
100.0	1	100.00	100.30							
100.0	2	199.90	100.30							
100.0	3	299.20	99.39							