

CERTIFICATE OF CALIBRATION

Ref. No.: MSM003
Applicant : MEDICITY SPECIALITY MEDICAL LABORATORY
Device Calibrated : CENTRIFUGE
Make/Model: LABTECH/C-POULO
Sl.No: —

Condition at receipt : Functional
Calibrated at : Onsite

Date of Calibration : 10/01/2021

Calibration due date : 09/01/2022

Certified that the above Instrument has been calibrated by trained technical personnel of this centre using a calibration system traceable to National Standards. The calibration details attached with the certificate are authentic quantitative analysis of the instrument's credibility on the date of calibration and relate only to the Instrument calibrated.

Date of Issue : 11/01/2021

Approved Signatory



Cyrix Healthcare Pvt. Ltd.,
30/641 B Petta Junction, Poonithura, Kochi-38.
Ph: +91 484 6544033, Mobile: 97455 55073
Email: sales@cyrixhealthcare.com
service@cyrixhealthcare.com

CYRIX
HEALTH CARE PVT LTD

www.cyrixhealthcare.com



ISO 9001:2008 ISO 13485:2012
Certified Company

AERB Approved
QA Agency

CERTIFICATE OF CALIBRATION

Customer Name & Address	MEDICITY SPECIALITY MEDICAL LABORATORY THALIPARAMBA, KERALA 670141	Certificate No.	MSM003
		Date of Calibration	10/01/2021
		Valid Upto	09/01/2022
		Date of Issue	11/01/2021

Instruments Details

Instrument Name	CENTRIFUGE	S No.	----
Make/Model No.	LABTECH/C-POULO	Location	LAB
ID No.	-----	Visual Inspection	OK
RANGE	0-3500 RPM		

Details of Reference Standards & Major Equipments Used

Equipment Name	Make/Model	Serial No:	Calibration Validity	Calibration By
DIGITAL TACHOMETER	METRAVI/NCTM 1000	170603632	09/01/2022	UNIQUE CALIBRATION
DIGITAL THERMOHYGROMETER	-/HTC-2	...	22/09/2021	NENO TECHNICAL SERVICES

Environmental Condition

Temperature	24.8°C	Relative Humidity	54.5 %
-------------	--------	-------------------	--------

Calibration Results

SL NO	SET READING (RPM)	MEASURED READING (RPM)	TOLERANCE	STATUS
1	1000	993	±5%	PASS
2	1500	1512	±5%	PASS
3	2500	2543	±5%	PASS
4	3000	3042	±5%	PASS
5	3500	3521	±5%	PASS

REMARKS:-

- ❖ This result of calibration refers to only to the particular items submitted for calibration.
- ❖ The above result is valid at the time of and under the stated conditions measurement.
- ❖ Calibration certificate issued for weight & measure parameters like mass, weighting balance, volumetric equipment, measuring scales/tapes etc. Are scientific purpose only and should not used for commercial.

Calibrated By
(SE)

Approved By
(QM)
SARATH K B



Cyrix Healthcare Pvt. Ltd.,
30/641 B Petta Junction, Poonithura, Kochi-39.
Ph: +91 484 6544033, Mobile: 97455 55073
Email: sales@cyrixhealthcare.com
service@cyrixhealthcare.com

CYRIX
HEALTH CARE PVT LTD

www.cyrixhealthcare.com