

# CERTIFICATE OF CALIBRATION

Customer Name & Address	<b>MEDICITY SPECIALITY MEDICAL LABORATORY, THALIPARAMBA</b>	<b>Certificate No.</b>	CHPL/PC/MSL/001
		Date of Calibration	10/01/2021
		Valid Upto	09/06/2021
		Date of Issue	10/01/2021

### Instruments Details

Instrument Name	MICRO PIPETTE	S No.	..
Make/Model No.	AVANTOR	Visual Inspection	OK
Range	10-100 µl Variable	Least Count	1 µl

### Details of Reference Standards & Major Equipments Used

Equipment Name	ANALYTICAL BALANCE	THERMO HYGROMETER
Make	WENSAR	FANEM
Model/SR No.	MAB182/39736	MOD3620/RAH85394
Calibration Validity	13-02-2021	25/09/2021
Calibration By	STIC	FCRI PALAKKAD

### Environmental Condition

Temperature	25.6°C	Relative Humidity	40.1%
-------------	--------	-------------------	-------

### Calibration Results

SL NO	SET READING (µl)	MEASURED READING (µl)					STATUS
		I	II	III	IV	V	
1	10	9.99	10.02	9.99	10.08	9.98	P
2	50	50.15	49.95	49.89	50.25	50.08	P
3	100	100.01	100.01	99.99	100.00	99.98	P

#### REMARKS:- Tolerance value $\pm 2\%$

- ❖ This result of calibration refers to only to the particular items submitted for calibration.
- ❖ The above result is valid at the time of and under the stated conditions measurement.
- ❖ Calibration certificate issued for weight & measure parameters like mass, weighting balance, volumetric equipment, measuring scales/tapes etc. Are scientific purpose only and should not used for commercial.

Calibrated By

(SE)



Approved By

(QM)

SARATHK B

**Cyrix Healthcare Pvt. Ltd.,**  
30/641 B Petta Junction, Poonithura, Kochi-39.  
Ph: +91 484 6544033, Mobile: 97455 55073  
Email: sales@cyrixhealthcare.com  
service@cyrixhealthcare.com

**CYRIX**  
HEALTH CARE PVT LTD

[www.cyrixhealthcare.com](http://www.cyrixhealthcare.com)