

CERTIFICATE OF CALIBRATION

Customer Name & Address	MEDICITY SPECIALITY MEDICAL LABORATORY, THALIPARAMBA	Certificate No.	CHPL/PC/MSL/002
		Date of Calibration	10/01/2021
		Valid Upto	09/06/2021
		Date of Issue	10/01/2021

Instruments Details

Instrument Name	MICRO PIPETTE	S No.	..
Make/Model No.	AVANTOR	Visual Inspection	OK
Range	100-1000 µl Variable	Least Count	5µl

Details of Reference Standards & Major Equipments Used

Equipment Name	ANALYTICAL BALANCE	THERMO HYGROMETER
Make	WENSAR	FANEM
Model/SR No.	MAB182/39736	MOD3620/RAH85394
Calibration Validity	13-02-2021	25/09/2021
Calibration By	STIC	FCRI PALAKKAD

Environmental Condition

Temperature	25.6°C	Relative Humidity	40.1%
-------------	--------	-------------------	-------

Calibration Results

SL NO	SET READING (µl)	MEASURED READING (µl)					STATUS
		I	II	III	IV	V	
1	100	99.96	100.08	99.98	100.10	99.98	P
2	500	500.25	499.95	499.88	500.24	500.18	P
3	1000	1000.01	1000.00	1000.18	1000.48	1000.45	P

REMARKS:- **Tolerance value ±8%**

- ❖ This result of calibration refers to only to the particular items submitted for calibration.
- ❖ The above result is valid at the time of and under the stated conditions measurement.
- ❖ Calibration certificate issued for weight & measure parameters like mass, weighting balance, volumetric equipment, measuring scales/tapes etc. Are scientific purpose only and should not used for commercial.

Calibrated By

(SE)



Approved By

(QM)

SARATHK B

Cyrix Healthcare Pvt. Ltd.,
30/641 B Potta Junction, Poonithura, Kochi-39.
Ph: +91 484 6544033, Mobile: 97455 55073
Email: sales@cyrixhealthcare.com
service@cyrixhealthcare.com

CYRIX
HEALTH CARE PVT LTD

www.cyrixhealthcare.com