## **CERTIFICATE OF CALIBRATION**

Customer Name & Address	MEDICITY SPECIALITY MEDICAL LABORATORY,THALIPARAMBA	Certificate No.	CHPL/PC/MSL/003
		Date of Calibration	10/01/2021
		Valid Upto	09/06/2021
		Date of Issue	10/01/2021

Instruments Details					
Instrument Name	MICRO PIPETTE	S No.			
Make/Model No.	THERMO SCIENTIFIC	Visual Inspection	ОК		
Range	1000 μl Fixed				

Details of Reference Standards & Major Equipments Used				
Equipment Name	ANALYTICAL BALANCE	THERMO HYGROMETER		
Make	WENSAR	FANEM		
Model/SR No.	MAB182/39736	MOD3620/RAH85394		
Calibration Validity	13-02-2021	25/09/2021		
Calibration By	STIC	FCRI PALAKKAD		

Environmental Condition				
Temperature	25.6°C	Relative Humidity	40.1%	

Calibration Results							
SL NO SET READING ΜΕΑSURED READING (μΙ)					STATUS		
SLINU	(μl)	I	Π		IV	V	STATUS
3	1000	1000.81	999.45	999.56	1000.38	1000.78	Р

## REMARKS:- Tolerance value ±4%

- This result of calibration refers to only to the particular items submitted for calibration.
- ♦ The above result is valid at the time of and under the stated conditions measurement.
- Calibration certificate issued for weight & measure parameters like mass, weighting balance, volumetric equipment, measuring scales/tapes etc. Are scientific purpose only and should not used for commercial.

## **Calibrated By**





Cyrix Healthcare Pvt. Ltd., 30/641 B Petta Junction, Poonithura, Kochi-38.

Ph: +91 484 6544033, Mobile: 97455 55073 Email: sales@cyrixhealthcare.com service@cyrixhealthcare.com



www.cyrixhealthcare.com