

LAB MONTHLY SUMMARY



Lab Name **MEDICITY MEDICAL LABORATORY** Lab No **5948**
 Month **May** Year **2021**
 Constituent Group **Chemistry II**



Date of Result Entered : 17/05/2021

Date of Report Published : 05/06/2021

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE II	GOD-POD II	Mindray	129	133.59	8.67	11.58	145 mg/dl	0.99	2.04
2	UREA II	Urease UV / GLDH II	Mindray	107	64.54	11.57	7.47	72.5 mg/dl	1.07	1.44
3	CREATININE II	Jaffes End point II	Any Analyser (Automation / Semi Automation)	601	1.85	15.60	0.29	2 mg/dl	0.52	0.02
4	T.BILIRUBIN II	Diazonium Salt (Colorimetric) / Jendrassik II	Mindray	110	2.52	17.62	0.44	3 mg/dl	1.08	0.08
5	T-PROTEIN II	Biuret - Colorimetric II	Mindray	120	5.42	11.27	0.61	4.9 g/dl	-0.85	0.11
6	ALBUMIN II	BCG - Colorimetric II	Mindray	119	3.36	11.48	0.39	3.9 g/dl	1.40	0.07
7	URIC ACID II	Enzymatic / Uricase Colorimetric II	Mindray	116	6.89	14.60	1.01	6.8 mg/dl	-0.09	0.19
8	CHOLESTROL II	CHOD-PAP II	Mindray	136	112.55	11.60	13.06	124 mg/dl	0.88	2.24
9	TRIGLYCERIDE II	GPO-PAP / Enzymatic Colorimetric / End Point II	Mindray	123	223.20	9.92	22.13	235 mg/dl	0.53	3.99

SDI Range	Interpretation
Within -1.0 to +1.0	Excellent.
Between ± 1.0 to ± 2.0	Good.
Between ± 2.0 to ± 3.0	Accept with caution. Warning Signal.
Beyond ± 3.0	Unacceptable performance. Action Signal.

Homogeneity and Stability of the sample is passed.

Data in CMC EQAS reports is confidential

Contact details:

Email: clinqc@cmcvellore.ac.in

Contact Number: 0416-2283102

Pamela Christudoss

Dr. Pamela Christudoss
 CMC EQAS Co-Ordinator

Christian Medical College, Vellore

***** End of Report *****