



ASIAN TECHNOLOGY

(A HOUSE OF CALIBRATION)
Cert. No. : CC-2239



CALIBRATION CERTIFICATE

Customer Name And Address	M/S: Ashok Nursing Home. F-3/13-16, Vijay Chowk, Krishan Nagar, Delhi-110051.	Certificate No.	CC223920600023521 F
		Service request No & date	P/01/20/08/2020
		Date of calibration	20/08/2020
		Valid up to	19/08/2021
		Date of issue	21/08/2020

Instrument Details			
Instrument name	Digital Weighing Balance	Sr. No.	1402086
Make /Model No	Kerro/BLC 3002	Location	Path Lab
Range / Size	0 to 300 gm	Accuracy	-----
Least Count	0.01 gm	Visual Inspection	OK
I.D. No.	DWB-01		

Detail of reference standards & Major equipments used			
Equipment Name	A set of weight box		
Make	Weigh India		
Model / SR No.	-----		
Certificate No.	WMCL/E/2018-12/2533		
Calibration Validity	26/12/2021		
Calibration By	WEIGHTRONICS		


Environmental Condition	Temperature	23± 3 °C	Calibration Reference	OIML R 76
	Relative Humidity	50±20%	Calibration Procedure	CP-27

Calibration Results			
Serial No.	Applied Weight in	Actual Value in (g)	Uncertainty At 95% C.L. (coverage factor K=2)
01.	500 mg	0.50	±0.36 mg
02.	1 gm	0.99	±0.36 mg
03.	2 gm	2.00	±0.36 mg
04.	5 gm	5.00	±0.36 mg
05.	10 gm	10.00	±0.36 mg
06.	20 gm	20.00	±0.36 mg
07.	50 gm	50.00	±0.36 mg
08.	100 gm	100.01	±0.36 mg
09.	200 gm	199.98	±0.36 mg
10.	250 gm	249.98	±6 g

Remarks:

<ul style="list-style-type: none"> ❖ This result of calibration refer to only to the particular item submitted for calibration. ❖ The above result are valid at the time of and under the stated conditions measurement. ❖ Calibration certificate issued for weight & measure parameters like mass, weighting balance, volumetric equipment, measuring scales/tapes etc. are scientific purpose only and should not used for commercial.
--

Calibrated By
(Calibration Engg./TM)
(MAMPAL)
Form No.-32



Approved By
(QM/TM)
(NEERAJ TYAGI)
Page No. 1 of 1

---End Of Report---