



# ASIAN TECHNOLOGY

(A HOUSE OF CALIBRATION)  
Cert. No. : CC-2239



## CALIBRATION CERTIFICATE

|                                 |   |                           |                     |
|---------------------------------|---|---------------------------|---------------------|
| Customer Name<br>And<br>Address | M/S: Ashok Nursing Home.<br>F-3/13-16, Vijay Chowk, Krishan<br>Nagar, Delhi-110051. | Certificate No.           | CC223920600023523 F |
|                                 |   | Service request No & date | P/01/20/08/2020     |
|                                 |   | Date of calibration       | 20/08/2020          |
|                                 |   | Valid up to               | 19/08/2021          |
|                                 |   | Date of issue             | 21/08/2020          |

| Instrument Details |               |                   |                       |
|--------------------|---------------|-------------------|-----------------------|
| Instrument name    | RPM Meter     | Sr. No.           | CLLC-2421             |
| Make               | Remi          | Location          | Centrifuge (Path Lab) |
| Range / Size       | 0 to 5000 rpm | Accuracy          | -----                 |
| Least Count        | 10 rpm        | Visual Inspection | OK                    |
| I.D. No.           | RPM-01        |                   |                       |

| Detail of reference standards & Major equipments used |                    |  |  |
|---|--------------------|--|--|
| Equipment Name  | Digital Tachometer |  |  |
| Make  | Lutron             |  |  |
| Model / SR No.  | AC92792            |  |  |
| Certificate No.                                       | CC237919000001774F |  |  |
| Calibration Validity                                  | 30.10.2020         |  |  |
| Calibration by  | KCL                |  |  |


|                            |                   |         |                       |           |
|----------------------------|-------------------|---------|-----------------------|-----------|
| Environmental<br>Condition | Temperature       | 20±2°C  | Calibration Reference | IS: 12508 |
|                            | Relative Humidity | 50±10 % | Calibration Procedure | CP-37     |

| Calibration Results |                            |                            |  |
|---------------------|----------------------------|----------------------------|--|
| Serial No.          | Standard value in<br>(rpm) | Measured value in<br>(rpm) | Uncertainty At 95% C.L.<br>(coverage factor K=2) |
| 01.                 | 200                        | 199                        | ±0.20%rdg  |
| 02.                 | 500                        | 501                        |  |
| 03.                 | 1000                       | 1002                       |  |
| 04.                 | 2000                       | 2004                       |  |
| 05.                 | 2995                       | 3006                       |  |
| 06.                 | 4950                       | 4954                       |  |

### Remarks:

- ❖ This result of calibration refers to only to the particular item submitted for calibration.
- ❖ The above result is valid at the time of and under the stated conditions measurement.
- ❖ Calibration certificate issued for weight & measure parameters like mass, weighting balance, volumetric equipment, measuring scales/tapes etc. are scientific purpose only and should not used for commercial.

Calibrated By  
(Calibration Engg./TM)  
(MAHPAL)  
Form No.-32



Approved By  
(OM/TM)  
(NEERAJ TVAGI)  
Page No. 1 of 1

---End Of Report---